

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 10/4/2018

Week #	1	Dates:	12/30/2018-1/5/2019	Counted On	1/4/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.		uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100	
Limit of quantification (LOQ)=		8.89E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	

Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	28	*****	*****	
2.	BACKGROUND 2	29	*****	*****	
3.	BACKGROUND 3	31	*****	*****	
4.	BACKGROUND 4	29	*****	*****	
5.	BACKGROUND 5	32	*****	*****	
6.	BACKGROUND 6	27	*****	*****	
7.	ELEVATOR ENTRANCE	32	M	M	
8.	FENCED GATE	30	M	M	
9.	EAST FLOOR	29	M	M	
10.	NORTH BENCH TOP	33	M	M	
11.	WEST LAB FLOOR	39	1.35E-05	1.35E-07	
12.	WORK PLATFORM	29	M	M	
13.	CALORIMETER	30	M	M	
14.	COMPUTERS	32	M	M	
15.	DSC/TGA	36	M	M	
16.	SOUTH BENCH TOP	32	M	M	

Notes: NONE

³⁶ CI SOURCE 09/14/1971	50116	*****	*****	
	101.05%	CALIBRATION CHECK OK?		YES

Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1×10^{-6} uCi/cm².

SURVEY PREPARED BY: M. Heine	REVIEWED BY: <i>C. Hines</i>
SIGN/DATE: <i>M. Heine</i> 1/4/2019	SIGN/DATE: <i>C. Hines</i> 1/7/2019

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190104_1425

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190104_1425\20190104_1425.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	1/4/2019	2:26:52 PM	1	5.00		21	7	28	272.22	
1	1/4/2019	2:33:00 PM	2	5.00		23	6	29	311.75	
1	1/4/2019	2:39:18 PM	3	5.00		26	6	31	224.31	
1	1/4/2019	2:45:27 PM	4	5.00		24	4	29	212.23	
1	1/4/2019	2:51:43 PM	5	5.00		26	6	32	231.96	
1	1/4/2019	2:57:59 PM	6	5.00		22	5	27	237.45	
1	1/4/2019	3:04:18 PM	7	5.00		23	9	32	368.46	
1	1/4/2019	3:10:31 PM	8	5.00		24	6	30	237.05	
1	1/4/2019	3:16:50 PM	9	5.00		25	5	29	191.57	
1	1/4/2019	3:23:03 PM	10	5.00		26	7	33	260.37	
1	1/4/2019	3:29:26 PM	11	5.00		32	7	39	135.63	
1	1/4/2019	3:35:40 PM	12	5.00		24	5	29	238.10	

1	1/4/2019	3:42:02 PM	13	5.00	22	8	30	247.32
1	1/4/2019	3:48:16 PM	14	5.00	25	6	32	206.88
1	1/4/2019	3:54:25 PM	15	5.00	29	7	36	241.26
1	1/4/2019	4:00:39 PM	16	5.00	26	7	32	279.11
1	1/4/2019	4:06:18 PM	17	5.00	14477	35639	50116	999.10


LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 1/4/19	Time 0910	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) <input checked="" type="checkbox"/> ON <input type="checkbox"/> OFF <input type="checkbox"/>					

LSC Printout Date: 1/4/2019

NOTE SWIPE AND SURVEY LOCATIONS ON MAP



Surveyed By: CMH	Date: 1/4/2019	Notes Reactor operation at 10 Watts for maintenance	
Swiped By: CS	Date: 1/4/2019		
Reviewed By:	Date:		

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 10/4/2018

Week #	1	Dates:	1/6/2019-1/12/2019	Counted On	1/8/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		9.03E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	31	*****	*****	
2.	BACKGROUND 2	29	*****	*****	
3.	BACKGROUND 3	34	*****	*****	
4.	BACKGROUND 4	29	*****	*****	
5.	BACKGROUND 5	30	*****	*****	
6.	BACKGROUND 6	34	*****	*****	
7.	ELEVATOR ENTRANCE	30	M	M	
8.	FENCED GATE	31	M	M	
9.	EAST FLOOR	33	M	M	
10.	NORTH BENCH TOP	31	M	M	
11.	WEST LAB FLOOR	31	M	M	
12.	WORK PLATFORM	32	M	M	
13.	CALORIMETER	29	M	M	
14.	COMPUTERS	31	M	M	
15.	DSC/TGA	34	M	M	
16.	SOUTH BENCH TOP	30	M	M	
Notes: NONE					
³⁶ CI SOURCE 09/14/1971		50121		*****	*****
		101.06%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: M. Heine			REVIEWED BY: <i>a.Hines</i>		
SIGN/DATE: <i>Madison Heine</i> 1/10/19			SIGN/DATE: <i>allison</i> 1/11/2019		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190108_1343

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190108_1343\20190108_1343.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	1/8/2019	1:44:46 PM	1	5.00	26	5	31	178.38		
1	1/8/2019	1:51:03 PM	2	5.00	23	5	29	267.42		
1	1/8/2019	1:57:19 PM	3	5.00	29	5	34	235.54		
1	1/8/2019	2:03:34 PM	4	5.00	22	7	29	299.34		
1	1/8/2019	2:09:43 PM	5	5.00	25	5	30	226.86		
1	1/8/2019	2:15:52 PM	6	5.00	28	6	34	225.94		
1	1/8/2019	2:22:10 PM	7	5.00	23	6	30	192.95		
1	1/8/2019	2:28:24 PM	8	5.00	25	6	31	190.68		
1	1/8/2019	2:34:42 PM	9	5.00	27	5	33	206.91		
1	1/8/2019	2:40:50 PM	10	5.00	27	4	31	144.31		
1	1/8/2019	2:47:07 PM	11	5.00	27	4	31	164.83		
1	1/8/2019	2:53:23 PM	12	5.00	25	7	32	236.84		

1	1/8/2019	2:59:43	PM	13	5.00	23	6	29	194.64
1	1/8/2019	3:05:57	PM	14	5.00	26	5	31	193.33
1	1/8/2019	3:12:08	PM	15	5.00	29	5	34	189.27
1	1/8/2019	3:18:22	PM	16	5.00	23	7	30	303.42
1	1/8/2019	3:24:01	PM	17	5.00	14479	35642	50121	1007.29


LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 1/8/19	Time 1316	Swipe Instrument Quantasmart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) <input type="checkbox"/> ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>					

LSC Printout Date 1/10/2019

NOTE SWIPE AND SURVEY LOCATIONS ON MAP



Survey #	Location	CPM
1	Elevator Entrance	≤ 100
2	South Floor	≤ 100
3	East Floor	1000
4	North Bench Top	≤ 100
5	West Floor	≤ 100
6	Work Platform	≤ 100
7	Calorimeter	≤ 100
8	Computers	≤ 100
9	DSC/TGA	200
10	South Bench Top	150
11	Background	
12	Elevator Entrance	
13	South Floor	
14	East Floor	
15	North Bench Top	
16	West Lab Floor	
17	Work Platform	
18	Calorimeter	
19	Computers	
20	DSC/TGA	

Notes: High survey reading at location 3 because of radioactive material safe

Surveyed By: AG/PW Date: 1/8/2019

Swiped By: PW Date: 1/8/2019

Reviewed By: [Signature] Date: 1/11/2019

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 10/4/2018

Week #	3	Dates:	1/13/2019-1/19/2019	Counted On	1/17/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm2 = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.95E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	

Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	32	*****	*****	
2.	BACKGROUND 2	29	*****	*****	
3.	BACKGROUND 3	26	*****	*****	
4.	BACKGROUND 4	30	*****	*****	
5.	BACKGROUND 5	34	*****	*****	
6.	BACKGROUND 6	30	*****	*****	
7.	B021 ELEVATOR ENTRANCE	28	M	M	
8.	B021 FENCED GATE	32	M	M	
9.	B021 EAST FLOOR	31	M	M	
10.	B021 NORTH BENCH TOP	27	M	M	
11.	B021 WEST LAB FLOOR	30	M	M	
12.	B021 WORK PLATFORM	29	M	M	
13.	B021 CALORIMETER	32	M	M	
14.	B021 COMPUTERS	33	M	M	
15.	B021 DSC/TGA	34	M	M	
16.	B021 SOUTH BENCH TOP	35	M	M	

Notes: NONE

³⁶ Cl SOURCE 09/14/1971	50145	*****	*****		
	101.11%	CALIBRATION CHECK OK?		YES	

Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10⁻⁶ uCi/cm².

SURVEY PREPARED BY: M. Heine	REVIEWED BY: <i>c. l. bauer</i>
------------------------------	---------------------------------

SIGN/DATE: <i>Modellias/Heine 1/17/2019</i>	SIGN/DATE: <i>awjdm 1/18/2019</i>
---	-----------------------------------

Assay Definition-

Assay Description:

Assay Type: CPM
 Report Name: Report1
 Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190117_1454
 Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190117_1454\20190117_1454.results
 Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor
 Quench Indicator: tSIE/AEC
 External Std Terminator (sec): 0.5 2s%
 Pre-Count Delay (min): 0.00
 Quench Set: n/a
 Count Time (min): 5.00
 Count Mode: Normal
 Assay Count Cycles: 1 Repeat Sample Count: 1
 #Vials/Sample: 1 Calculate % Reference: Off

Background Subtract: Off
 Low CPM Threshold: Off
 2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On Luminescence Correction: Off
 Colored Samples: n/a Heterogeneity Monitor: n/a
 Coincidence Time (nsec): 18 Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	1/17/2019	2:55:22 PM	1	5.00		25	6	32	298.01	
1	1/17/2019	3:01:31 PM	2	5.00		23	6	29	254.06	
1	1/17/2019	3:07:48 PM	3	5.00		20	7	26	330.76	
1	1/17/2019	3:13:58 PM	4	5.00		24	6	30	225.66	
1	1/17/2019	3:20:12 PM	5	5.00		26	7	34	261.80	
1	1/17/2019	3:26:29 PM	6	5.00		24	5	30	195.18	
1	1/17/2019	3:32:48 PM	7	5.00		23	5	28	198.02	
1	1/17/2019	3:39:02 PM	8	5.00		25	7	32	257.42	
1	1/17/2019	3:45:19 PM	9	5.00		25	6	31	162.63	
1	1/17/2019	3:51:32 PM	10	5.00		19	8	27	284.03	
1	1/17/2019	3:57:49 PM	11	5.00		24	6	30	186.11	
1	1/17/2019	4:04:04 PM	12	5.00		22	7	29	181.36	

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	1/17/2019	4:10:22 PM	13	5.00	23	9	32	280.15
1	1/17/2019	4:16:40 PM	14	5.00	26	7	33	227.88
1	1/17/2019	4:22:57 PM	15	5.00	28	6	34	140.81
1	1/17/2019	4:29:12 PM	16	5.00	27	7	35	206.46
1	1/17/2019	4:34:52 PM	17	5.00	14542	35603	50145	1007.54

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building	Dodger Research Facility	Room	B0021 - GUO	Lab Class	D	Survey Instrument	GM Detector Model 3	Serial No.	58680	Radiation Detected	Gamma
Authorized User	Xiaofeng Guo	Date	1/17/2019	Time	1431	Swipe Instrument	Quantasmat	Serial No.	073396	Radiation Detected	Beta/Gamma
Reactor Status (Check One)	<input checked="" type="checkbox"/> ON <input type="checkbox"/> OFF										

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP								
1	Elevator Entrance	140	<p style="text-align: center;">DODG B0021 GUO LAB</p>								
2	South Floor	< 100									
3	East Floor	200									
4	North Bench Top	< 100									
5	West Floor	< 100									
6	Work Platform	< 100									
7	Calorimeter	< 100									
8	Computers	< 100									
9	DSC/TGA	< 100									
10	South Bench Top	< 100									
Swipe # and Locations 1-6 Background 16 South Bench Top 7 Elevator Entrance 8 South Floor 9 East Floor 10 North Bench Top 11 West Lab Floor 12 Work Platform 13 Calorimeter 14 Computers 15 DSC/TGA											
			Surveyed By: <i>SW</i> Date: 1/17/19 Swiped By: <i>BT</i> Date: 1/17/19 Reviewed By: <i>caufman</i> Date: 1/18/2019								
			Notes								

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 10/4/2018

Week #	4	Dates:	1/20/2019-1/27/2019	Counted On	1/22/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.80E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	

Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	29	*****	*****	
2.	BACKGROUND 2	27	*****	*****	
3.	BACKGROUND 3	23	*****	*****	
4.	BACKGROUND 4	26	*****	*****	
5.	BACKGROUND 5	30	*****	*****	
6.	BACKGROUND 6	34	*****	*****	
7.	B021 ELEVATOR ENTRANCE	29	M	M	
8.	B021 FENCED GATE	30	M	M	
9.	B021 EAST FLOOR	27	M	M	
10.	B021 NORTH BENCH TOP	31	M	M	
11.	B021 WEST LAB FLOOR	34	M	M	
12.	B021 WORK PLATFORM	30	M	M	
13.	B021 CALORIMETER	26	M	M	
14.	B021 COMPUTERS	23	M	M	
15.	B021 DSC/TGA	30	M	M	
16.	B021 SOUTH BENCH TOP	26	M	M	

Notes: NONE

³⁶ Cl SOURCE 09/14/1971	50210	*****	*****	
	101.24%	CALIBRATION CHECK OK?		YES

Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10⁻⁶ uCi/cm².

SURVEY PREPARED BY: M. Heine	REVIEWED BY: <i>C. Hines</i>
SIGN/DATE: <i>Madeline Heine 1/23/19</i>	SIGN/DATE: <i>anyone 1/23/2019</i>

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190122_1819

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190122_1819\20190122_1819.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	1/22/2019	6:20:55 PM	1		5.00	23	6	29	204.52	
1	1/22/2019	6:27:05 PM	2		5.00	22	5	27	185.38	
1	1/22/2019	6:33:20 PM	3		5.00	20	3	23	114.51	
1	1/22/2019	6:39:33 PM	4		5.00	21	5	26	212.34	
1	1/22/2019	6:45:41 PM	5		5.00	24	6	30	296.60	
1	1/22/2019	6:51:49 PM	6		5.00	29	5	34	188.19	
1	1/22/2019	6:58:03 PM	7		5.00	23	6	29	247.91	
1	1/22/2019	7:04:16 PM	8		5.00	26	4	30	154.92	
1	1/22/2019	7:10:34 PM	9		5.00	22	4	27	148.45	
1	1/22/2019	7:16:43 PM	10		5.00	26	5	31	202.85	
1	1/22/2019	7:22:59 PM	11		5.00	29	5	34	168.26	
1	1/22/2019	7:29:09 PM	12		5.00	24	6	30	187.95	

1	1/22/2019	7:35:27 PM	13	5.00	21	5	26	214.66
1	1/22/2019	7:41:36 PM	14	5.00	19	5	23	195.50
1	1/22/2019	7:47:51 PM	15	5.00	25	6	30	207.67
1	1/22/2019	7:54:04 PM	16	5.00	21	5	26	272.36
1	1/22/2019	7:59:43 PM	17	5.00	14472	35739	50210	1009.26

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 10/4/2018

Week #	5	Dates:	1/27/2019-2/02/2019	Counted On	1/29/2019
INSTRUMENT USED: Quantasart (Serial# 073396)			CONTROL FACTOR: 1.40E-06		
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		9.16E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	

Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	34	*****	*****	
2.	BACKGROUND 2	30	*****	*****	
3.	BACKGROUND 3	28	*****	*****	
4.	BACKGROUND 4	30	*****	*****	
5.	BACKGROUND 5	36	*****	*****	
6.	BACKGROUND 6	41	*****	*****	
7.	B021 ELEVATOR ENTRANCE	36	M	M	
8.	B021 FENCED GATE	23	M	M	
9.	B021 EAST FLOOR	31	M	M	
10.	B021 NORTH BENCH TOP	28	M	M	
11.	B021 WEST LAB FLOOR	34	M	M	
12.	B021 WORK PLATFORM	33	M	M	
13.	B021 CALORIMETER	30	M	M	
14.	B021 COMPUTERS	27	M	M	
15.	B021 DSC/TGA	37	M	M	
16.	B021 SOUTH BENCH TOP	29	M	M	

Notes: NONE

³⁶ Cl SOURCE 09/14/1971	50228	*****	*****	
	101.27%	CALIBRATION CHECK OK?		YES

Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10⁻⁶ uCi/cm².

SURVEY PREPARED BY: M. Heine	REVIEWED BY: <i>C. Heine</i>
SIGN/DATE: <i>M. Heine</i> 1/30/19	SIGN/DATE: <i>C. Heine</i> 1/30/2019

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190129_1458

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190129_1458\20190129_1458.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	1/29/2019	2:59:27 PM	1		5.00	27	7	34	275.64	
1	1/29/2019	3:05:42 PM	2		5.00	23	7	30	340.82	
1	1/29/2019	3:11:58 PM	3		5.00	23	5	28	266.33	
1	1/29/2019	3:18:12 PM	4		5.00	23	7	30	294.57	
1	1/29/2019	3:24:21 PM	5		5.00	30	6	36	244.42	
1	1/29/2019	3:30:30 PM	6		5.00	33	8	41	264.59	
1	1/29/2019	3:36:44 PM	7		5.00	27	9	36	335.45	
1	1/29/2019	3:42:58 PM	8		5.00	18	5	23	261.14	
1	1/29/2019	3:49:15 PM	9		5.00	25	6	31	253.40	
1	1/29/2019	3:55:23 PM	10		5.00	24	4	28	181.01	
1	1/29/2019	4:01:40 PM	11		5.00	27	7	34	222.14	
1	1/29/2019	4:07:52 PM	12		5.00	26	8	33	249.16	

1	1/29/2019	4:14:12 PM	13	5.00	25	5	30	233.83
1	1/29/2019	4:20:30 PM	14	5.00	21	6	27	217.17
1	1/29/2019	4:26:43 PM	15	5.00	27	10	37	328.37
1	1/29/2019	4:32:57 PM	16	5.00	23	6	29	262.26
1	1/29/2019	4:38:37 PM	17	5.00	14462	35766	50228	1012.51


LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 1/29/19	Time 1424	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) <input checked="" type="checkbox"/> ON <input type="checkbox"/> OFF					

LSC Printout Date 1/29/2019

NOTE SWIPE AND SURVEY LOCATIONS ON MAP



Survey #	Location	CPM
1	Elevator Entrance	≤ 100
2	South Floor	≤ 100
3	East Floor	700
4	North Bench Top	≤ 100
5	West Floor	150
6	Work Platform	≤ 100
7	Calorimeter	≤ 100
8	Computers	≤ 100
9	DSC/TGA	≤ 100
10	South Bench Top	≤ 100
11	Background	
12	Elevator Entrance	South Bench Top
13	South Floor	
14	East Floor	
15	North Bench Top	
16	West Lab Floor	
17	Work Platform	
18	Calorimeter	
19	Computers	
20	DSC/TGA	

Surveyed By: AG/MJ	Date: 1/29/19
Swiped By: AG	Date: 1/29/19
Reviewed By: <i>[Signature]</i>	Date: 1/30/2019

Notes At location 3 (survey) there was 2-3x10⁴ High CPM due to radioactive material in safe

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 10/4/2018

Week #	6	Dates:	2/03/2019-2/09/2019	Counted On	2/5/2019
INSTRUMENT USED: Quantasart (Serial# 073396)			CONTROL FACTOR: 1.40E-06		
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		9.10E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	

Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	32	*****	*****	
2.	BACKGROUND 2	31	*****	*****	
3.	BACKGROUND 3	34	*****	*****	
4.	BACKGROUND 4	32	*****	*****	
5.	BACKGROUND 5	29	*****	*****	
6.	BACKGROUND 6	35	*****	*****	
7.	B021 ELEVATOR ENTRANCE	34	M	M	
8.	B021 FENCED GATE	36	M	M	
9.	B021 EAST FLOOR	37	M	M	
10.	B021 NORTH BENCH TOP	36	M	M	
11.	B021 WEST LAB FLOOR	35	M	M	
12.	B021 WORK PLATFORM	32	M	M	
13.	B021 CALORIMETER	26	M	M	
14.	B021 COMPUTERS	26	M	M	
15.	B021 DSC/TGA	29	M	M	
16.	B021 SOUTH BENCH TOP	38	M	M	

Notes: NONE

³⁶ CI SOURCE 09/14/1971	50106	*****	*****	
------------------------------------	-------	-------	-------	--

	101.03%	CALIBRATION CHECK OK?	YES
--	---------	-----------------------	-----

Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10⁻⁶ uCi/cm².

SURVEY PREPARED BY: <i>m. Heine</i>	REVIEWED BY: <i>c. Hines</i>
-------------------------------------	------------------------------

SIGN/DATE: <i>[Signature]</i> 2/5/19	SIGN/DATE: <i>[Signature]</i>
--------------------------------------	-------------------------------

Assay Definition-

Assay Description:

Assay Type: CPM
 Report Name: Report1
 Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190205_1441
 Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190205_1441\20190205_1441.results
 Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor
 Quench Indicator: tSIE/AEC
 External Std Terminator (sec): 0.5 2s%
 Pre-Count Delay (min): 0.00
 Quench Set: n/a
 Count Time (min): 5.00
 Count Mode: Normal
 Assay Count Cycles: 1 Repeat Sample Count: 1
 #Vials/Sample: 1 Calculate % Reference: Off

Background Subtract: Off
 Low CPM Threshold: Off
 2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On Luminescence Correction: Off
 Colored Samples: n/a Heterogeneity Monitor: n/a
 Coincidence Time (nsec): 18 Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	2/5/2019	2:42:18 PM	1	5.00		23	9	32	323.30	
1	2/5/2019	2:48:36 PM	2	5.00		26	5	31	223.82	
1	2/5/2019	2:54:52 PM	3	5.00		27	7	34	261.73	
1	2/5/2019	3:01:06 PM	4	5.00		25	6	32	269.14	
1	2/5/2019	3:07:15 PM	5	5.00		24	5	29	204.58	
1	2/5/2019	3:13:23 PM	6	5.00		28	7	35	279.37	
1	2/5/2019	3:19:39 PM	7	5.00		25	9	34	270.80	
1	2/5/2019	3:25:53 PM	8	5.00		28	8	36	273.19	
1	2/5/2019	3:32:11 PM	9	5.00		31	6	37	193.06	
1	2/5/2019	3:38:20 PM	10	5.00		29	7	36	183.74	
1	2/5/2019	3:44:38 PM	11	5.00		30	5	35	198.38	
1	2/5/2019	3:50:51 PM	12	5.00		25	7	32	217.27	

1	2/5/2019	3:57:13 PM	13	5.00	20	6	26	197.88
1	2/5/2019	4:03:22 PM	14	5.00	19	7	26	236.68
1	2/5/2019	4:09:37 PM	15	5.00	22	8	29	268.98
1	2/5/2019	4:15:50 PM	16	5.00	26	12	38	343.69
1	2/5/2019	4:21:29 PM	17	5.00	14445	35660	50106	1014.64


LABORATORY SURVEY

Radiation Safety Office
Washington State University
Pullman, WA 99164-1302
(509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 2/5/2019	Time 1412	Swipe Instrument Quantasmat	Radiation Detected Beta/Gamma
Reactor Status (Check One)				
ON <input checked="" type="checkbox"/> OFF <input type="checkbox"/>				

LSC Printout Date 2/5/2019

NOTE SWIPE AND SURVEY LOCATIONS ON MAP



Survey #	Location	CPM
1	Elevator Entrance	≤ 100
2	South Floor	≤ 100
3	East Floor	700
4	North Bench Top	≤ 100
5	West Floor	≤ 100
6	Work Platform	≤ 100
7	Calorimeter	≤ 100
8	Computers	≤ 100
9	DSC/TGA	≤ 100
10	South Bench Top	≤ 100
11	Background	
12	Elevator Entrance	South Bench Top
13	South Floor	
14	East Floor	
15	North Bench Top	
16	West Lab Floor	
17	Work Platform	
18	Calorimeter	
19	Computers	
20	DSC/TGA	

Notes 1) Survey #3 high CPM reading due to radioactive material safe

Surveyed By: AG	Date: 2/5/2019
Swiped By: AG	Date: 2/5/2019
Reviewed By: <i>[Signature]</i>	Date: 2/8/2019

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 10/4/2018

Week #	7	Dates:	2/10/2019-2/16/2019	Counted On	2/15/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.93E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	

Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	28	*****	*****	
2.	BACKGROUND 2	31	*****	*****	
3.	BACKGROUND 3	31	*****	*****	
4.	BACKGROUND 4	27	*****	*****	
5.	BACKGROUND 5	30	*****	*****	
6.	BACKGROUND 6	32	*****	*****	
7.	B021 ELEVATOR ENTRANCE	29	M	M	
8.	B021 FENCED GATE	30	M	M	
9.	B021 EAST FLOOR	29	M	M	
10.	B021 NORTH BENCH TOP	27	M	M	
11.	B021 WEST LAB FLOOR	26	M	M	
12.	B021 WORK PLATFORM	30	M	M	
13.	B021 CALORIMETER	30	M	M	
14.	B021 COMPUTERS	26	M	M	
15.	B021 DSC/TGA	28	M	M	
16.	B021 SOUTH BENCH TOP	43	1.84E-05	1.84E-07	

Notes: NONE

³⁶ Cl SOURCE 09/14/1971	50217	*****	*****	
	101.25%	CALIBRATION CHECK OK?		YES

Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10⁻⁶ uCi/cm².

SURVEY PREPARED BY: B.Tanner	REVIEWED BY: <i>C. Hines</i>
SIGN/DATE: <i>[Signature]</i> / 2/15/2019	SIGN/DATE: <i>[Signature]</i> 2/15/2019

Protocol# 1 - Reactor Weekly.lsa

User: CSE

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190215_0802

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190215_0802\20190215_0802.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results


P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	2/15/2019	8:03:08 AM	1	5.00		24	4	28	152.57	
1	2/15/2019	8:09:17 AM	2	5.00		24	7	31	250.75	
1	2/15/2019	8:15:32 AM	3	5.00		26	5	31	269.96	
1	2/15/2019	8:21:39 AM	4	5.00		23	5	27	260.38	
1	2/15/2019	8:27:54 AM	5	5.00		24	6	30	242.14	
1	2/15/2019	8:34:08 AM	6	5.00		27	5	32	233.22	
1	2/15/2019	8:40:27 AM	7	5.00		25	4	29	168.06	
1	2/15/2019	8:46:42 AM	8	5.00		23	7	30	245.71	
1	2/15/2019	8:53:00 AM	9	5.00		25	4	29	177.74	
1	2/15/2019	8:59:12 AM	10	5.00		22	5	27	258.67	
1	2/15/2019	9:05:26 AM	11	5.00		21	5	26	210.54	
1	2/15/2019	9:11:39 AM	12	5.00		24	7	30	336.70	

1	2/15/2019	9:17:56 AM	13	5.00	24	6	30	298.41
1	2/15/2019	9:24:12 AM	14	5.00	19	7	26	351.72
1	2/15/2019	9:30:26 AM	15	5.00	23	5	28	252.10
1	2/15/2019	9:36:40 AM	16	5.00	32	11	43	309.37
1	2/15/2019	9:42:20 AM	17	5.00	14317	35900	50217	1016.07

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 2/14/19	Time 1435	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input checked="" type="checkbox"/> OFF <input type="checkbox"/>		LSC Printout Date 2/15/2019			



Survey #	Location	CPM
1	Elevator Entrance	170
2	South Floor	180
3	East Floor	170 UMH 800
4	North Bench Top	160
5	West Floor	≤100
6	Work Platform	≤100
7	Calorimeter	≤100
8	Computers	≤100
9	DSC/TGA	350
10	South Bench Top	≤100
11	Background	
12	Elevator Entrance	
13	South Floor	
14	East Floor	
15	North Bench Top	
16	West Lab Floor	
17	Work Platform	
18	Calorimeter	
19	Computers	
20	DSC/TGA	

Surveyed By: CMH/AG/CS	Date: 2/14/19
Swiped By: CS/AG	Date: 2/14/19
Reviewed By: <i>[Signature]</i>	Date: 2/15/2019

Notes: Survey location 3 has high CPM due to radioactive material safe.

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 10/4/2018

Week #	8	Dates:	2/17/2019-2/23/2019	Counted On	2/19/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		9.59E-08 uCi/cm ²	QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK		

Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	40	*****	*****	
2.	BACKGROUND 2	30	*****	*****	
3.	BACKGROUND 3	36	*****	*****	
4.	BACKGROUND 4	33	*****	*****	
5.	BACKGROUND 5	77	*****	*****	
6.	BACKGROUND 6	30	*****	*****	
7.	B021 ELEVATOR ENTRANCE	30	M	M	
8.	B021 FENCED GATE	35	M	M	
9.	B021 EAST FLOOR	35	M	M	
10.	B021 NORTH BENCH TOP	30	M	M	
11.	B021 WEST LAB FLOOR	32	M	M	
12.	B021 WORK PLATFORM	30	M	M	
13.	B021 CALORIMETER	32	M	M	
14.	B021 COMPUTERS	29	M	M	
15.	B021 DSC/TGA	33	M	M	
16.	B021 SOUTH BENCH TOP	37	M	M	

Notes: NONE

³⁶ CI SOURCE 09/14/1971	50543	*****	*****		
	101.91%	CALIBRATION CHECK OK?		YES	

Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10⁻⁶ uCi/cm².

SURVEY PREPARED BY: M.Heine	REVIEWED BY: C. Heine
SIGN/DATE: <i>Maddison Heine</i> 2/19/19	SIGN/DATE: <i>amysain</i>

Assay Definition-

Assay Description:

Assay Type: CPM
 Report Name: Report1
 Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190219_1253
 Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190219_1253\20190219_1253.results
 Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor
 Quench Indicator: tSIE/AEC
 External Std Terminator (sec): 0.5 2s%
 Pre-Count Delay (min): 0.00
 Quench Set: n/a
 Count Time (min): 5.00
 Count Mode: Normal
 Assay Count Cycles: 1 Repeat Sample Count: 1
 #Vials/Sample: 1 Calculate % Reference: Off

Background Subtract: Off
 Low CPM Threshold: Off
 2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On Luminescence Correction: Off
 Colored Samples: n/a Heterogeneity Monitor: n/a
 Coincidence Time (nsec): 18 Delay Before Burst (nsec): 75

Half Life-

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	2/19/2019	12:54:26 PM	1	5.00		34	7	40	213.54	
1	2/19/2019	1:00:35 PM	2	5.00		24	6	30	271.18	
1	2/19/2019	1:06:50 PM	3	5.00		31	5	36	278.18	
1	2/19/2019	1:12:58 PM	4	5.00		27	6	33	275.24	
1	2/19/2019	1:19:14 PM	5	5.00		68	9	77	141.89	
1	2/19/2019	1:25:30 PM	6	5.00		25	5	30	248.58	
1	2/19/2019	1:31:48 PM	7	5.00		24	6	30	226.22	
1	2/19/2019	1:38:03 PM	8	5.00		26	10	35	270.38	
1	2/19/2019	1:44:23 PM	9	5.00		26	9	35	243.24	
1	2/19/2019	1:50:37 PM	10	5.00		22	7	30	245.70	
1	2/19/2019	1:56:55 PM	11	5.00		26	7	32	180.99	
1	2/19/2019	2:03:08 PM	12	5.00		24	6	30	272.32	

1	2/19/2019	2:09:29 PM	13	5.00	25	7	32	215.85
1	2/19/2019	2:15:48 PM	14	5.00	25	4	29	170.33
1	2/19/2019	2:22:04 PM	15	5.00	26	7	33	194.74
1	2/19/2019	2:28:26 PM	16	5.00	30	7	37	217.92
1	2/19/2019	2:34:06 PM	17	5.00	14641	35902	50543	1010.60


LABORATORY SURVEY

Radiation Safety Office
Washington State University
Pullman, WA 99164-1302
(509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 2/19/19	Time 12:18	Swipe Instrument Quantasmart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input checked="" type="checkbox"/> OFF <input type="checkbox"/>					

LSC Printout Date **2/19/2019**

NOTE SWIPE AND SURVEY LOCATIONS ON MAP



Survey #	Location	CPM
1	Elevator Entrance	≤ 100
2	South Floor	≤ 100
3	East Lab Floor	700
4	North Bench Top	≤ 100
5	West Floor	≤ 100
6	Work Platform	≤ 100
7	Calorimeter	≤ 100
8	Computers	≤ 100
9	DSC/TGA	≤ 100
10	South Bench Top	≤ 100
11	Background	
12	Elevator Entrance	
13	South Floor	
14	East Lab Floor	
15	North Bench Top	
16	West Lab Floor	
17	Work Platform	
18	Calorimeter	
19	Computers	
20	DSC/TGA	

Surveyed By: AG/JS Date: 2/19/19

Swiped By: JS/MS Date: 2/19/19



Reviewed By: [Signature] Date: 2/19/2019

Notes: Survey location #3 had high CPM because radioactive materials safe

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	9	Dates:	2/24/2019-3/2/2019	Counted On	2/26/2019
INSTRUMENT USED: Quantasmart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.		uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100	
Limit of quantification (LOQ)=		8.93E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	33	*****	*****	
2.	BACKGROUND 2	25	*****	*****	
3.	BACKGROUND 3	27	*****	*****	
4.	BACKGROUND 4	31	*****	*****	
5.	BACKGROUND 5	30	*****	*****	
6.	BACKGROUND 6	33	*****	*****	
7.	B021 ELEVATOR ENTRANCE	31	M	M	
8.	B021 FENCED GATE	28	M	M	
9.	B021 EAST FLOOR	26	M	M	
10.	B021 NORTH BENCH TOP	29	M	M	
11.	B021 WEST LAB FLOOR	26	M	M	
12.	B021 WORK PLATFORM	26	M	M	
13.	B021 CALORIMETER	32	M	M	
14.	B021 COMPUTERS	31	M	M	
15.	B021 DSC/TGA	36	M	M	
16.	B021 SOUTH BENCH TOP	28	M	M	
17.	121 DOORWAY	33	M	M	
18.	121 GLOVE BOX- FLOOR	29	M	M	
19.	121 GLOVE BOX	25	M	M	
20.	121 WEST BENCH TOP	25	M	M	
21.	121 CENTER OF LAB	32	M	M	
22.	121 EAST BENCH TOP	26	M	M	
Notes: NONE					
³⁶ Cl SOURCE 09/14/1971		50124		*****	*****
		101.07%		CALIBRATION CHECK OK? YES	
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B.Tanner			REVIEWED BY: C. Hines		
SIGN/DATE:  3/1/2019			SIGN/DATE:  3/1/2019		

Assay Definition-

Assay Description:

Assay Type: CPM
Report Name: Report1
Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190226_1454
Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190226_1454\20190226_1454.results
Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor
Quench Indicator: tSIE/AEC
External Std Terminator (sec): 0.5 2s%
Pre-Count Delay (min): 0.00
Quench Set: n/a
Count Time (min): 5.00
Count Mode: Normal
Assay Count Cycles: 1 Repeat Sample Count: 1
#Vials/Sample: 1 Calculate % Reference: Off

Background Subtract: Off
Low CPM Threshold: Off
2 Sigma % Terminator: Off

Table with 3 columns: Regions, LL, UL. Rows A, B, C.

Count Corrections-

Static Controller: On Luminescence Correction: Off
Colored Samples: n/a Heterogeneity Monitor: n/a
Coincidence Time (nsec): 18 Delay Before Burst (nsec): 75

Half Life-

Table with 5 columns: Half Life Correction, Regions, Half Life, Units, Reference Date, Reference Time.

Cycle 1 Results

Table with 10 columns: P#, DATE, TIME, S#, Count, Time, CPMA, CPMB, CPMC, SIS, MESSAGES. Contains 12 rows of data.

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	2/26/2019	4:10:32 PM	13	5.00	25	6	32	161.88
1	2/26/2019	4:16:47 PM	14	5.00	25	6	31	203.51
1	2/26/2019	4:23:00 PM	15	5.00	28	8	36	308.14
1	2/26/2019	4:29:09 PM	16	5.00	24	4	28	227.71
1	2/26/2019	4:35:22 PM	17	5.00	28	4	33	185.70
1	2/26/2019	4:41:32 PM	18	5.00	24	4	29	165.81
1	2/26/2019	4:47:48 PM	19	5.00	21	4	25	190.55
1	2/26/2019	4:54:01 PM	20	5.00	20	5	25	289.96
1	2/26/2019	5:03:33 PM	21	5.00	26	6	32	45.02
1	2/26/2019	5:09:46 PM	22	5.00	20	7	26	312.83
1	2/26/2019	5:15:26 PM	23	5.00	14342	35782	50124	1009.65

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916


Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 2/26/19	Time 1221	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input checked="" type="checkbox"/> OFF <input type="checkbox"/>		LSC Printout Date 2/26/2019			

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP		
11	121 Doorway	5100	<p>DODG 121 GUO LAB</p> <p>East Bench Top</p> <p>Fume Hood</p> <p>West Bench Top</p> <p>Glove Box</p>		
12	121 Center of lab	5100			
13	121 Glove Box	600			
Swipe # and Locations			Notes Survey # 13 highest reading came from inside left rubber arm		
17	121 Doorway		Surveyed By: AG	Date: 2/26/19	
18	121 Glove Box-floor		Swiped By: MS	Date: 2/26/19	
19	121 Glove Box		Reviewed By: [Signature]	Date: 3/1/2019	
20	West Bench Top				
21	121 Center of lab				
22	East Bench Top				

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	10	Dates:	3/3/2019-3/9/2019	Counted On	3/8/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.94E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	23	*****	*****	
2.	BACKGROUND 2	31	*****	*****	
3.	BACKGROUND 3	33	*****	*****	
4.	BACKGROUND 4	27	*****	*****	
5.	BACKGROUND 5	33	*****	*****	
6.	BACKGROUND 6	33	*****	*****	
7.	B021 ELEVATOR ENTRANCE	29	M	M	
8.	B021 FENCED GATE	31	M	M	
9.	B021 EAST FLOOR	30	M	M	
10.	B021 NORTH BENCH TOP	27	M	M	
11.	B021 WEST LAB FLOOR	32	M	M	
12.	B021 WORK PLATFORM	32	M	M	
13.	B021 CALORIMETER	26	M	M	
14.	B021 COMPUTERS	25	M	M	
15.	B021 DSC/TGA	34	M	M	
16.	B021 SOUTH BENCH TOP	49	2.66E-05	2.66E-07	
17.	121 DOORWAY	28	M	M	
18.	121 GLOVE BOX- FLOOR	32	M	M	
19.	121 GLOVE BOX	27	M	M	
20.	121 WEST BENCH TOP	30	M	M	
21.	121 CENTER OF LAB	26	M	M	
22.	121 EAST BENCH TOP	31	M	M	
Notes: NONE					
³⁶ CI SOURCE 09/14/1971		50342		*****	*****
		101.50%		CALIBRATION CHECK OK? YES	
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B.Tanner			REVIEWED BY: C. Hines		
SIGN/DATE:  3/8/2019			SIGN/DATE:  3/13/2019		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190308_0803

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190308_0803\20190308_0803.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	3/8/2019	8:04:16 AM	1	5.00		18	6	23	368.44	
1	3/8/2019	8:10:30 AM	2	5.00		26	4	31	195.25	
1	3/8/2019	8:16:39 AM	3	5.00		27	6	33	231.52	
1	3/8/2019	8:22:53 AM	4	5.00		22	5	27	206.58	
1	3/8/2019	8:29:01 AM	5	5.00		27	6	33	257.59	
1	3/8/2019	8:35:13 AM	6	5.00		26	7	33	236.36	
1	3/8/2019	8:41:28 AM	7	5.00		24	5	29	201.41	
1	3/8/2019	8:47:41 AM	8	5.00		24	8	31	251.22	
1	3/8/2019	8:53:56 AM	9	5.00		22	8	30	291.69	
1	3/8/2019	9:00:08 AM	10	5.00		22	5	27	180.32	
1	3/8/2019	9:06:20 AM	11	5.00		25	7	32	256.65	
1	3/8/2019	9:12:36 AM	12	5.00		25	7	32	214.81	

Protocol# 1 - Reactor Weekly.lsa


User: CSE

1	3/8/2019	9:18:53 AM	13	5.00	22	4	26	175.22
1	3/8/2019	9:25:10 AM	14	5.00	20	5	25	281.64
1	3/8/2019	9:31:23 AM	15	5.00	24	9	34	354.75
1	3/8/2019	9:37:31 AM	16	5.00	33	16	49	331.36
1	3/8/2019	9:43:44 AM	17	5.00	22	7	28	317.36
1	3/8/2019	9:49:52 AM	18	5.00	27	5	32	192.02
1	3/8/2019	9:56:10 AM	19	5.00	23	4	27	227.39
1	3/8/2019	10:02:23 AM	20	5.00	25	5	30	205.42
1	3/8/2019	10:08:39 AM	21	5.00	22	4	26	231.62
1	3/8/2019	10:14:54 AM	22	5.00	25	6	31	198.38
1	3/8/2019	10:20:34 AM	23	5.00	14544	35798	50342	1016.67

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 3/5/19	Time 1319	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) <input checked="" type="checkbox"/> ON <input type="checkbox"/> OFF <input type="checkbox"/>		LSC Printout Date 3/8/2019			



Survey #	Location	CPM
1	Elevator Entrance	≤ 100
2	South Floor	≤ 100
3	East Lab Floor	900
4	North Bench Top	≤ 100
5	West Floor	≤ 100
6	Work Platform	≤ 100
7	Calorimeter	≤ 100
8	Computers	≤ 100
9	DSC/TGA	≤ 100
10	South Bench Top	≤ 100
11	Background	
12	Work Platform	
13	Calorimeter	
14	Computers	
15	DSC/TGA	
16	South Bench Top	

Surveys By: <i>MD</i>		Date: 3/5/19	Notes
Swiped By: <i>AG</i>		Date: 3/5/19	
Reviewed By: <i>Carroll</i>		Date: 3/8/2019	

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.


LABORATORY SURVEY

Radiation Safety Office
Washington State University
Pullman, WA 99164-1302
(509) 335 8916

Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 3/5/19	Time 1314	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input checked="" type="checkbox"/> OFF <input type="checkbox"/>					

LSC Printout Date 3/8/2019

NOTE SWIPE AND SURVEY LOCATIONS ON MAP



Swipe # and Locations		Notes	
17	121 Doorway	Surveyed By: <u>AG</u>	Date: <u>3/5/19</u>
18	121 Glove Box-floor	Swiped By: <u>MD</u>	Date: <u>3/5/19</u>
19	121 Glove Box	Reviewed By: <u>[Signature]</u>	Date: <u>3/13/2019</u>
20	West Bench Top		
21	121 Center of lab		
22	East Bench Top		

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	11	Dates:	3/10/2019-3/16/2019	Counted On	3/12/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.96E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	29	*****	*****	
2.	BACKGROUND 2	30	*****	*****	
3.	BACKGROUND 3	33	*****	*****	
4.	BACKGROUND 4	29	*****	*****	
5.	BACKGROUND 5	29	*****	*****	
6.	BACKGROUND 6	31	*****	*****	
7.	B021 ELEVATOR ENTRANCE	30	M	M	
8.	B021 FENCED GATE	29	M	M	
9.	B021 EAST FLOOR	28	M	M	
10.	B021 NORTH BENCH TOP	29	M	M	
11.	B021 WEST LAB FLOOR	28	M	M	
12.	B021 WORK PLATFORM	30	M	M	
13.	B021 CALORIMETER	29	M	M	
14.	B021 COMPUTERS	27	M	M	
15.	B021 DSC/TGA	31	M	M	
16.	B021 SOUTH BENCH TOP	29	M	M	
17.	121 DOORWAY	32	M	M	
18.	121 GLOVE BOX- FLOOR	29	M	M	
19.	121 GLOVE BOX	30	M	M	
20.	121 WEST BENCH TOP	29	M	M	
21.	121 CENTER OF LAB	30	M	M	
22.	121 EAST BENCH TOP	31	M	M	
Notes: NONE					
³⁶ CI SOURCE 09/14/1971		50051		*****	*****
		100.92%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: M.Heine			REVIEWED BY: C. Hines		
SIGN/DATE: <i>Mandelson/Heine</i> / 3/13/19			SIGN/DATE: <i>auger</i> — 3/13/2019		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190312_0954

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190312_0954\20190312_0954.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	3/12/2019	9:55:17 AM	1	5.00	23	6	29	330.19		
1	3/12/2019	10:01:34 AM	2	5.00	23	7	30	304.28		
1	3/12/2019	10:07:47 AM	3	5.00	28	5	33	256.80		
1	3/12/2019	10:14:04 AM	4	5.00	22	6	29	273.18		
1	3/12/2019	10:20:18 AM	5	5.00	22	7	29	303.33		
1	3/12/2019	10:26:34 AM	6	5.00	25	6	31	292.25		
1	3/12/2019	10:32:50 AM	7	5.00	25	5	30	250.13		
1	3/12/2019	10:39:02 AM	8	5.00	23	5	29	232.00		
1	3/12/2019	10:45:21 AM	9	5.00	24	4	28	169.46		
1	3/12/2019	10:51:34 AM	10	5.00	24	5	29	220.72		
1	3/12/2019	10:57:50 AM	11	5.00	23	5	28	240.02		
1	3/12/2019	11:04:07 AM	12	5.00	25	5	30	222.67		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	3/12/2019	11:10:26 AM	13	5.00	23	6	29	299.14
1	3/12/2019	11:16:44 AM	14	5.00	22	6	27	251.41
1	3/12/2019	11:23:00 AM	15	5.00	24	7	31	265.35
1	3/12/2019	11:29:10 AM	16	5.00	25	4	29	199.04
1	3/12/2019	11:35:23 AM	17	5.00	26	7	32	269.56
1	3/12/2019	11:41:34 AM	18	5.00	23	6	29	242.99
1	3/12/2019	11:47:55 AM	19	5.00	25	5	30	222.68
1	3/12/2019	11:54:10 AM	20	5.00	23	7	29	265.86
1	3/12/2019	12:00:26 PM	21	5.00	26	4	30	191.11
1	3/12/2019	12:06:44 PM	22	5.00	25	6	31	235.85
1	3/12/2019	12:12:24 PM	23	5.00	14495	35556	50051	1016.62

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 3/12/19	Time 1026	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input checked="" type="checkbox"/> OFF <input type="checkbox"/>					

LSC Printout Date **3/12/2019**

NOTE SWIPE AND SURVEY LOCATIONS ON MAP

DODG B0021 GUO LAB

Survey #	Location	CPM
1	Elevator Entrance	<100
2	South Floor	<100
3	East Lab Floor	600
4	North Bench Top	<100
5	West Floor	<100
6	Work Platform	<100
7	Calorimeter	<100
8	Computers	<100
9	DSC/TGA	<100
10	South Bench Top	<100
11	Background	
12	Elevator Entrance	South Bench Top
13	South Floor	
14	East Lab Floor	
15	North Bench Top	
16	West Lab Floor	
17	Work Platform	
18	Calorimeter	
19	Computers	
20	DSC/TGA	

Notes
 Survey location # 3 had high CPM because RAM safe.

Surveyed By: LD Date: 3/12/19

Swiped By: MJD Date: 3/12/19

Reviewed By: [Signature] Date: 3/12/19

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	12	Dates:	3/17/2019-3/23/2019	Counted On	3/21/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.75E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	26	*****	*****	
2.	BACKGROUND 2	30	*****	*****	
3.	BACKGROUND 3	30	*****	*****	
4.	BACKGROUND 4	26	*****	*****	
5.	BACKGROUND 5	28	*****	*****	
6.	BACKGROUND 6	25	*****	*****	
7.	B021 ELEVATOR ENTRANCE	24	M	M	
8.	B021 FENCED GATE	28	M	M	
9.	B021 EAST FLOOR	26	M	M	
10.	B021 NORTH BENCH TOP	25	M	M	
11.	B021 WEST LAB FLOOR	28	M	M	
12.	B021 WORK PLATFORM	29	M	M	
13.	B021 CALORIMETER	29	M	M	
14.	B021 COMPUTERS	27	M	M	
15.	B021 DSC/TGA	30	M	M	
16.	B021 SOUTH BENCH TOP	43	2.17E-05	2.17E-07	
17.	121 DOORWAY	31	M	M	
18.	121 GLOVE BOX- FLOOR	23	M	M	
19.	121 GLOVE BOX	20	M	M	
20.	121 WEST BENCH TOP	29	M	M	
21.	121 CENTER OF LAB	28	M	M	
22.	121 EAST BENCH TOP	26	M	M	
Notes: NONE					
³⁶ CI SOURCE 09/14/1971		50202	*****	*****	
		101.22%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: <i>C. Hines</i>		
SIGN/DATE: <i>B. Tanner</i> / 3/22/19			SIGN/DATE: <i>amjgum</i> 3/22/2019		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190321_0713

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190321_0713\20190321_0713.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	3/21/2019	7:14:09 AM	1	5.00	20	6	26	309.63		
1	3/21/2019	7:20:23 AM	2	5.00	24	7	30	320.06		
1	3/21/2019	7:26:34 AM	3	5.00	24	6	30	221.81		
1	3/21/2019	7:32:45 AM	4	5.00	21	5	26	216.45		
1	3/21/2019	7:38:56 AM	5	5.00	24	5	28	205.90		
1	3/21/2019	7:45:05 AM	6	5.00	20	5	25	247.61		
1	3/21/2019	7:51:19 AM	7	5.00	18	6	24	214.71		
1	3/21/2019	7:57:31 AM	8	5.00	24	4	28	230.45		
1	3/21/2019	8:03:47 AM	9	5.00	21	6	26	254.66		
1	3/21/2019	8:09:57 AM	10	5.00	21	4	25	289.45		
1	3/21/2019	8:16:08 AM	11	5.00	22	6	28	270.83		
1	3/21/2019	8:22:23 AM	12	5.00	27	3	29	175.52		

Protocol# 1 - Reactor Weekly.lsa

User: CSE


1	3/21/2019	8:28:39 AM	13	5.00	23	5	29	229.36
1	3/21/2019	8:34:52 AM	14	5.00	23	5	27	204.64
1	3/21/2019	8:41:04 AM	15	5.00	24	5	30	248.21
1	3/21/2019	8:47:19 AM	16	5.00	31	12	43	243.86
1	3/21/2019	8:53:30 AM	17	5.00	25	6	31	273.79
1	3/21/2019	8:59:43 AM	18	5.00	20	3	23	217.80
1	3/21/2019	9:05:59 AM	19	5.00	15	5	20	400.36
1	3/21/2019	9:12:11 AM	20	5.00	23	6	29	238.81
1	3/21/2019	9:18:23 AM	21	5.00	22	6	28	264.76
1	3/21/2019	9:24:37 AM	22	5.00	21	5	26	278.33
1	3/21/2019	9:30:17 AM	23	5.00	14445	35757	50202	1012.54

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 85010	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 3/19/19	Time 12:14	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input checked="" type="checkbox"/> OFF <input type="checkbox"/>					

LSC Printout Date _____



Survey #	Location	CPM	Notes
1-6	Background	16	South Bench Top
7	Elevator Entrance		
8	South Floor		
9	East Lab Floor		
10	North Bench Top		
11	West Lab Floor		
12	Work Platform		
13	Calorimeter		
14	Computers		
15	DSC/TGA		

Surveyed By: AG Date: 3/19/19



Swiped By: MJ Date: 3/19/19

Reviewed By: AMJ Date: 3/22/19

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	13	Dates:	3/24/2019-3/30/2019	Counted On	4/1/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.		uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100	
Limit of quantification (LOQ)=		8.90E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	27	*****	*****	
2.	BACKGROUND 2	29	*****	*****	
3.	BACKGROUND 3	30	*****	*****	
4.	BACKGROUND 4	32	*****	*****	
5.	BACKGROUND 5	33	*****	*****	
6.	BACKGROUND 6	26	*****	*****	
7.	B021 ELEVATOR ENTRANCE	28	M	M	
8.	B021 FENCED GATE	30	M	M	
9.	B021 EAST FLOOR	31	M	M	
10.	B021 NORTH BENCH TOP	29	M	M	
11.	B021 WEST LAB FLOOR	25	M	M	
12.	B021 WORK PLATFORM	28	M	M	
13.	B021 CALORIMETER	Note 1	#VALUE!	#VALUE!	###
14.	B021 COMPUTERS	29	M	M	
15.	B021 DSC/TGA	27	M	M	
16.	B021 SOUTH BENCH TOP	26	M	M	
17.	121 DOORWAY	34	M	M	
18.	121 GLOVE BOX- FLOOR	25	M	M	
19.	121 GLOVE BOX	31	M	M	
20.	121 WEST BENCH TOP	31	M	M	
21.	121 CENTER OF LAB	25	M	M	
22.	121 EAST BENCH TOP	26	M	M	
Note 1: Calorimeter swipe not taken due to experiment in progress.					
³⁶ CI SOURCE 09/14/1971		50060	*****	*****	
		100.94%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: C. Hines		
SIGN/DATE:  4/1/2019			SIGN/DATE:  4/1/2019		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190401_0713

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190401_0713\20190401_0713.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	4/1/2019	7:14:33 AM	1	5.00	20	7	27	262.90	
1	4/1/2019	7:20:47 AM	2	5.00	24	5	29	253.86	
1	4/1/2019	7:27:01 AM	3	5.00	25	5	30	246.32	
1	4/1/2019	7:33:16 AM	4	5.00	26	5	32	254.95	
1	4/1/2019	7:39:22 AM	5	5.00	27	6	33	190.65	
1	4/1/2019	7:45:34 AM	6	5.00	22	5	26	245.82	
1	4/1/2019	7:51:49 AM	7	5.00	22	6	28	227.85	
1	4/1/2019	7:58:02 AM	8	5.00	23	6	30	279.24	
1	4/1/2019	8:04:18 AM	9	5.00	25	6	31	193.00	
1	4/1/2019	8:10:25 AM	10	5.00	22	7	29	356.29	
1	4/1/2019	8:16:38 AM	11	5.00	19	6	25	230.52	
1	4/1/2019	8:22:51 AM	12	5.00	23	5	28	231.90	


Missing vial 13.

1	4/1/2019	8:29:09 AM	14	5.00	21	6	27	372.23
1	4/1/2019	8:35:17 AM	15	5.00	25	4	29	188.10
1	4/1/2019	8:41:28 AM	16	5.00	21	7	27	320.92
1	4/1/2019	8:47:42 AM	17	5.00	20	6	26	276.52
1	4/1/2019	8:53:55 AM	18	5.00	27	7	34	330.61
1	4/1/2019	9:00:10 AM	19	5.00	20	5	25	237.82
1	4/1/2019	9:06:26 AM	20	5.00	25	6	31	263.45
1	4/1/2019	9:12:31 AM	21	5.00	25	7	31	251.65
1	4/1/2019	9:18:39 AM	22	5.00	19	6	25	377.60
1	4/1/2019	9:24:58 AM	23	5.00	20	6	26	320.66
1	4/1/2019	9:30:38 AM	24	5.00	14513	35547	50060	1016.14

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 3/29/2019	Time 1539	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One)		LSC Printout Date 4/1/2019			
ON <input type="checkbox"/>	OFF <input checked="" type="checkbox"/>				





Surveyed By: BT	Date: 3/29/19	Notes NOTE 1: EXPERIMENT IN PROGRESS, SWAPE AND SURVEY NOT TAKEN.
Swiped By: CMH	Date: 3/29/19	
Reviewed By: <i>[Signature]</i>	Date: 4/1/19	

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	14	Dates:	3/31/2019-4/6/2019	Counted On	4/8/2019
INSTRUMENT USED: Quantasart (Serial# 073396)			CONTROL FACTOR: 1.40E-06		
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		9.02E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	32	*****	*****	
2.	BACKGROUND 2	32	*****	*****	
3.	BACKGROUND 3	31	*****	*****	
4.	BACKGROUND 4	28	*****	*****	
5.	BACKGROUND 5	32	*****	*****	
6.	BACKGROUND 6	31	*****	*****	
7.	B021 ELEVATOR ENTRANCE	24	M	M	
8.	B021 FENCED GATE	27	M	M	
9.	B021 EAST FLOOR	29	M	M	
10.	B021 NORTH BENCH TOP	34	M	M	
11.	B021 WEST LAB FLOOR	38	9.80E-06	9.80E-08	
12.	B021 WORK PLATFORM	27	M	M	
13.	B021 CALORIMETER	28	M	M	
14.	B021 COMPUTERS	33	M	M	
15.	B021 DSC/TGA	27	M	M	
16.	B021 SOUTH BENCH TOP	29	M	M	
17.	121 DOORWAY	52	2.94E-05	2.94E-07	
18.	121 GLOVE BOX- FLOOR	28	M	M	
19.	121 GLOVE BOX	26	M	M	
20.	121 WEST BENCH TOP	25	M	M	
21.	121 CENTER OF LAB	21	M	M	
22.	121 EAST BENCH TOP	24	M	M	
None					
³⁶ CI SOURCE 09/14/1971		50157	*****	*****	
		101.13%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: C. Hines		
SIGN/DATE:  4/9/2019			SIGN/DATE:  4/8/2019		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190408_0912

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190408_0912\20190408_0912.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	4/8/2019	9:13:34 AM	1	5.00	26	6	32	276.94		
1	4/8/2019	9:19:47 AM	2	5.00	26	6	32	282.54		
1	4/8/2019	9:25:59 AM	3	5.00	25	6	31	276.79		
1	4/8/2019	9:32:15 AM	4	5.00	23	5	28	313.26		
1	4/8/2019	9:38:27 AM	5	5.00	25	7	32	270.67		
1	4/8/2019	9:44:42 AM	6	5.00	25	7	31	277.28		
1	4/8/2019	9:50:59 AM	7	5.00	20	3	24	194.78		
1	4/8/2019	9:57:15 AM	8	5.00	22	5	27	265.48		
1	4/8/2019	10:03:29 AM	9	5.00	21	8	29	323.41		
1	4/8/2019	10:09:45 AM	10	5.00	26	8	34	281.87		
1	4/8/2019	10:16:03 AM	11	5.00	30	8	38	265.15		
1	4/8/2019	10:22:20 AM	12	5.00	22	5	27	208.25		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	4/8/2019	10:28:38 AM	13	5.00	23	5	28	217.94
1	4/8/2019	10:34:54 AM	14	5.00	28	5	33	221.88
1	4/8/2019	10:41:13 AM	15	5.00	21	6	27	296.24
1	4/8/2019	10:47:26 AM	16	5.00	22	7	29	301.29
1	4/8/2019	10:53:42 AM	17	5.00	36	16	52	266.68
1	4/8/2019	10:59:51 AM	18	5.00	20	7	28	351.96
1	4/8/2019	11:06:07 AM	19	5.00	22	4	26	276.50
1	4/8/2019	11:12:16 AM	20	5.00	20	4	25	241.63
1	4/8/2019	11:18:29 AM	21	5.00	17	4	21	202.06
1	4/8/2019	11:24:47 AM	22	5.00	19	5	24	277.41
1	4/8/2019	11:30:28 AM	23	5.00	14616	35541	50157	1015.78

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodger Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 4/5/19	Time 1543	Swipe Instrument Quantasmart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input checked="" type="checkbox"/> OFF <input type="checkbox"/>		LSC Printout Date 4/9/2019			

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP		
1	Elevator Entrance	60	<p>DODG B0021 GUO LAB</p>		
2	South Floor	70			
3	East Lab Floor	220			
4	North Bench Top	100			
5	West Floor	80			
6	Work Platform	70			
7	Calorimeter	60			
8	Computers	100			
9	DSC/TGA	80			
10	South Bench Top	60			
Swipe # and Locations			Surveyed By: <i>BT</i> Date: 4/5/19 Swiped By: <i>BT</i> Date: 4/5/19 Reviewed By: <i>[Signature]</i> Date: 4/8/2019		
1-6	Background	16	Notes		
7	Elevator Entrance				
8	South Floor				
9	East Lab Floor				
10	North Bench Top				
11	West Lab Floor				
12	Work Platform				
13	Calorimeter				
14	Computers				
15	DSC/TGA				

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building	Dodgen Research Facility	Room	121-GUO	Lab Class	D	Survey Instrument	GM Detector Model 3	Serial No.	58680	Radiation Detected	Gamma
Authorized User	Xiaofeng Guo	Date	4/5/19	Time	1555	Swipe Instrument	Quantasmat	Serial No.	073396	Radiation Detected	Beta/Gamma
Reactor Status (Check One)	<input checked="" type="checkbox"/> ON <input type="checkbox"/> OFF										

NOTE SWIPE AND SURVEY LOCATIONS ON MAP



LSC Printout Date 4/8/2019

Survey #	Location	CPM	DODG 121 GUO LAB			
11	121 Doorway	50				
12	121 Center of lab	80				
13	121 Glove Box	140				
Swipe # and Locations						
17	121 Doorway		Notes			
18	121 Glove Box-floor					
19	121 Glove Box					
20	West Bench Top					
21	121 Center of lab					
22	East Bench Top					

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	15	Dates:	4/7/2019-4/13/2019	Counted On	4/9/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.		uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100	
Limit of quantification (LOQ)=		8.90E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	33	*****	*****	
2.	BACKGROUND 2	28	*****	*****	
3.	BACKGROUND 3	30	*****	*****	
4.	BACKGROUND 4	30	*****	*****	
5.	BACKGROUND 5	26	*****	*****	
6.	BACKGROUND 6	30	*****	*****	
7.	B021 ELEVATOR ENTRANCE	29	M	M	
8.	B021 FENCED GATE	29	M	M	
9.	B021 EAST FLOOR	33	M	M	
10.	B021 NORTH BENCH TOP	25	M	M	
11.	B021 WEST LAB FLOOR	27	M	M	
12.	B021 WORK PLATFORM	28	M	M	
13.	B021 CALORIMETER	29	M	M	
14.	B021 COMPUTERS	29	M	M	
15.	B021 DSC/TGA	33	M	M	
16.	B021 SOUTH BENCH TOP	32	M	M	
17.	121 DOORWAY	27	M	M	
18.	121 GLOVE BOX- FLOOR	28	M	M	
19.	121 GLOVE BOX	26	M	M	
20.	121 WEST BENCH TOP	29	M	M	
21.	121 CENTER OF LAB	32	M	M	
22.	121 EAST BENCH TOP	28	M	M	
None					
³⁶ CI SOURCE 09/14/1971		50075		*****	*****
		100.97%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: C. Hines		
SIGN/DATE:  / 4/9/2019			SIGN/DATE: 		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190409_1317

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190409_1317\20190409_1317.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	4/9/2019	1:18:44 PM	1	5.00	27	6	33	225.79		
1	4/9/2019	1:24:58 PM	2	5.00	22	6	28	240.21		
1	4/9/2019	1:31:13 PM	3	5.00	25	5	30	213.34		
1	4/9/2019	1:37:30 PM	4	5.00	23	6	30	272.86		
1	4/9/2019	1:43:44 PM	5	5.00	21	5	26	279.87		
1	4/9/2019	1:49:57 PM	6	5.00	26	4	30	167.75		
1	4/9/2019	1:56:13 PM	7	5.00	23	6	29	284.52		
1	4/9/2019	2:02:29 PM	8	5.00	23	5	29	202.13		
1	4/9/2019	2:08:44 PM	9	5.00	27	6	33	166.62		
1	4/9/2019	2:15:02 PM	10	5.00	19	6	25	263.79		
1	4/9/2019	2:21:20 PM	11	5.00	23	4	27	184.52		
1	4/9/2019	2:27:38 PM	12	5.00	24	5	28	206.94		

1	4/9/2019	2:33:57 PM	13	5.00	23	5	29	196.10
1	4/9/2019	2:40:14 PM	14	5.00	25	5	29	187.86
1	4/9/2019	2:46:30 PM	15	5.00	25	8	33	211.82
1	4/9/2019	2:52:43 PM	16	5.00	25	7	32	136.42
1	4/9/2019	2:58:57 PM	17	5.00	23	5	27	174.17
1	4/9/2019	3:05:07 PM	18	5.00	22	5	28	189.38
1	4/9/2019	3:11:24 PM	19	5.00	21	5	26	209.19
1	4/9/2019	3:17:34 PM	20	5.00	25	4	29	175.62
1	4/9/2019	3:23:48 PM	21	5.00	27	5	32	147.51
1	4/9/2019	3:30:07 PM	22	5.00	21	7	28	291.87
1	4/9/2019	3:35:47 PM	23	5.00	14561	35514	50075	1013.22

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 4/9/19	Time 1237	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>	LSC Printout Date 4/9/2019				

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP		
1	Elevator Entrance	<100	<p>DODG B0021 GUO LAB</p>		
2	South Floor	<100			
3	East Lab Floor	900			
4	North Bench Top	150			
5	West Floor	<100			
6	Work Platform	<100			
7	Calorimeter	<100			
8	Computers	<100			
9	DSC/TGA	<100			
10	South Bench Top	<100			
Swipe # and Locations			Surveyed By: AG Date: 4/9/19 Swiped By: AL/BT Date: 4/9/19 Reviewed By: [Signature] Date: 4/10/2019		
1-6	Background	16	Notes		
7	Elevator Entrance				
8	South Floor				
9	East Lab Floor				
10	North Bench Top				
11	West Lab Floor				
12	Work Platform				
13	Calorimeter				
14	Computers				
15	DSC/TGA				

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room 121 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 4/9/19	Time 1221	Swipe Instrument Quantasart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>					

LSC Printout Date 4/9/2019

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP		
11	121 Doorway	5100			
12	121 Center of lab	5100			
13	121 Glove Box	600			
Swipe # and Locations					
17	121 Doorway		West Bench Top		
18	121 Glove Box-floor				
19	121 Glove Box				
20	West Bench Top				
21	121 Center of lab		Surveyed By: AG Date: 4/9/19		
22	East Bench Top		Swiped By: AG Date: 4/9/19		
			Reviewed By: [Signature] Date: 4/10/2019		
			Notes		

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	16	Dates:	4/14/2019-4/20/2019	Counted On	4/16/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.		uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100	
Limit of quantification (LOQ)=		.8.75E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	27	*****	*****	
2.	BACKGROUND 2	27	*****	*****	
3.	BACKGROUND 3	26	*****	*****	
4.	BACKGROUND 4	26	*****	*****	
5.	BACKGROUND 5	28	*****	*****	
6.	BACKGROUND 6	31	*****	*****	
7.	B021 ELEVATOR ENTRANCE	32	M	M	
8.	B021 FENCED GATE	31	M	M	
9.	B021 EAST FLOOR	27	M	M	
10.	B021 NORTH BENCH TOP	30	M	M	
11.	B021 WEST LAB FLOOR	29	M	M	
12.	B021 WORK PLATFORM	51	3.29E-05	3.29E-07	
13.	B021 CALORIMETER	31	M	M	
14.	B021 COMPUTERS	25	M	M	
15.	B021 DSC/TGA	32	M	M	
16.	B021 SOUTH BENCH TOP	29	M	M	
17.	121 DOORWAY	28	M	M	
18.	121 GLOVE BOX- FLOOR	28	M	M	
19.	121 GLOVE BOX	29	M	M	
20.	121 WEST BENCH TOP	29	M	M	
21.	121 CENTER OF LAB	28	M	M	
22.	121 EAST BENCH TOP	34	9.10E-06	9.10E-08	
None					
³⁶ CI SOURCE 09/14/1971		50275		*****	*****
		101.37%		CALIBRATION CHECK OK? YES	
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: M. Heine			REVIEWED BY: C. Hines		
SIGN/DATE: <i>Maddalena Heine</i> 4/18/19			SIGN/DATE: <i>C. Hines</i> 4/18/2019		

Assay Definition-

Assay Description:

Assay Type: CPM
 Report Name: Report1
 Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190416_1403
 Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190416_1403\20190416_1403.results
 Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor
 Quench Indicator: tSIE/AEC
 External Std Terminator (sec): 0.5 2s%
 Pre-Count Delay (min): 0.00
 Quench Set: n/a
 Count Time (min): 5.00
 Count Mode: Normal
 Assay Count Cycles: 1 Repeat Sample Count: 1
 #Vials/Sample: 1 Calculate % Reference: Off

Background Subtract: Off
 Low CPM Threshold: Off
 2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On Luminescence Correction: Off
 Colored Samples: n/a Heterogeneity Monitor: n/a
 Coincidence Time (nsec): 18 Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	4/16/2019	2:04:20 PM	1	5.00		21	6	27	330.41	
1	4/16/2019	2:10:34 PM	2	5.00		22	4	27	237.24	
1	4/16/2019	2:16:50 PM	3	5.00		22	5	26	217.29	
1	4/16/2019	2:23:03 PM	4	5.00		21	5	26	304.32	
1	4/16/2019	2:29:17 PM	5	5.00		23	5	28	307.23	
1	4/16/2019	2:35:31 PM	6	5.00		26	5	31	217.90	
1	4/16/2019	2:41:47 PM	7	5.00		26	6	32	232.29	
1	4/16/2019	2:48:04 PM	8	5.00		22	9	31	270.16	
1	4/16/2019	2:54:21 PM	9	5.00		22	5	27	293.71	
1	4/16/2019	3:00:36 PM	10	5.00		25	4	30	171.74	
1	4/16/2019	3:06:55 PM	11	5.00		22	6	29	254.04	
1	4/16/2019	3:13:09 PM	12	5.00		45	6	51	233.78	


1	4/16/2019	3:19:29 PM	13	5.00	25	6	31	276.39
1	4/16/2019	3:25:45 PM	14	5.00	21	4	25	247.64
1	4/16/2019	3:31:54 PM	15	5.00	25	7	32	243.83
1	4/16/2019	3:38:10 PM	16	5.00	23	7	29	275.67
1	4/16/2019	3:44:26 PM	17	5.00	23	5	28	249.86
1	4/16/2019	3:50:37 PM	18	5.00	23	5	28	182.83
1	4/16/2019	3:56:54 PM	19	5.00	23	6	29	197.07
1	4/16/2019	4:03:06 PM	20	5.00	25	5	29	211.97
1	4/16/2019	4:09:21 PM	21	5.00	23	5	28	227.77
1	4/16/2019	4:15:37 PM	22	5.00	27	7	34	233.82
1	4/16/2019	4:21:18 PM	23	5.00	14584	35691	50275	1010.23

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument Ludlum Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 4/16/19	Time 12:40	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input checked="" type="checkbox"/> OFF <input type="checkbox"/>					

LSC Printout Date 4/16/19





Survey #	Location	CPM	Surveyed By:	Date:	Notes
11	121 Doorway	250	AH	4/16/2019	
12	121 Center of lab	± 100	JS	4/16/19	
13	121 Glove Box	150	awp	4/18/2019	
17	121 Doorway				
18	121 Glove Box-floor				
19	121 Glove Box				
20	West Bench Top				
21	121 Center of lab				
22	East Bench Top				

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	17	Dates:	4/21/2019-4/27/2019	Counted On	4/23/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm2 = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.94E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	28	*****	*****	
2.	BACKGROUND 2	30	*****	*****	
3.	BACKGROUND 3	32	*****	*****	
4.	BACKGROUND 4	31	*****	*****	
5.	BACKGROUND 5	27	*****	*****	
6.	BACKGROUND 6	32	*****	*****	
7.	B021 ELEVATOR ENTRANCE	27	M	M	
8.	B021 FENCED GATE	27	M	M	
9.	B021 EAST FLOOR	26	M	M	
10.	B021 NORTH BENCH TOP	25	M	M	
11.	B021 WEST LAB FLOOR	22	M	M	
12.	B021 WORK PLATFORM	27	M	M	
13.	B021 CALORIMETER	31	M	M	
14.	B021 COMPUTERS	27	M	M	
15.	B021 DSC/TGA	23	M	M	
16.	B021 SOUTH BENCH TOP	26	M	M	
17.	121 DOORWAY	31	M	M	
18.	121 GLOVE BOX- FLOOR	39	1.26E-05	1.26E-07	
19.	121 GLOVE BOX	24	M	M	
20.	121 WEST BENCH TOP	27	M	M	
21.	121 CENTER OF LAB	35	M	M	
22.	121 EAST BENCH TOP	27	M	M	
None					
³⁶ CI SOURCE 09/14/1971		50046	*****	*****	
		100.91%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: C. Hines		
SIGN/DATE:  / 4/24/2019			SIGN/DATE:  4/24/2019		

Assay Definition-

Assay Description:

Assay Type: CPM
 Report Name: Report1
 Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190423_1443
 Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190423_1443\20190423_1443.results
 Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor
 Quench Indicator: tSIE/AEC
 External Std Terminator (sec): 0.5 2s%
 Pre-Count Delay (min): 0.00
 Quench Set: n/a
 Count Time (min): 5.00
 Count Mode: Normal
 Assay Count Cycles: 1 Repeat Sample Count: 1
 #Vials/Sample: 1 Calculate % Reference: Off

Background Subtract: Off
 Low CPM Threshold: Off
 2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On Luminescence Correction: Off
 Colored Samples: n/a Heterogeneity Monitor: n/a
 Coincidence Time (nsec): 18 Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	4/23/2019	2:45:01 PM	1	5.00		22	6	28	312.85	
1	4/23/2019	2:51:17 PM	2	5.00		21	9	30	331.88	
1	4/23/2019	2:57:35 PM	3	5.00		25	6	32	230.63	
1	4/23/2019	3:03:46 PM	4	5.00		24	6	31	255.99	
1	4/23/2019	3:10:01 PM	5	5.00		21	6	27	249.46	
1	4/23/2019	3:16:15 PM	6	5.00		24	7	32	268.27	
1	4/23/2019	3:22:31 PM	7	5.00		22	5	27	290.85	
1	4/23/2019	3:28:50 PM	8	5.00		21	5	27	230.23	
1	4/23/2019	3:35:07 PM	9	5.00		23	4	26	141.01	
1	4/23/2019	3:41:21 PM	10	5.00		19	7	25	287.48	
1	4/23/2019	3:47:39 PM	11	5.00		16	6	22	308.11	
1	4/23/2019	3:53:57 PM	12	5.00		22	5	27	239.04	

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	4/23/2019	4:00:16 PM	13	5.00	26	5	31	215.40
1	4/23/2019	4:06:30 PM	14	5.00	22	4	27	198.77
1	4/23/2019	4:12:43 PM	15	5.00	20	4	23	167.50
1	4/23/2019	4:19:00 PM	16	5.00	21	5	26	242.29
1	4/23/2019	4:25:14 PM	17	5.00	25	6	31	264.33
1	4/23/2019	4:31:32 PM	18	5.00	29	10	39	204.54
1	4/23/2019	4:37:49 PM	19	5.00	19	5	24	287.52
1	4/23/2019	4:44:03 PM	20	5.00	20	7	27	227.15
1	4/23/2019	4:50:17 PM	21	5.00	27	8	35	172.94
1	4/23/2019	4:56:36 PM	22	5.00	22	5	27	126.42
1	4/23/2019	5:02:16 PM	23	5.00	14378	35669	50046	1016.44

LABORATORY SURVEY

Radiation Safety Office
Washington State University
Pullman, WA 99164-1302
(509) 335 8916

Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument GM Detector Model 12	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 4/23/19	Time 1413	Swipe Instrument Quantasmat	Radiation Detected Beta/Gamma
Reactor Status (Check One)		LSC Printout Date 4/23/2019		
ON <input type="checkbox"/>	OFF <input checked="" type="checkbox"/>			

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP		
11	121 Doorway	≤ 100	<p style="text-align: center;">DODG 121 GUO LAB</p>		
12	121 Center of lab	≤ 100			
13	121 Glove Box	≤ 100			
Swipe # and Locations			<p style="text-align: center;">West Bench Top</p>		
17	121 Doorway				
18	121 Glove Box-floor				
19	121 Glove Box				
20	West Bench Top				
21	121 Center of lab				
22	East Bench Top		Glove Box		
			Notes		
			Surveyed By: JS	Date: 4/23/19	
			Swiped By: AG	Date: 4/23/19	
			Reviewed By: JS	Date: 4/23/19	
			Signature: <i>[Handwritten Signature]</i> Date: 4/24/2019		

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	18	Dates:	4/28/2019-5/4/2019	Counted On	4/30/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.		uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100	
Limit of quantification (LOQ)=		8.84E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	29	*****	*****	
2.	BACKGROUND 2	30	*****	*****	
3.	BACKGROUND 3	28	*****	*****	
4.	BACKGROUND 4	30	*****	*****	
5.	BACKGROUND 5	29	*****	*****	
6.	BACKGROUND 6	26	*****	*****	
7.	B021 ELEVATOR ENTRANCE	29	M	M	
8.	B021 FENCED GATE	29	M	M	
9.	B021 EAST FLOOR	27	M	M	
10.	B021 NORTH BENCH TOP	25	M	M	
11.	B021 WEST LAB FLOOR	29	M	M	
12.	B021 WORK PLATFORM	33	M	M	
13.	B021 CALORIMETER	32	M	M	
14.	B021 COMPUTERS	26	M	M	
15.	B021 DSC/TGA	27	M	M	
16.	B021 SOUTH BENCH TOP	26	M	M	
17.	121 DOORWAY	29	M	M	
18.	121 GLOVE BOX- FLOOR	43	2.01E-05	2.01E-07	
19.	121 GLOVE BOX	29	M	M	
20.	121 WEST BENCH TOP	32	M	M	
21.	121 CENTER OF LAB	30	M	M	
22.	121 EAST BENCH TOP	32	M	M	
None					
³⁶ CI SOURCE 09/14/1971		50025		*****	*****
		100.87%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: M.Heine			REVIEWED BY: C. Hines		
SIGN/DATE: <i>Maddison Heine 5/1/19</i>			SIGN/DATE: <i>amgum 5/1/2019</i>		

Assay Definition-

Assay Description:

Assay Type: CPM
 Report Name: Report1
 Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190430_1423
 Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190430_1423\20190430_1423.results
 Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor
 Quench Indicator: tSIE/AEC
 External Std Terminator (sec): 0.5 2s%
 Pre-Count Delay (min): 0.00
 Quench Set: n/a
 Count Time (min): 5.00
 Count Mode: Normal
 Assay Count Cycles: 1 Repeat Sample Count: 1
 #Vials/Sample: 1 Calculate % Reference: Off

Background Subtract: Off
 Low CPM Threshold: Off
 2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On Luminescence Correction: Off
 Colored Samples: n/a Heterogeneity Monitor: n/a
 Coincidence Time (nsec): 18 Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	4/30/2019	2:24:15 PM	1		5.00	23	6	29	277.08	
1	4/30/2019	2:30:24 PM	2		5.00	22	8	30	413.67	
1	4/30/2019	2:36:40 PM	3		5.00	24	4	28	220.34	
1	4/30/2019	2:42:50 PM	4		5.00	24	7	30	316.16	
1	4/30/2019	2:49:04 PM	5		5.00	24	5	29	232.66	
1	4/30/2019	2:55:21 PM	6		5.00	19	7	26	341.36	
1	4/30/2019	3:01:40 PM	7		5.00	24	5	29	202.94	
1	4/30/2019	3:07:55 PM	8		5.00	24	5	29	223.61	
1	4/30/2019	3:14:14 PM	9		5.00	22	5	27	208.10	
1	4/30/2019	3:20:28 PM	10		5.00	21	4	25	258.55	
1	4/30/2019	3:26:44 PM	11		5.00	23	6	29	199.53	
1	4/30/2019	3:32:57 PM	12		5.00	26	7	33	299.69	

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	4/30/2019	3:39:19 PM	13	5.00	27	6	32	291.45
1	4/30/2019	3:45:35 PM	14	5.00	21	5	26	299.55
1	4/30/2019	3:51:51 PM	15	5.00	22	5	27	254.41
1	4/30/2019	3:58:07 PM	16	5.00	21	5	26	290.09
1	4/30/2019	4:04:16 PM	17	5.00	23	6	29	251.45
1	4/30/2019	4:10:32 PM	18	5.00	30	13	43	253.84
1	4/30/2019	4:16:48 PM	19	5.00	22	7	29	355.97
1	4/30/2019	4:23:03 PM	20	5.00	23	9	32	366.42
1	4/30/2019	4:29:22 PM	21	5.00	24	5	30	158.81
1	4/30/2019	4:35:35 PM	22	5.00	25	7	32	317.13
1	4/30/2019	4:41:15 PM	23	5.00	14544	35481	50025	1015.95

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	19	Dates:	5/5/2019-5/11/2019	Counted On	5/13/2019
INSTRUMENT USED: Quantasart (Serial# 073396)			CONTROL FACTOR: 1.40E-06		
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		9.03E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	

Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	34	*****	*****	
2.	BACKGROUND 2	29	*****	*****	
3.	BACKGROUND 3	29	*****	*****	
4.	BACKGROUND 4	32	*****	*****	
5.	BACKGROUND 5	31	*****	*****	
6.	BACKGROUND 6	32	*****	*****	
7.	B021 ELEVATOR ENTRANCE	30	M	M	
8.	B021 FENCED GATE	28	M	M	
9.	B021 EAST FLOOR	34	M	M	
10.	B021 NORTH BENCH TOP	29	M	M	
11.	B021 WEST LAB FLOOR	33	M	M	
12.	B021 WORK PLATFORM	30	M	M	
13.	B021 CALORIMETER	33	M	M	
14.	B021 COMPUTERS	34	M	M	
15.	B021 DSC/TGA	38	9.57E-06	9.57E-08	
16.	B021 SOUTH BENCH TOP	29	M	M	
17.	121 DOORWAY	39	1.10E-05	1.10E-07	
18.	121 GLOVE BOX- FLOOR	30	M	M	
19.	121 GLOVE BOX	33	M	M	
20.	121 WEST BENCH TOP	32	M	M	
21.	121 CENTER OF LAB	30	M	M	
22.	121 EAST BENCH TOP	32	M	M	

Swipe 13 not taken due to an experiment in progress with the calorimeter. A background swipe was put in its place.

³⁶ CI SOURCE 09/14/1971	49988	*****	*****	
		100.79%	CALIBRATION CHECK OK?	YES

Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10⁻⁶ uCi/cm².

SURVEY PREPARED BY: B. Tanner	REVIEWED BY: C. Hines
-------------------------------	-----------------------

SIGN/DATE: 	SIGN/DATE:  5/13/2019
--	--

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190513_0810

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190513_0810\20190513_0810.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	5/13/2019	8:11:15 AM	1	5.00	28	6	34	197.72		
1	5/13/2019	8:17:22 AM	2	5.00	23	6	29	270.09		
1	5/13/2019	8:23:34 AM	3	5.00	23	6	29	257.44		
1	5/13/2019	8:29:42 AM	4	5.00	26	5	32	227.26		
1	5/13/2019	8:35:55 AM	5	5.00	25	6	31	220.46		
1	5/13/2019	8:42:08 AM	6	5.00	28	5	32	154.74		
1	5/13/2019	8:48:23 AM	7	5.00	24	6	30	300.83		
1	5/13/2019	8:54:36 AM	8	5.00	24	4	28	170.87		
1	5/13/2019	9:00:50 AM	9	5.00	28	6	34	218.10		
1	5/13/2019	9:07:00 AM	10	5.00	24	5	29	294.33		
1	5/13/2019	9:13:12 AM	11	5.00	27	6	33	180.58		
1	5/13/2019	9:19:26 AM	12	5.00	23	7	30	273.74		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	5/13/2019	9:25:44 AM	13	5.00	27	6	33	239.13
1	5/13/2019	9:31:58 AM	14	5.00	27	7	34	282.05
1	5/13/2019	9:38:05 AM	15	5.00	29	8	38	278.82
1	5/13/2019	9:44:18 AM	16	5.00	25	4	29	193.70
1	5/13/2019	9:50:33 AM	17	5.00	32	7	39	207.05
1	5/13/2019	9:56:47 AM	18	5.00	26	5	30	213.86
1	5/13/2019	10:03:01 AM	19	5.00	25	8	33	339.13
1	5/13/2019	10:09:14 AM	20	5.00	25	7	32	261.28
1	5/13/2019	10:15:22 AM	21	5.00	22	7	30	267.38
1	5/13/2019	10:21:36 AM	22	5.00	26	6	32	258.17
1	5/13/2019	10:27:17 AM	23	5.00	14247	35742	49988	1014.17

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building	Dodgen Research Facility	Room	B0021 - GUO	Lab Class	D	Survey Instrument	GM Detector Model 12	Serial No.	172178	Radiation Detected	Gamma
Authorized User	Xiaofeng Guo	Date	5/9/19	Time	1130	Swipe Instrument	Quantasmat	Serial No.	073396	Radiation Detected	Beta/Gamma
Reactor Status (Check One)	ON <input checked="" type="checkbox"/> OFF <input type="checkbox"/>	LSC Printout Date <u>5/13/2019</u> <u>5/19/2019 BT</u>									

Survey #	Location	CPM	Notes
1	Elevator Entrance	80	
2	South Floor	120	
3	East Lab Floor	200 CPH 600	
4	North Bench Top	120	
5	West Floor	80	
6	Work Platform	100	
7	Calorimeter	100	
8	Computers	100	
9	DSC/TGA	120	
10	South Bench Top	120	
Swipe # and Locations			
1-6	Background	16	South Bench Top
7	Elevator Entrance		
8	South Floor		
9	East Lab Floor		
10	North Bench Top		
11	West Lab Floor		
12	Work Platform		
13	Calorimeter		
14	Computers		
15	DSC/TGA		

NOTE SWIPE AND SURVEY LOCATIONS ON MAP

DODG B0021 GUO LAB

Surveys are marked with circled numbers on the map: 1 (Elevator Entrance), 2 (South Bench Top), 3 (East Lab Floor), 4 (North Bench Top), 5 (West Floor), 6 (Work Platform), 7 (Calorimeter), 8 (Computers), 9 (DSC/TGA), 10 (South Bench Top), 11 (South Floor), 12 (Work Platform), 13 (Calorimeter), 14 (Computers), 15 (DSC/TGA), 16 (South Bench Top).

Surveyed By: CPH Date: 5/9/2019
 Swiped By: MS Date: 5/9/19
 Reviewed By: [Signature] Date: 5/13/2019

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building	Dodgen Research Facility	Room	121- GUO	Lab Class	D	Survey Instrument	GM Detector Model 12	Serial No.	172178	Radiation Detected	Gamma
Authorized User	Xiaofeng Guo	Date	5/9/19	Time	1146	Swipe Instrument	Quantasart	Serial No.	073396	Radiation Detected	Beta/Gamma
Reactor Status (Check One)	<input checked="" type="checkbox"/> ON <input type="checkbox"/> OFF										



LSC Printout Date 5/13/2019

Survey #	Location	CPM	Notes
11	121 Doorway	110	<p style="text-align: center;">NOTE SWIPE AND SURVEY LOCATIONS ON MAP</p> <p style="text-align: center;">DODG 121 GUO LAB</p>
12	121 Center of lab	120	
13	121 Glove Box	240	
Swipe # and Locations			
17	121 Doorway		
18	121 Glove Box-floor		
19	121 Glove Box		
20	West Bench Top		
21	121 Center of lab		
22	East Bench Top		
Surveyed By: <u>CMH</u> Date: <u>5/9/2019</u>			
Swiped By: <u>MS</u> Date: <u>5/9/19</u>			
Reviewed By: <u>[Signature]</u> Date: <u>5/13/2019</u>			

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	20	Dates:	5/12/2019-5/19/2019	Counted On	5/14/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.		uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100	
Limit of quantification (LOQ)=		9.00E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	31	*****	*****	
2.	BACKGROUND 2	27	*****	*****	
3.	BACKGROUND 3	32	*****	*****	
4.	BACKGROUND 4	35	*****	*****	
5.	BACKGROUND 5	28	*****	*****	
6.	BACKGROUND 6	32	*****	*****	
7.	B021 ELEVATOR ENTRANCE	22	M	M	
8.	B021 FENCED GATE	27	M	M	
9.	B021 RAM SAFE	26	M	M	
10.	B021 EAST FLOOR	33	M	M	
11.	B021 NORTH BENCH TOP	22	M	M	
12.	B021 WEST LAB FLOOR	29	M	M	
13.	B021 WORK PLATFORM	27	M	M	
14.	B021 CALORIMETER	30	M	M	
15.	B021 COMPUTERS	29	M	M	
16.	B021 DSC/TGA	29	M	M	
17.	B021 SOUTH BENCH TOP	31	M	M	
18.	121 DOORWAY	38	1.00E-05	1.00E-07	
19.	121 GLOVE BOX- FLOOR	26	M	M	
20.	121 GLOVE BOX	33	M	M	
21.	121 WEST BENCH TOP	27	M	M	
22.	121 CENTER OF LAB	30	M	M	
23.	122 EAST BENCH TOP	25	M	M	
None					
³⁶ CI SOURCE 09/14/1971		50154		*****	*****
		101.13%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: C. Hines		
SIGN/DATE:  5/15/2019			SIGN/DATE:  5/15/2019		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190514_1527

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190514_1527\20190514_1527.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	5/14/2019	3:28:48 PM	1	5.00	24	7	31	210.79		
1	5/14/2019	3:34:57 PM	2	5.00	21	6	27	287.79		
1	5/14/2019	3:41:15 PM	3	5.00	26	6	32	281.03		
1	5/14/2019	3:47:24 PM	4	5.00	28	6	35	203.63		
1	5/14/2019	3:53:40 PM	5	5.00	21	6	28	326.34		
1	5/14/2019	3:59:57 PM	6	5.00	25	7	32	339.26		
1	5/14/2019	4:06:15 PM	7	5.00	17	5	22	262.75		
1	5/14/2019	4:12:31 PM	8	5.00	21	6	27	247.73		
1	5/14/2019	4:18:50 PM	9	5.00	21	5	26	182.35		
1	5/14/2019	4:25:02 PM	10	5.00	28	5	33	164.16		
1	5/14/2019	4:31:19 PM	11	5.00	17	5	22	189.41		
1	5/14/2019	4:37:34 PM	12	5.00	24	5	29	203.69		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	5/14/2019	4:43:56 PM	13	5.00	22	5	27	240.34
1	5/14/2019	4:50:14 PM	14	5.00	25	5	30	237.69
1	5/14/2019	4:56:24 PM	15	5.00	23	5	29	203.23
1	5/14/2019	5:02:41 PM	16	5.00	23	6	29	243.92
1	5/14/2019	5:08:59 PM	17	5.00	23	7	31	318.34
1	5/14/2019	5:15:17 PM	18	5.00	25	13	38	324.38
1	5/14/2019	5:21:34 PM	19	5.00	20	5	26	231.92
1	5/14/2019	5:27:51 PM	20	5.00	27	6	33	221.48
1	5/14/2019	5:33:59 PM	21	5.00	22	5	27	177.92
1	5/14/2019	5:40:16 PM	22	5.00	24	6	30	197.88
1	5/14/2019	5:46:28 PM	23	5.00	20	6	25	223.97
1	5/14/2019	5:52:08 PM	24	5.00	14410	35744	50154	1009.53

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 5/14/19	Time 1402	Swipe Instrument Quantasart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>		LSC Printout Date 5/14/19			

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP	
1	Elevator Entrance	120	<p style="text-align: center;">DODG B0021 GUO LAB</p>	
2	South Floor	120		
3	RAM Safe	1600		
4	East Lab Floor	80		
5	North Bench Top	166		
6	West Floor	100		
7	Work Platform	80		
8	Calorimeter	160		
9	Computers	50		
10	DSC/TGA	120		
11	South Bench Top	100		
Swipe # and Locations				
1-6	Background	16		DSC/TGA
7	Elevator Entrance	17		South Bench Top
8	South Floor			
9	RAM Safe			
10	East Lab Floor			
11	North Bench Top			
12	West Lab Floor			
13	Work Platform			
14	Calorimeter			
15	Computers			

Surveyed By: BT	Date: 5/14/19	Notes
Swiped By: BT	Date: 5/14/19	
Reviewed By: <i>[Signature]</i>	Date: 5/15/2019	

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building	Dodgen Research Facility	Room	121-GUO	Lab Class	D	Survey Instrument	GM Detector Model 12	Serial No.	172178	Radiation Detected	Gamma
Authorized User	Xiaofeng Guo	Date	5/14/19	Time	1418	Swipe Instrument	Quantasart	Serial No.	073396	Radiation Detected	Beta/Gamma
Reactor Status (Check One)	<input type="checkbox"/> ON <input checked="" type="checkbox"/> OFF										

LSC Printout Date 5/14/19

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP			
12	121 Doorway	100				
13	121 Center of lab	120				
14	121 Glove Box	120				
Swipe # and Locations						
18	121 Doorway					
19	121 Glove Box-floor					
20	121 Glove Box					
21	West Bench Top					
22	121 Center of lab					
23	East Bench Top					

Surveyed By: BT Date: 5/14/19
 Swiped By: BT Date: 5/14/19
 Reviewed By: [Signature] Date: 5/15/2019

Notes

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	21	Dates:	5/19/2019-5/25/2019	Counted On	5/21/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.98E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	



Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	23	*****	*****	
2.	BACKGROUND 2	26	*****	*****	
3.	BACKGROUND 3	23	*****	*****	
4.	BACKGROUND 4	28	*****	*****	
5.	BACKGROUND 5	66	*****	*****	
6.	BACKGROUND 6	25	*****	*****	
7.	B021 ELEVATOR ENTRANCE	30	M	M	
8.	B021 FENCED GATE	31	M	M	
9.	B021 RAM SAFE	21	M	M	
10.	B021 EAST FLOOR	27	M	M	
11.	B021 NORTH BENCH TOP	26	M	M	
12.	B021 WEST LAB FLOOR	31	M	M	
13.	B021 WORK PLATFORM	34	M	M	
14.	B021 CALORIMETER	29	M	M	
15.	B021 COMPUTERS	35	M	M	
16.	B021 DSC/TGA	38	M	M	
17.	B021 SOUTH BENCH TOP	33	M	M	
18.	121 DOORWAY	33	M	M	
19.	121 GLOVE BOX- FLOOR	29	M	M	
20.	121 GLOVE BOX	122	1.26E-04	1.26E-06	
21.	121 WEST BENCH TOP	29	M	M	
22.	121 CENTER OF LAB	35	M	M	
23.	122 EAST BENCH TOP	32	M	M	

Swipe 20 came back 99 over the lowest background. Additional swipes taken of the glovebox to ensure no contamination and the results are attached to this form.

³⁶ CI SOURCE 09/14/1971	49886	*****	*****	
	100.59%	CALIBRATION CHECK OK?		YES

Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10⁻⁶ uCi/cm².

SURVEY PREPARED BY: B. Tanner	REVIEWED BY: C. Hines
-------------------------------	-----------------------

SIGN/DATE:  / 5/22/19	SIGN/DATE:  5/31/2019
--	---

Assay Definition-

Assay Description:

Assay Type: CPM
 Report Name: Report1
 Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190521_1400
 Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190521_1400\20190521_1400.results
 Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor
 Quench Indicator: tSIE/AEC
 External Std Terminator (sec): 0.5 2s%
 Pre-Count Delay (min): 0.00
 Quench Set: n/a
 Count Time (min): 5.00
 Count Mode: Normal
 Assay Count Cycles: 1 Repeat Sample Count: 1
 #Vials/Sample: 1 Calculate % Reference: Off

Background Subtract: Off
 Low CPM Threshold: Off
 2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On Luminescence Correction: Off
 Colored Samples: n/a Heterogeneity Monitor: n/a
 Coincidence Time (nsec): 18 Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	5/21/2019	2:01:36 PM	1	5.00	20	3	23	224.37		
1	5/21/2019	2:07:53 PM	2	5.00	21	5	26	268.65		
1	5/21/2019	2:14:10 PM	3	5.00	19	4	23	220.97		
1	5/21/2019	2:20:25 PM	4	5.00	23	5	28	269.29		
1	5/21/2019	2:26:40 PM	5	5.00	50	16	66	266.00		
1	5/21/2019	2:32:54 PM	6	5.00	21	3	25	243.48		
1	5/21/2019	2:39:11 PM	7	5.00	24	6	30	196.28		
1	5/21/2019	2:45:28 PM	8	5.00	26	6	31	242.76		
1	5/21/2019	2:51:45 PM	9	5.00	17	4	21	247.69		
1	5/21/2019	2:58:01 PM	10	5.00	21	6	27	242.41		
1	5/21/2019	3:04:18 PM	11	5.00	21	4	26	295.47		
1	5/21/2019	3:10:38 PM	12	5.00	25	6	31	240.05		

1	5/21/2019	3:17:01 PM	13	5.00	26	8	34	259.68
1	5/21/2019	3:23:19 PM	14	5.00	23	6	29	246.66
1	5/21/2019	3:29:29 PM	15	5.00	30	5	35	236.89
1	5/21/2019	3:35:46 PM	16	5.00	31	7	38	206.05
1	5/21/2019	3:42:00 PM	17	5.00	25	8	33	329.32
1	5/21/2019	3:48:16 PM	18	5.00	26	7	33	293.91
1	5/21/2019	3:54:35 PM	19	5.00	22	7	29	276.89
1	5/21/2019	4:00:53 PM	20	5.00	114	8	122	149.15
1	5/21/2019	4:07:12 PM	21	5.00	22	6	29	248.25
1	5/21/2019	4:13:32 PM	22	5.00	28	6	35	141.58
1	5/21/2019	4:19:48 PM	23	5.00	27	5	32	194.32
1	5/21/2019	4:25:28 PM	24	5.00	14345	35541	49886	1016.55

Cycle 1 Results

S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1		1.00	10	11	28	42.48	
2		1.00	6	12	21	39.45	
3		1.00	6	12	21	42.74	
4		1.00	4	14	26	37.86	
5		1.00	5	9	17	33.75	

- 1) BACKGROUND
- 2) SOUTH HALF OF GLASS FACE
- 3) SOUTH GLOVE (LEFT HAND)
- 4) NORTH HALF OF GLASS FACE
- 5) NORTH GLOVE (RIGHT HAND)

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	22	Dates:	5/26/2019-6/1/2019	Counted On	5/29/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.		uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100	
Limit of quantification (LOQ)=		8.84E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	26	*****	*****	
2.	BACKGROUND 2	30	*****	*****	
3.	BACKGROUND 3	29	*****	*****	
4.	BACKGROUND 4	29	*****	*****	
5.	BACKGROUND 5	28	*****	*****	
6.	BACKGROUND 6	30	*****	*****	
7.	B021 ELEVATOR ENTRANCE	34	M	M	
8.	B021 FENCED GATE	26	M	M	
9.	B021 RAM SAFE	30	M	M	
10.	B021 EAST FLOOR	29	M	M	
11.	B021 NORTH BENCH TOP	29	M	M	
12.	B021 WEST LAB FLOOR	31	M	M	
13.	B021 WORK PLATFORM	27	M	M	
14.	B021 CALORIMETER	43	2.01E-05	2.01E-07	
15.	B021 COMPUTERS	31	M	M	
16.	B021 DSC/TGA	26	M	M	
17.	B021 SOUTH BENCH TOP	29	M	M	
18.	121 DOORWAY	26	M	M	
19.	121 GLOVE BOX- FLOOR	28	M	M	
20.	121 GLOVE BOX	28	M	M	
21.	121 WEST BENCH TOP	32	M	M	
22.	121 CENTER OF LAB	25	M	M	
23.	122 EAST BENCH TOP	31	M	M	
³⁶ CI SOURCE 09/14/1971		50202		*****	*****
		101.22%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: C. Hines		
SIGN/DATE: <i>[Signature]</i> / 5/30/2019			SIGN/DATE: <i>[Signature]</i> / 5/31/2019		

Assay Definition-

Assay Description:

Assay Type: CPM
Report Name: Report1
Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190529_1317
Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190529_1317\20190529_1317.results
Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor
Quench Indicator: tSIE/AEC
External Std Terminator (sec): 0.5 2s%
Pre-Count Delay (min): 0.00
Quench Set: n/a
Count Time (min): 5.00
Count Mode: Normal
Assay Count Cycles: 1 Repeat Sample Count: 1
#Vials/Sample: 1 Calculate % Reference: Off

Background Subtract: Off
Low CPM Threshold: Off
2 Sigma % Terminator: Off

Table with 3 columns: Regions, LL, UL. Rows A, B, C with values for LL and UL.

Count Corrections-

Static Controller: On Luminescence Correction: Off
Colored Samples: n/a Heterogeneity Monitor: n/a
Coincidence Time (nsec): 18 Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Table with 5 columns: Regions, Half Life, Units, Reference Date, Reference Time. Rows A, B, C.

Cycle 1 Results

Table with 10 columns: P#, DATE, TIME, S#, Count, Time, CPMA, CPMB, CPMC, SIS, MESSAGES. Contains 12 rows of cycle results.

Protocol# 1 - Reactor Weekly.lsa

User: CSE


1	5/29/2019	2:33:40 PM	13	5.00	22	5	27	215.37
1	5/29/2019	2:39:57 PM	14	5.00	32	10	43	341.43
1	5/29/2019	2:46:07 PM	15	5.00	25	6	31	305.29
1	5/29/2019	2:52:15 PM	16	5.00	21	5	26	247.96
1	5/29/2019	2:58:30 PM	17	5.00	24	4	29	181.88
1	5/29/2019	3:04:45 PM	18	5.00	21	5	26	203.14
1	5/29/2019	3:11:03 PM	19	5.00	23	5	28	190.34
1	5/29/2019	3:17:18 PM	20	5.00	23	5	28	206.27
1	5/29/2019	3:23:34 PM	21	5.00	27	5	32	237.59
1	5/29/2019	3:29:45 PM	22	5.00	20	5	25	179.14
1	5/29/2019	3:36:02 PM	23	5.00	25	6	31	270.33
1	5/29/2019	3:41:42 PM	24	5.00	14539	35663	50202	1011.12

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 5/29/19	Time 1:23	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) <input type="checkbox"/> ON <input checked="" type="checkbox"/> OFF					

LSC Printout Date 5/29/2019



Swipe # and Locations	CPM	Surveyed By:	Date:	Notes
1-6 Background	150	MJ	5/29/19	14 ms Swipe #15 not taken since calorimeter was remaining MJ
7 Elevator Entrance	150			
8 South Floor	1300			
9 RAM Safe	<100			
10 East Lab Floor	175			
11 North Bench Top	<100			
12 West Lab Floor	<100			
13 Work Platform	<100			
14 Calorimeter	<100			
15 Computers	<100			
Reviewed By:		MJ	5/29/19	
Date:			5/29/19	
Reviewed By:			5/31/2019	
Date:				


Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 5/29/19	Time 1317	Swipe Instrument Quantasart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One)					
ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>					

LSC Printout Date **5/29/2019**



Survey #	Location	CPM	Surveyed By:	Date:	Notes
12	121 Doorway	<100	MJ	5/29/19	
13	121 Center of lab	<100	MJ	5/29/19	
14	121 Glove Box	<100	MJ	5/29/19	
15					
16					
17					
18	121 Doorway				
19	121 Glove Box-floor				
20	121 Glove Box				
21	West Bench Top				
22	121 Center of lab				
23	East Bench Top				

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	23	Dates:	6/2/2019-6/8/2019	Counted On	6/4/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.93E-08 uCi/cm ²	QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK		

Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	31	*****	*****	
2.	BACKGROUND 2	29	*****	*****	
3.	BACKGROUND 3	28	*****	*****	
4.	BACKGROUND 4	31	*****	*****	
5.	BACKGROUND 5	30	*****	*****	
6.	BACKGROUND 6	30	*****	*****	
7.	B021 ELEVATOR ENTRANCE	24	M	M	
8.	B021 FENCED GATE	26	M	M	
9.	B021 RAM SAFE	30	M	M	
10.	B021 EAST FLOOR	31	M	M	
11.	B021 NORTH BENCH TOP	26	M	M	
12.	B021 WEST LAB FLOOR	27	M	M	
13.	B021 WORK PLATFORM	31	M	M	
14.	B021 CALORIMETER	37	1.00E-05	1.00E-07	
15.	B021 COMPUTERS	31	M	M	
16.	B021 DSC/TGA	33	M	M	
17.	B021 SOUTH BENCH TOP	31	M	M	
18.	121 DOORWAY	27	M	M	
19.	121 GLOVE BOX- FLOOR	31	M	M	
20.	121 GLOVE BOX	29	M	M	
21.	121 WEST BENCH TOP	27	M	M	
22.	121 CENTER OF LAB	25	M	M	
23.	122 EAST BENCH TOP	27	M	M	



--	--	--	--	--

³⁶ Cl SOURCE 09/14/1971	50223	*****	*****	
------------------------------------	-------	-------	-------	--

	101.26%	CALIBRATION CHECK OK?	YES
--	---------	-----------------------	-----

Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10⁻⁶ uCi/cm².

SURVEY PREPARED BY: B. Tanner	REVIEWED BY: C. Hines
-------------------------------	-----------------------

SIGN/DATE:  6/5/2019	SIGN/DATE:  6/5/2019
---	---

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190604_1529

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190604_1529\20190604_1529.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	6/4/2019	3:30:51 PM	1	5.00	26	5	31	238.36	
1	6/4/2019	3:37:01 PM	2	5.00	22	7	29	310.51	
1	6/4/2019	3:43:19 PM	3	5.00	21	6	28	361.88	
1	6/4/2019	3:49:29 PM	4	5.00	25	6	31	323.98	
1	6/4/2019	3:55:43 PM	5	5.00	23	6	30	441.13	
1	6/4/2019	4:02:00 PM	6	5.00	25	5	30	268.72	
1	6/4/2019	4:08:19 PM	7	5.00	20	4	24	220.74	
1	6/4/2019	4:14:34 PM	8	5.00	19	7	26	352.98	
1	6/4/2019	4:20:49 PM	9	5.00	22	7	30	315.93	
1	6/4/2019	4:27:03 PM	10	5.00	26	5	31	180.84	
1	6/4/2019	4:33:18 PM	11	5.00	20	6	26	283.29	
1	6/4/2019	4:39:34 PM	12	5.00	21	6	27	224.60	

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	6/4/2019	4:45:54 PM	13	5.00	26	5	31	234.83
1	6/4/2019	4:52:11 PM	14	5.00	29	7	37	292.14
1	6/4/2019	4:58:21 PM	15	5.00	23	8	31	333.25
1	6/4/2019	5:04:31 PM	16	5.00	27	6	33	238.51
1	6/4/2019	5:10:47 PM	17	5.00	24	7	31	245.59
1	6/4/2019	5:17:04 PM	18	5.00	21	6	27	316.08
1	6/4/2019	5:23:21 PM	19	5.00	25	6	31	202.50
1	6/4/2019	5:29:36 PM	20	5.00	22	7	29	277.66
1	6/4/2019	5:35:53 PM	21	5.00	22	5	27	224.45
1	6/4/2019	5:42:04 PM	22	5.00	21	4	25	161.61
1	6/4/2019	5:48:21 PM	23	5.00	21	5	27	263.41
1	6/4/2019	5:54:01 PM	24	5.00	14454	35770	50223	1014.28

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 6/4/19	Time 1501	Swipe Instrument Quantasart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>		LSC Printout Date 6/4/2019			

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP					
1	Background	100						
2	Elevator Entrance	120						
3	South Floor	140						
4	RAM Safe	1200						
5	East Lab Floor	130						
6	North Bench Top	110						
7	West Floor	140						
8	Work Platform	< 100						
9	Calorimeter	< 100						
10	Computers	120						
11	DSC/TGA	< 100						
12	South Bench Top	< 100						
Swipe # and Locations								
1-6	Background	16				DSC/TGA		
7	Elevator Entrance	17				South Bench Top		
8	South Floor							
9	RAM Safe							
10	East Lab Floor							
11	North Bench Top							
12	West Lab Floor							
13	Work Platform							
14	Calorimeter							
15	Computers							

Surveyed By: LD
 Date: 6/4/19

Swiped By: BR
 Date: 6/4/19

Reviewed By: *[Signature]*
 Date: 6/5/2019

Notes

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 6/4/2019	Time 1524	Swipe Instrument Quantasart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) <input type="checkbox"/> ON <input checked="" type="checkbox"/> OFF		LSC Printout Date 6/4/2019			

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP		
13	Backgroud	130		DODG 121 GUO LAB	Notes
14	121 Doorway	<100			
15	121 Center of lab	120			
16	121 Glove Box	300			
Swipe # and Locations					
18	121 Doorway		West Bench Top		
19	121 Glove Box-floor				
20	121 Glove Box				
21	West Bench Top				
22	121 Center of lab				
23	East Bench Top				
			Surveyed By: <i>CD</i>	Date: 6/4/19	
			Swiped By: <i>B T</i>	Date: 6/4/19	
			Reviewed By: <i>Airgum</i>	Date: 6/5/2019	

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	24	Dates:	6/9/2019-6/15/2019	Counted On	6/12/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		9.12E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	34	*****	*****	
2.	BACKGROUND 2	34	*****	*****	
3.	BACKGROUND 3	36	*****	*****	
4.	BACKGROUND 4	30	*****	*****	
5.	BACKGROUND 5	32	*****	*****	
6.	BACKGROUND 6	29	*****	*****	
7.	B021 ELEVATOR ENTRANCE	27	M	M	
8.	B021 FENCED GATE	30	M	M	
9.	B021 RAM SAFE	31	M	M	
10.	B021 EAST FLOOR	24	M	M	
11.	B021 NORTH BENCH TOP	30	M	M	
12.	B021 WEST LAB FLOOR	25	M	M	
13.	B021 WORK PLATFORM	27	M	M	
14.	B021 CALORIMETER	25	M	M	
15.	B021 COMPUTERS	28	M	M	
16.	B021 DSC/TGA	28	M	M	
17.	B021 SOUTH BENCH TOP	31	M	M	
18.	121 DOORWAY	25	M	M	
19.	121 GLOVE BOX- FLOOR	30	M	M	
20.	121 GLOVE BOX	27	M	M	
21.	121 WEST BENCH TOP	29	M	M	
22.	121 CENTER OF LAB	30	M	M	
23.	122 EAST BENCH TOP	30	M	M	
³⁶ Cl SOURCE 09/14/1971		50283	*****	*****	
		101.39%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: M. Heine			REVIEWED BY: C. Hines		
SIGN/DATE: <i>Madalisen Heine</i> 6/14/19			SIGN/DATE: <i>Chris Hines</i> 6/14/2019		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190612_1203

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190612_1203\20190612_1203.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	6/12/2019	12:04:57 PM	1	5.00	28	6	34	275.84		
1	6/12/2019	12:11:15 PM	2	5.00	26	8	34	263.36		
1	6/12/2019	12:17:31 PM	3	5.00	29	7	36	274.12		
1	6/12/2019	12:23:47 PM	4	5.00	23	7	30	333.43		
1	6/12/2019	12:30:02 PM	5	5.00	26	6	32	295.78		
1	6/12/2019	12:36:14 PM	6	5.00	24	5	29	302.39		
1	6/12/2019	12:42:27 PM	7	5.00	23	5	27	337.41		
1	6/12/2019	12:48:45 PM	8	5.00	24	6	30	271.97		
1	6/12/2019	12:55:02 PM	9	5.00	27	4	31	199.17		
1	6/12/2019	1:01:19 PM	10	5.00	21	4	24	154.02		
1	6/12/2019	1:07:36 PM	11	5.00	24	6	30	291.85		
1	6/12/2019	1:13:54 PM	12	5.00	20	5	25	301.48		

Protocol# 1 - Reactor Weekly.lsa

User: CSE


1	6/12/2019	1:20:17 PM	13	5.00	22	5	27	272.94
1	6/12/2019	1:26:36 PM	14	5.00	19	6	25	359.00
1	6/12/2019	1:32:47 PM	15	5.00	20	8	28	415.65
1	6/12/2019	1:39:03 PM	16	5.00	24	4	28	237.56
1	6/12/2019	1:45:17 PM	17	5.00	26	5	31	191.19
1	6/12/2019	1:51:32 PM	18	5.00	20	4	25	273.09
1	6/12/2019	1:57:51 PM	19	5.00	25	6	30	233.79
1	6/12/2019	2:04:05 PM	20	5.00	22	5	27	292.05
1	6/12/2019	2:10:23 PM	21	5.00	24	5	29	228.84
1	6/12/2019	2:16:40 PM	22	5.00	26	4	30	177.81
1	6/12/2019	2:22:57 PM	23	5.00	24	6	30	286.83
1	6/12/2019	2:28:37 PM	24	5.00	14396	35887	50283	1009.47

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 6/12/19	Time 1139	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) <input checked="" type="checkbox"/> ON <input type="checkbox"/> OFF <input type="checkbox"/>					

LSC Printout Date 6/12/19



Survey #	Location	CPM
1	Background	≤ 100
2	Elevator Entrance	≤ 100
3	South Floor	≤ 100
4	RAM Safe	1400
5	East Lab Floor	35-400 ≤ 100
6	North Bench Top	≤ 100
7	West Floor	≤ 100
8	Work Platform	≤ 100
9	Calorimeter	≤ 100
10	Computers	≤ 100
11	DSC/TGA	≤ 100
12	South Bench Top	≤ 100
13		
14		
15		

Swipe # and Locations	
1-6	Background 16 DSC/TGA
7	Elevator Entrance 17 South Bench Top
8	South Floor
9	RAM Safe
10	East Lab Floor
11	North Bench Top
12	West Lab Floor
13	Work Platform
14	Calorimeter
15	Computers

Surveyed By: <i>SS</i>	Date: <u>6/12/19</u>
Swiped By: <i>MJ</i>	Date: <u>6/12/19</u>
Reviewed By: <i>awg</i>	Date: <u>6/14/2019</u>

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	25	Dates:	6/16/2019-6/22/2019	Counted On	6/19/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.84E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	30	*****	*****	
2.	BACKGROUND 2	33	*****	*****	
3.	BACKGROUND 3	25	*****	*****	
4.	BACKGROUND 4	29	*****	*****	
5.	BACKGROUND 5	31	*****	*****	
6.	BACKGROUND 6	24	*****	*****	
7.	B021 ELEVATOR ENTRANCE	30	M	M	
8.	B021 FENCED GATE	30	M	M	
9.	B021 RAM SAFE	33	M	M	
10.	B021 EAST FLOOR	29	M	M	
11.	B021 NORTH BENCH TOP	29	M	M	
12.	B021 WEST LAB FLOOR	27	M	M	
13.	B021 WORK PLATFORM	29	M	M	
14.	B021 CALORIMETER	34	M	M	
15.	B021 COMPUTERS	26	M	M	
16.	B021 DSC/TGA	29	M	M	
17.	B021 SOUTH BENCH TOP	27	M	M	
18.	121 DOORWAY	31	M	M	
19.	121 GLOVE BOX- FLOOR	25	M	M	
20.	121 GLOVE BOX	25	M	M	
21.	121 WEST BENCH TOP	25	M	M	
22.	121 CENTER OF LAB	31	M	M	
23.	122 EAST BENCH TOP	29	M	M	
³⁶ CI SOURCE 09/14/1971		50134	*****	*****	
		101.09%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: M. Heine			REVIEWED BY: <i>C. Heine</i>		
SIGN/DATE: <i>Maddison Heine 6/20/19</i>			SIGN/DATE: <i>Augustine 6/20/2019</i>		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190619_0847

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190619_0847\20190619_0847.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	6/19/2019	8:48:46 AM	1	5.00		25	6	30	301.03	
1	6/19/2019	8:55:02 AM	2	5.00		26	6	33	239.30	
1	6/19/2019	9:01:20 AM	3	5.00		21	4	25	268.44	
1	6/19/2019	9:07:33 AM	4	5.00		22	7	29	358.90	
1	6/19/2019	9:13:46 AM	5	5.00		24	7	31	374.56	
1	6/19/2019	9:20:01 AM	6	5.00		17	6	24	340.85	
1	6/19/2019	9:26:15 AM	7	5.00		21	9	30	383.24	
1	6/19/2019	9:32:31 AM	8	5.00		23	7	30	268.30	
1	6/19/2019	9:38:48 AM	9	5.00		27	6	33	253.01	
1	6/19/2019	9:45:02 AM	10	5.00		23	6	29	258.31	
1	6/19/2019	9:51:21 AM	11	5.00		21	7	29	293.47	
1	6/19/2019	9:57:42 AM	12	5.00		22	4	27	215.86	

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	6/19/2019	10:04:06 AM	13	5.00	25	4	29	175.75
1	6/19/2019	10:10:25 AM	14	5.00	26	8	34	297.26
1	6/19/2019	10:16:35 AM	15	5.00	22	5	26	198.41
1	6/19/2019	10:22:49 AM	16	5.00	24	5	29	236.04
1	6/19/2019	10:29:05 AM	17	5.00	20	6	27	295.79
1	6/19/2019	10:35:21 AM	18	5.00	26	5	31	169.67
1	6/19/2019	10:41:38 AM	19	5.00	19	6	25	237.02
1	6/19/2019	10:47:52 AM	20	5.00	20	5	25	252.26
1	6/19/2019	10:54:10 AM	21	5.00	22	2	25	110.50
1	6/19/2019	11:00:29 AM	22	5.00	26	5	31	165.73
1	6/19/2019	11:06:47 AM	23	5.00	23	6	29	266.78
1	6/19/2019	11:12:27 AM	24	5.00	14364	35770	50134	1015.75


LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 6/19/19	Time 0822	Swipe Instrument Quantasart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>					

LSC Printout Date 6/19/19

NOTE SWIPE AND SURVEY LOCATIONS ON MAP



Survey #	Location	CPM
1	Background	≤ 100
2	Elevator Entrance	≤ 100
3	South Floor	≤ 100
4	RAM Safe	1400
5	East Lab Floor	≤ 100
6	North Bench Top	≤ 100
7	West Floor	≤ 100
8	Work Platform	≤ 100
9	Calorimeter	≤ 100
10	Computers	≤ 100
11	DSC/TGA	≤ 100
12	South Bench Top	≤ 100

Swipe # and Locations		Notes
1-6	Background	16 DSC/TGA
7	Elevator Entrance	17 South Bench Top
8	South Floor	
9	RAM Safe	
10	East Lab Floor	
11	North Bench Top	
12	West Lab Floor	
13	Work Platform	
14	Calorimeter	
15	Computers	

Surveyed By: MH	Date: 6/19/19
Swiped By: MH	Date: 6/19/19
Reviewed By: <i>[Signature]</i>	Date: 6/20/2019

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	26	Dates:	6/23/2019-6/29/2019	Counted On	6/25/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		9.27E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	37	*****	*****	
2.	BACKGROUND 2	37	*****	*****	
3.	BACKGROUND 3	36	*****	*****	
4.	BACKGROUND 4	33	*****	*****	
5.	BACKGROUND 5	35	*****	*****	
6.	BACKGROUND 6	30	*****	*****	
7.	B021 ELEVATOR ENTRANCE	35	M	M	
8.	B021 FENCED GATE	29	M	M	
9.	B021 RAM SAFE	31	M	M	
10.	B021 EAST FLOOR	29	M	M	
11.	B021 NORTH BENCH TOP	32	M	M	
12.	B021 WEST LAB FLOOR	28	M	M	
13.	B021 WORK PLATFORM	28	M	M	
14.	B021 CALORIMETER	31	M	M	
15.	B021 COMPUTERS	31	M	M	
16.	B021 DSC/TGA	27	M	M	
17.	B021 SOUTH BENCH TOP	27	M	M	
18.	121 DOORWAY	30	M	M	
19.	121 GLOVE BOX- FLOOR	31	M	M	
20.	121 GLOVE BOX	25	M	M	
21.	121 WEST BENCH TOP	33	M	M	
22.	121 CENTER OF LAB	29	M	M	
23.	122 EAST BENCH TOP	23	M	M	
³⁶ CI SOURCE 09/14/1971		49994	*****	*****	
		100.80%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: M. Heine			REVIEWED BY: C. Hines		
SIGN/DATE: <i>Maddison Heine</i> 6/26/19			SIGN/DATE: <i>Cliff Hines</i> 6/26/2019		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190625_1637

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190625_1637\20190625_1637.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	6/25/2019	4:38:08 PM	1	5.00	33	4	37	187.34		
1	6/25/2019	4:44:18 PM	2	5.00	29	9	37	316.46		
1	6/25/2019	4:50:37 PM	3	5.00	31	6	36	208.48		
1	6/25/2019	4:56:46 PM	4	5.00	26	7	33	304.83		
1	6/25/2019	5:03:02 PM	5	5.00	28	7	35	288.39		
1	6/25/2019	5:09:20 PM	6	5.00	23	6	30	275.37		
1	6/25/2019	5:15:39 PM	7	5.00	25	10	35	368.91		
1	6/25/2019	5:21:53 PM	8	5.00	22	8	29	314.32		
1	6/25/2019	5:28:11 PM	9	5.00	25	5	31	205.48		
1	6/25/2019	5:34:26 PM	10	5.00	24	5	29	184.00		
1	6/25/2019	5:40:42 PM	11	5.00	27	5	32	278.03		
1	6/25/2019	5:46:58 PM	12	5.00	23	5	28	183.14		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	6/25/2019	5:53:17 PM	13	5.00	23	5	28	232.95
1	6/25/2019	5:59:32 PM	14	5.00	25	5	31	245.64
1	6/25/2019	6:05:42 PM	15	5.00	25	6	31	262.06
1	6/25/2019	6:11:57 PM	16	5.00	24	3	27	177.09
1	6/25/2019	6:18:16 PM	17	5.00	22	5	27	232.89
1	6/25/2019	6:24:30 PM	18	5.00	23	6	30	268.32
1	6/25/2019	6:30:47 PM	19	5.00	25	6	31	208.31
1	6/25/2019	6:37:05 PM	20	5.00	20	5	25	299.77
1	6/25/2019	6:43:15 PM	21	5.00	26	7	33	237.39
1	6/25/2019	6:49:26 PM	22	5.00	24	5	29	183.91
1	6/25/2019	6:55:48 PM	23	5.00	18	5	23	231.29
1	6/25/2019	7:01:28 PM	24	5.00	14358	35636	49994	1013.28

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 6/25/19	Time 1615	Swipe Instrument Quantasmart	Serial No. 073396	Radiation Detected Beta/Gamma

Reactor Status (Check One)
 ON OFF

LSC Printout Date 6/25/2019

NOTE SWIPE AND SURVEY LOCATIONS ON MAP

DODG B0021 GUO LAB

The map shows the following locations and their corresponding CPM readings:

Survey #	Location	CPM
1	Background	≤ 100
2	Elevator Entrance	≤ 100
3	South Floor	≤ 100
4	RAM Safe	1225
5	East Lab Floor	≤ 100
6	North Bench Top	≤ 100
7	West Floor	≤ 100
8	Work Platform	≤ 100
9	Calorimeter	≤ 100
10	Computers	≤ 100
11	DSC/TGA	≤ 100
12	South Bench Top	≤ 100

Surveyed By: MH	Date: 6/25/19	Notes
Swiped By: MH	Date: 6/25/19	
Reviewed By: <i>[Signature]</i>	Date: 6/26/2019	

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 6/25/19	Time 1630	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>		LSC Printout Date 6/25/19			

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP		
13	Background	≤ 100	<p>DODG 121 GUO LAB</p> <p>Fume Hood</p> <p>East Bench Top</p> <p>West Bench Top</p> <p>Glove Box</p> <p>14, 15, 16, 18, 19, 20, 21, 22, 23</p> <p>← N</p>		
14	121 Doorway	≤ 100			
15	121 Center of lab	≤ 100			
16	121 Glove Box	≤ 100			
			Swipe # and Locations		
18	121 Doorway		<p>Surveyed By: MH Date: 6/25/19</p> <p>Swiped By: MH Date: 6/25/19</p> <p>Reviewed By: [Signature] Date: 6/26/2019</p>		
19	121 Glove Box-floor				
20	121 Glove Box				
21	West Bench Top				
22	121 Center of lab				
23	East Bench Top				

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	27	Dates:	6/30/2019-7/06/2019	Counted On	7/3/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.		uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100	
Limit of quantification (LOQ)=		9.01E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	25	*****	*****	
2.	BACKGROUND 2	29	*****	*****	
3.	BACKGROUND 3	41	*****	*****	
4.	BACKGROUND 4	27	*****	*****	
5.	BACKGROUND 5	35	*****	*****	
6.	BACKGROUND 6	30	*****	*****	
7.	B021 ELEVATOR ENTRANCE	33	M	M	
8.	B021 FENCED GATE	31	M	M	
9.	B021 RAM SAFE	25	M	M	
10.	B021 EAST FLOOR	32	M	M	
11.	B021 NORTH BENCH TOP	27	M	M	
12.	B021 WEST LAB FLOOR	30	M	M	
13.	B021 WORK PLATFORM	28	M	M	
14.	B021 CALORIMETER	34	M	M	
15.	B021 COMPUTERS	28	M	M	
16.	B021 DSC/TGA	27	M	M	
17.	B021 SOUTH BENCH TOP	29	M	M	
18.	121 DOORWAY	28	M	M	
19.	121 GLOVE BOX- FLOOR	25	M	M	
20.	121 GLOVE BOX	29	M	M	
21.	121 WEST BENCH TOP	32	M	M	
22.	121 CENTER OF LAB	35	M	M	
23.	122 EAST BENCH TOP	30	M	M	
³⁶ Cl SOURCE 09/14/1971		50127	*****	*****	
		101.07%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: M. Heine			REVIEWED BY: C. Hines		
SIGN/DATE: <i>Maddison Heine</i> 7/5/19			SIGN/DATE: <i>C. Hines</i> 7/5/2019		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190703_1634

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190703_1634\20190703_1634.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	7/3/2019	4:35:56 PM	1	5.00		19	6	25	360.79	
1	7/3/2019	4:42:05 PM	2	5.00		24	5	29	239.99	
1	7/3/2019	4:48:22 PM	3	5.00		33	8	41	311.18	
1	7/3/2019	4:54:32 PM	4	5.00		22	5	27	337.89	
1	7/3/2019	5:00:48 PM	5	5.00		28	7	35	316.65	
1	7/3/2019	5:07:04 PM	6	5.00		24	6	30	304.20	
1	7/3/2019	5:13:23 PM	7	5.00		28	5	33	216.81	
1	7/3/2019	5:19:40 PM	8	5.00		25	7	31	205.93	
1	7/3/2019	5:25:58 PM	9	5.00		20	5	25	223.81	
1	7/3/2019	5:32:15 PM	10	5.00		26	6	32	204.43	
1	7/3/2019	5:38:30 PM	11	5.00		23	4	27	177.63	
1	7/3/2019	5:44:47 PM	12	5.00		24	7	30	242.88	

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	7/3/2019	5:51:06 PM	13	5.00	23	5	28	341.07
1	7/3/2019	5:57:22 PM	14	5.00	28	7	34	216.59
1	7/3/2019	6:03:29 PM	15	5.00	23	5	28	243.60
1	7/3/2019	6:09:43 PM	16	5.00	21	6	27	285.70
1	7/3/2019	6:16:00 PM	17	5.00	23	6	29	374.69
1	7/3/2019	6:22:16 PM	18	5.00	24	4	28	217.41
1	7/3/2019	6:28:32 PM	19	5.00	20	5	25	214.43
1	7/3/2019	6:34:51 PM	20	5.00	22	7	29	284.69
1	7/3/2019	6:40:58 PM	21	5.00	25	7	32	267.62
1	7/3/2019	6:47:09 PM	22	5.00	28	7	35	190.23
1	7/3/2019	6:53:28 PM	23	5.00	25	5	30	260.74
1	7/3/2019	6:59:08 PM	24	5.00	14531	35597	50127	1011.19


LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 7/3/19	Time 16:14	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One)					
ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>					

LSC Printout Date **7/3/2019**

NOTE SWIPE AND SURVEY LOCATIONS ON MAP



Survey #	Location	CPM	Swipe # and Locations	Notes
1	Background	150		
2	Elevator Entrance	130		
3	South Floor	≤ 100		
4	RAM Safe	950		
5	East Lab Floor	≤ 100		
6	North Bench Top	≤ 100		
7	West Floor	110		
8	Work Platform	≤ 100		
9	Calorimeter	≤ 100		
10	Computers	≤ 100		
11	DSC/TGA	≤ 100		
12	South Bench Top	≤ 100		
13	Work Platform			
14	Calorimeter			
15	Computers			
16	Background		16 DSC/TGA	
17	Elevator Entrance		17 South Bench Top	
18	South Floor			
19	RAM Safe			
20	East Lab Floor			
21	North Bench Top			
22	West Lab Floor			
23	Work Platform			
24	Calorimeter			
25	Computers			

Surveyed By: *MH* Date: *7/3/19*

Swiped By: *AH* Date: *7/3/19*

Reviewed By: *[Signature]* Date: *7/5/2019*

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 7/3/19	Time 1424	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>		LSC Printout Date 7/3/2019			


Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP		
13	Background	≤ 100	<p>DODG 121 GUO LAB</p>		
14	121 Doorway	≤ 100			
15	121 Center of lab	150			
16	121 Glove Box	110			
Swipe # and Locations					
18	121 Doorway				
19	121 Glove Box-floor				
20	121 Glove Box				
21	West Bench Top				
22	121 Center of lab				
23	East Bench Top				

Surveyed By: MH	Date: 7/3/19	Notes
Swiped By: PAA	Date: 7/3/19	
Reviewed By: <i>[Signature]</i>	Date: 7/3/19	

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB B0021 WEEKLY SWIPES & SURVEYS

Revised 2/20/2019

Week #	28	Dates:	7/07/2019-7/13/2019	Counted On	7/9/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm2 = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.87E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	27	*****	*****	
2.	BACKGROUND 2	30	*****	*****	
3.	BACKGROUND 3	28	*****	*****	
4.	BACKGROUND 4	28	*****	*****	
5.	BACKGROUND 5	32	*****	*****	
6.	BACKGROUND 6	29	*****	*****	
7.	B021 ELEVATOR ENTRANCE	28	M	M	
8.	B021 FENCED GATE	30	M	M	
9.	B021 RAM SAFE	27	M	M	
10.	B021 EAST FLOOR	27	M	M	
11.	B021 NORTH BENCH TOP	25	M	M	
12.	B021 WEST LAB FLOOR	29	M	M	
13.	B021 WORK PLATFORM	30	M	M	
14.	B021 CALORIMETER	28	M	M	
15.	B021 COMPUTERS	31	M	M	
16.	B021 DSC/TGA	26	M	M	
17.	B021 SOUTH BENCH TOP	29	M	M	
18.	121 DOORWAY	25	M	M	
19.	121 GLOVE BOX- FLOOR	29	M	M	
20.	121 GLOVE BOX	26	M	M	
21.	121 WEST BENCH TOP	30	M	M	
22.	121 CENTER OF LAB	31	M	M	
23.	122 EAST BENCH TOP	25	M	M	
³⁶ Cl SOURCE 09/14/1971		50097	*****	*****	
		101.01%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: C. Hines		
SIGN/DATE:  7/9/2019			SIGN/DATE:  7/9/2019		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190709_1052

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190709_1052\20190709_1052.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	7/9/2019	10:53:40 AM	1	5.00		21	6	27	237.80	
1	7/9/2019	10:59:49 AM	2	5.00		24	6	30	336.03	
1	7/9/2019	11:06:07 AM	3	5.00		23	5	28	223.12	
1	7/9/2019	11:12:16 AM	4	5.00		25	3	28	202.83	
1	7/9/2019	11:18:30 AM	5	5.00		28	4	32	217.25	
1	7/9/2019	11:24:48 AM	6	5.00		24	5	29	202.31	
1	7/9/2019	11:31:08 AM	7	5.00		23	5	28	162.68	
1	7/9/2019	11:37:23 AM	8	5.00		26	4	30	133.99	
1	7/9/2019	11:43:41 AM	9	5.00		22	5	27	231.50	
1	7/9/2019	11:49:56 AM	10	5.00		21	6	27	218.35	
1	7/9/2019	11:56:13 AM	11	5.00		21	5	25	204.51	
1	7/9/2019	12:02:27 PM	12	5.00		24	5	29	167.66	

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	7/9/2019	12:08:46	PM	13	5.00	26	4	30	211.00
1	7/9/2019	12:14:58	PM	14	5.00	23	6	28	236.60
1	7/9/2019	12:21:07	PM	15	5.00	25	6	31	224.34
1	7/9/2019	12:27:22	PM	16	5.00	22	4	26	229.65
1	7/9/2019	12:33:41	PM	17	5.00	24	5	29	216.20
1	7/9/2019	12:39:57	PM	18	5.00	19	6	25	264.77
1	7/9/2019	12:46:13	PM	19	5.00	23	6	29	294.05
1	7/9/2019	12:52:34	PM	20	5.00	20	6	26	305.42
1	7/9/2019	12:58:43	PM	21	5.00	24	6	30	273.12
1	7/9/2019	1:04:53	PM	22	5.00	27	5	31	175.26
1	7/9/2019	1:11:13	PM	23	5.00	20	4	25	220.98
1	7/9/2019	1:16:53	PM	24	5.00	14501	35596	50097	1012.12

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 7/9/2019	Time 0903	Swipe Instrument Quantasart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>		LSC Printout Date 7/9/2019			

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP		
1	Background	80			
2	Elevator Entrance	86			
3	South Floor	120			
4	RAM Safe	1660			
5	East Lab Floor	100			
6	North Bench Top	120			
7	West Floor	100			
8	Work Platform	60			
9	Calorimeter	80			
10	Computers	160			
11	DSC/TGA	120			
12	South Bench Top	160			
13	South Floor				
14	RAM Safe				
15	East Lab Floor				
16	Background		Swipe # and Locations		
17	Elevator Entrance		16 DSC/TGA		
18	South Floor		17 South Bench Top		
19	RAM Safe		2 Elevator Entrance		
20	East Lab Floor		7		
21	North Bench Top		Notes		
22	West Lab Floor		Surveyed By: BT Date: 7/9/19		
23	Work Platform		Swiped By: BT Date: 7/9/19		
24	Calorimeter		Reviewed By: <i>anyone</i> Date: 7/9/2019		
25	Computers				

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.


LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 7/9/2019	Time 0927	Swipe Instrument Quantasart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>					

LSC Printout Date 7/9/2019

NOTE SWIPE AND SURVEY LOCATIONS ON MAP




Surveyed By: BT	Date: 7/9/19	Notes
Swiped By: BT	Date: 7/9/19	
Reviewed By: <i>[Signature]</i>	Date: 7/9/2019	

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	29	Dates:	7/14/2019-7/20/2019	Counted On	7/18/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.		uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100	
Limit of quantification (LOQ)=		8.75E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	25	*****	*****	
2.	BACKGROUND 2	27	*****	*****	
3.	BACKGROUND 3	27	*****	*****	
4.	BACKGROUND 4	25	*****	*****	
5.	BACKGROUND 5	34	*****	*****	
6.	BACKGROUND 6	27	*****	*****	
7.	B021 ELEVATOR ENTRANCE	25	M	M	
8.	B021 FENCED GATE	25	M	M	
9.	B021 RAM SAFE	24	M	M	
10.	B021 EAST FLOOR	24	M	M	
11.	B021 NORTH BENCH TOP	24	M	M	
12.	B021 WEST LAB FLOOR	26	M	M	
13.	B021 WORK PLATFORM	28	M	M	
14.	B021 CALORIMETER	43	2.17E-05	2.17E-07	
15.	B021 COMPUTERS	27	M	M	
16.	B021 DSC/TGA	27	M	M	
17.	B021 SOUTH BENCH TOP	27	M	M	
18.	121 DOORWAY	31	M	M	
19.	121 GLOVE BOX- FLOOR	30	M	M	
20.	121 GLOVE BOX	29	M	M	
21.	121 WEST BENCH TOP	31	M	M	
22.	121 CENTER OF LAB	40	1.75E-05	1.75E-07	
23.	122 EAST BENCH TOP	30	M	M	
³⁶ Cl SOURCE 09/14/1971		50301		*****	*****
		101.42%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: C. Hines		
SIGN/DATE:  / 7/19/19			SIGN/DATE: 		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190718_1507

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190718_1507\20190718_1507.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	7/18/2019	3:09:55 PM	1		5.00	21	4	25	187.19	
1	7/18/2019	3:16:10 PM	2		5.00	23	4	27	279.36	
1	7/18/2019	3:22:29 PM	3		5.00	21	6	27	329.72	
1	7/18/2019	3:30:13 PM	4		5.00	22	3	25	121.23	
1	7/18/2019	3:36:27 PM	5		5.00	27	7	34	289.56	
1	7/18/2019	3:42:47 PM	6		5.00	23	4	27	184.83	
1	7/18/2019	3:49:06 PM	7		5.00	21	4	25	268.10	
1	7/18/2019	3:55:26 PM	8		5.00	20	5	25	246.01	
1	7/18/2019	4:02:01 PM	9		5.00	22	2	24	133.18	
1	7/18/2019	4:08:20 PM	10		5.00	19	5	24	274.19	
1	7/18/2019	4:17:52 PM	11		5.00	22	3	24	100.69	
1	7/18/2019	4:24:06 PM	12		5.00	19	7	26	325.52	

1	7/18/2019	4:30:26 PM	13	5.00	23	5	28	284.56
1	7/18/2019	4:36:39 PM	14	5.00	29	14	43	414.16
1	7/18/2019	4:43:17 PM	15	5.00	24	3	27	133.00
1	7/18/2019	4:50:00 PM	16	5.00	23	4	27	166.47
1	7/18/2019	4:56:39 PM	17	5.00	25	2	27	168.03
1	7/18/2019	5:02:56 PM	18	5.00	25	6	31	294.14
1	7/18/2019	5:09:13 PM	19	5.00	23	7	30	295.82
1	7/18/2019	5:15:26 PM	20	5.00	24	5	29	208.29
1	7/18/2019	5:21:38 PM	21	5.00	25	6	31	291.65
1	7/18/2019	5:28:09 PM	22	5.00	36	4	40	165.81
1	7/18/2019	5:34:48 PM	23	5.00	27	3	30	237.66
1	7/18/2019	5:40:29 PM	24	5.00	14374	35927	50301	1025.33

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building	Dodgen Research Facility	Room	B0021 - GUO	Lab Class	D	Survey Instrument	GM Detector Model 12	Serial No.	172178	Radiation Detected	Gamma
Authorized User	Xiaofeng Guo	Date	7/18/2019	Time	1118	Swipe Instrument	Quantasart	Serial No.	073396	Radiation Detected	Beta/Gamma
Reactor Status (Check One)	ON <input checked="" type="checkbox"/> OFF <input type="checkbox"/>										

LSC Printout Date 7/18/2019

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP			
1	Background	60				
2	Elevator Entrance	80				
3	South Floor	60				
4	RAM Safe	100				
5	East Lab Floor	60				
6	North Bench Top	100				
7	West Floor	80				
8	Work Platform	100				
9	Calorimeter	60				
10	Computers	60				
11	DSC/TGA	80				
12	South Bench Top	60				
Swipe # and Locations						
1-6	Background	16			DSC/TGA	
7	Elevator Entrance	17			South Bench Top	
8	South Floor					
9	RAM Safe					
10	East Lab Floor					
11	North Bench Top					
12	West Lab Floor					
13	Work Platform					
14,	Calorimeter					
15	Computers					



Surveyed By: BT Date: 7/18/2019
 Swiped By: BT Date: 7/18/2019
 Reviewed By: [Signature] Date: 7/29/2019

Notes

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	30	Dates:	7/21/2019-7/27/2019	Counted On	7/24/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.		uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100	
Limit of quantification (LOQ)=		8.77E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDÉ 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	28	*****	*****	
2.	BACKGROUND 2	28	*****	*****	
3.	BACKGROUND 3	29	*****	*****	
4.	BACKGROUND 4	27	*****	*****	
5.	BACKGROUND 5	26	*****	*****	
6.	BACKGROUND 6	28	*****	*****	
7.	B021 ELEVATOR ENTRANCE	32	M	M	
8.	B021 FENCED GATE	31	M	M	
9.	B021 RAM SAFE	28	M	M	
10.	B021 EAST FLOOR	28	M	M	
11.	B021 NORTH BENCH TOP	46	2.57E-05	2.57E-07	
12.	B021 WEST LAB FLOOR	28	M	M	
13.	B021 WORK PLATFORM	29	M	M	
14.	B021 CALORIMETER	50	3.13E-05	3.13E-07	
15.	B021 COMPUTERS	25	M	M	
16.	B021 DSC/TGA	26	M	M	
17.	B021 SOUTH BENCH TOP	31	M	M	
18.	121 DOORWAY	33	M	M	
19.	121 GLOVE BOX- FLOOR	32	M	M	
20.	121 GLOVE BOX	29	M	M	
21.	121 WEST BENCH TOP	29	M	M	
22.	121 CENTER OF LAB	30	M	M	
23.	122 EAST BENCH TOP	28	M	M	
³⁶ CI SOURCE 09/14/1971		50091		*****	*****
		101.00%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: C. Hines		
SIGN/DATE:  / 7/25/19			SIGN/DATE:  7/29/2019		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190724_1550

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190724_1550\20190724_1550.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	7/24/2019	3:51:43 PM	1	5.00		24	4	28	220.57	
1	7/24/2019	3:57:53 PM	2	5.00		21	7	28	344.37	
1	7/24/2019	4:04:11 PM	3	5.00		22	7	29	293.81	
1	7/24/2019	4:10:21 PM	4	5.00		24	4	27	190.91	
1	7/24/2019	4:16:35 PM	5	5.00		22	4	26	212.12	
1	7/24/2019	4:22:52 PM	6	5.00		25	4	28	157.18	
1	7/24/2019	4:29:14 PM	7	5.00		26	6	32	256.25	
1	7/24/2019	4:35:31 PM	8	5.00		25	6	31	228.91	
1	7/24/2019	4:41:53 PM	9	5.00		23	5	28	150.19	
1	7/24/2019	4:48:13 PM	10	5.00		24	4	28	162.04	
1	7/24/2019	4:54:24 PM	11	5.00		35	11	46	243.53	
1	7/24/2019	5:00:42 PM	12	5.00		22	6	28	221.23	

Protocol# 1 - Reactor Weekly.lsa

User: CSE


1	7/24/2019	5:07:01 PM	13	5.00	23	6	29	147.03
1	7/24/2019	5:13:22 PM	14	5.00	37	13	50	272.79
1	7/24/2019	5:19:44 PM	15	5.00	19	6	25	238.63
1	7/24/2019	5:25:59 PM	16	5.00	20	6	26	336.43
1	7/24/2019	5:32:18 PM	17	5.00	24	7	31	326.40
1	7/24/2019	5:38:35 PM	18	5.00	27	6	33	178.56
1	7/24/2019	5:44:49 PM	19	5.00	28	4	32	101.98
1	7/24/2019	5:51:07 PM	20	5.00	24	4	29	207.08
1	7/24/2019	5:57:18 PM	21	5.00	23	6	29	212.55
1	7/24/2019	6:03:29 PM	22	5.00	24	6	30	234.87
1	7/24/2019	6:09:49 PM	23	5.00	22	6	28	229.78
1	7/24/2019	6:15:29 PM	24	5.00	14445	35646	50091	1025.06

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 7/24/19	Time 1446	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) <input checked="" type="checkbox"/> ON <input type="checkbox"/> OFF		<div style="text-align: right;">LSC Printout Date: 7/24/2019</div>			

Survey #	Location	CPM	Notes
1	Background	80	
2	Elevator Entrance	100	
3	South Floor	80	
4	RAM Safe	1100	
5	East Lab Floor	100	
6	North Bench Top	100	
7	West Floor	60	
8	Work Platform	80	
9	Calorimeter	120	
10	Computers	100	
11	DSC/TGA	80	
12	South Bench Top	60	
Swipe # and Locations			
1-6	Background	16	DSC/TGA
7	Elevator Entrance	17	South Bench Top
8	South Floor		
9	RAM Safe		
10	East Lab Floor		
11	North Bench Top		
12	West Lab Floor		
13	Work Platform		
14	Calorimeter		
15	Computers		



Surveyed By: BT Date: 7/24/19

Swiped By: CS Date: 7/24/19

Reviewed By: [Signature] Date: 7/24/2019

Elevator Entrance

Notes

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building	Dodgen Research Facility	Room	121- GUO	Lab Class	D	Survey Instrument	GM Detector Model 12	Serial No.	172178	Radiation Detected	Gamma
Authorized User	Xiaofeng Guo	Date	7/24/19	Time	1456	Swipe Instrument	Quantasart	Serial No.	073396	Radiation Detected	Beta/Gamma
Reactor Status (Check One)	<input checked="" type="checkbox"/> ON <input type="checkbox"/> OFF										


LSC Printout Date 7/24/2019

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP	
13	Background	80	<p style="text-align: center;">DODG 121 GUO LAB</p> <p style="text-align: center;">← N</p> <p style="text-align: center;">(22) 15</p> <p style="text-align: center;">(23) East Bench Top</p> <p style="text-align: center;">(20) Glove Box</p> <p style="text-align: center;">(19) 16</p> <p style="text-align: center;">(18) 14</p> <p style="text-align: center;">(21) West Bench Top</p> <p>Notes</p>	
14	121 Doorway	60		
15	121 Center of lab	80		
16	121 Glove Box	120		
Swipe # and Locations				
18	121 Doorway			
19	121 Glove Box-Floor			
20	121 Glove Box			
21	West Bench Top			
22	121 Center of lab			
23	East Bench Top			
Surveyed By: <i>BS</i> Date: 7/24/19				
Swiped By: <i>BS</i> Date: 7/24/19				
Reviewed By: <i>Carson</i> Date: 7/23/19				

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	31	Dates:	7/28/2019-8/03/2019	Counted On	7/29/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.94E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	32	*****	*****	
2.	BACKGROUND 2	31	*****	*****	
3.	BACKGROUND 3	28	*****	*****	
4.	BACKGROUND 4	28	*****	*****	
5.	BACKGROUND 5	31	*****	*****	
6.	BACKGROUND 6	30	*****	*****	
7.	B021 ELEVATOR ENTRANCE	27	M	M	
8.	B021 FENCED GATE	27	M	M	
9.	B021 RAM SAFE	31	M	M	
10.	B021 EAST FLOOR	24	M	M	
11.	B021 NORTH BENCH TOP	29	M	M	
12.	B021 WEST LAB FLOOR	28	M	M	
13.	B021 WORK PLATFORM	27	M	M	
14.	B021 CALORIMETER	23	M	M	
15.	B021 COMPUTERS	26	M	M	
16.	B021 DSC/TGA	25	M	M	
17.	B021 SOUTH BENCH TOP	24	M	M	
18.	121 DOORWAY	28	M	M	
19.	121 GLOVE BOX- FLOOR	28	M	M	
20.	121 GLOVE BOX	24	M	M	
21.	121 WEST BENCH TOP	26	M	M	
22.	121 CENTER OF LAB	27	M	M	
23.	122 EAST BENCH TOP	33	M	M	
³⁶ CI SOURCE 09/14/1971		50322		*****	*****
		101.46%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: C. Hines		
SIGN/DATE:  7/30/2019			SIGN/DATE:  7/30/2019		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190729_1644

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190729_1644\20190729_1644.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	7/29/2019	4:45:08 PM	1	5.00	26	6	32	279.89		
1	7/29/2019	4:51:19 PM	2	5.00	25	6	31	243.41		
1	7/29/2019	4:57:36 PM	3	5.00	23	5	28	293.92		
1	7/29/2019	5:03:46 PM	4	5.00	23	5	28	261.07		
1	7/29/2019	5:10:01 PM	5	5.00	27	5	31	205.10		
1	7/29/2019	5:16:19 PM	6	5.00	23	6	30	269.04		
1	7/29/2019	5:22:38 PM	7	5.00	21	6	27	300.45		
1	7/29/2019	5:28:51 PM	8	5.00	20	7	27	320.67		
1	7/29/2019	5:35:09 PM	9	5.00	27	4	31	77.13		
1	7/29/2019	5:41:28 PM	10	5.00	19	5	24	264.53		
1	7/29/2019	5:47:39 PM	11	5.00	23	6	29	285.31		
1	7/29/2019	5:53:54 PM	12	5.00	23	5	28	189.47		

1	7/29/2019	6:00:14 PM	13	5.00	22	5	27	214.80
1	7/29/2019	6:06:30 PM	14	5.00	19	5	23	300.07
1	7/29/2019	6:12:38 PM	15	5.00	20	6	26	245.00
1	7/29/2019	6:18:52 PM	16	5.00	18	7	25	323.95
1	7/29/2019	6:25:07 PM	17	5.00	20	4	24	264.73
1	7/29/2019	6:31:17 PM	18	5.00	22	6	28	258.29
1	7/29/2019	6:37:33 PM	19	5.00	25	3	28	41.85
1	7/29/2019	6:43:52 PM	20	5.00	18	5	24	268.93
1	7/29/2019	6:50:08 PM	21	5.00	22	4	26	238.95
1	7/29/2019	6:56:27 PM	22	5.00	22	5	27	243.72
1	7/29/2019	7:02:36 PM	23	5.00	26	7	33	304.08
1	7/29/2019	7:08:17 PM	24	5.00	14387	35935	50322	1015.06

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building	Dodgen Research Facility	Room	B0021 - GUO	Lab Class	D	Survey Instrument	GM Detector Model 12	Serial No.	172178	Radiation Detected	Gamma
Authorized User	Xiaofeng Guo	Date	7/29/19	Time	1616	Swipe Instrument	Quantasart	Serial No.	073396	Radiation Detected	Beta/Gamma
Reactor Status (Check One)	<input checked="" type="checkbox"/> ON <input type="checkbox"/> OFF										

LSC Printout Date 7/29/2019

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP	
1	Background	80		
2	Elevator Entrance	120		
3	South Floor	100		
4	RAM Safe	1000		
5	East Lab Floor	80		
6	North Bench Top	120		
7	West Floor	100		
8	Work Platform	120		
9	Calorimeter	100		
10	Computers	100		
11	DSC/TGA	80		
12	South Bench Top	80		
Swipe # and Locations				
1-6	Background	16		DSC/TGA
7	Elevator Entrance	17		South Bench Top
8	South Floor			
9	RAM Safe			
10	East Lab Floor			
11	North Bench Top			
12	West Lab Floor			
13	Work Platform			
14	Calorimeter			
15	Computers			

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 7/29/2019	Time 1637	Swipe Instrument Quantasart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) <input checked="" type="checkbox"/> ON <input type="checkbox"/> OFF					


LSC Printout Date 7/29/2019

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP		
13	Backgroud	180			
14	121 Doorway	100			
15	121 Center of lab	80			
16	121 Glove Box	140			
Swipe # and Locations					
18	121 Doorway		Notes		
19	121 Glove Box-floor				
20	121 Glove Box				
21	West Bench Top				
22	121 Center of lab				
23	East Bench Top				

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	32	Dates:	8/04/2019-8/10/2019	Counted On	8/7/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm2 = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		9.10E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	35	*****	*****	
2.	BACKGROUND 2	33	*****	*****	
3.	BACKGROUND 3	28	*****	*****	
4.	BACKGROUND 4	32	*****	*****	
5.	BACKGROUND 5	33	*****	*****	
6.	BACKGROUND 6	32	*****	*****	
7.	B021 ELEVATOR ENTRANCE	30	M	M	
8.	B021 FENCED GATE	28	M	M	
9.	B021 RAM SAFE	31	M	M	
10.	B021 EAST FLOOR	29	M	M	
11.	B021 NORTH BENCH TOP	29	M	M	
12.	B021 WEST LAB FLOOR	29	M	M	
13.	B021 WORK PLATFORM	30	M	M	
14.	B021 CALORIMETER	28	M	M	
15.	B021 COMPUTERS	36	M	M	
16.	B021 DSC/TGA	27	M	M	
17.	B021 SOUTH BENCH TOP	27	M	M	
18.	121 DOORWAY	25	M	M	
19.	121 GLOVE BOX- FLOOR	32	M	M	
20.	121 GLOVE BOX	32	M	M	
21.	121 WEST BENCH TOP	25	M	M	
22.	121 CENTER OF LAB	31	M	M	
23.	122 EAST BENCH TOP	29	M	M	
³⁶ CI SOURCE 09/14/1971		50043	*****	*****	
		100.90%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE:  8/28/2019			SIGN/DATE:  8/28/19		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190807_1414

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190807_1414\20190807_1414.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	8/7/2019	2:15:10 PM	1	5.00	28	7	35	248.67	
1	8/7/2019	2:21:19 PM	2	5.00	25	8	33	395.50	
1	8/7/2019	2:27:37 PM	3	5.00	23	5	28	265.74	
1	8/7/2019	2:33:46 PM	4	5.00	26	6	32	376.72	
1	8/7/2019	2:40:01 PM	5	5.00	27	5	33	220.45	
1	8/7/2019	2:46:18 PM	6	5.00	26	6	32	233.29	
1	8/7/2019	2:52:38 PM	7	5.00	25	5	30	197.57	
1	8/7/2019	2:58:54 PM	8	5.00	23	5	28	217.62	
1	8/7/2019	3:05:12 PM	9	5.00	26	5	31	124.34	
1	8/7/2019	3:11:31 PM	10	5.00	24	5	29	182.67	
1	8/7/2019	3:17:45 PM	11	5.00	23	6	29	323.50	
1	8/7/2019	3:24:03 PM	12	5.00	24	5	29	141.80	

1.	8/7/2019	3:30:23 PM	13	5.00	25	5	30	231.12
1	8/7/2019	3:36:38 PM	14	5.00	22	6	28	275.52
1	8/7/2019	3:42:46 PM	15	5.00	30	6	36	207.40
1	8/7/2019	3:48:59 PM	16	5.00	22	5	27	238.11
1	8/7/2019	3:55:15 PM	17	5.00	22	5	27	251.84
1	8/7/2019	4:01:27 PM	18	5.00	20	5	25	192.60
1	8/7/2019	4:07:48 PM	19	5.00	27	5	32	127.26
1	8/7/2019	4:14:04 PM	20	5.00	24	7	32	264.84
1	8/7/2019	4:20:23 PM	21	5.00	21	4	25	250.77
1	8/7/2019	4:26:42 PM	22	5.00	25	6	31	119.36
1	8/7/2019	4:32:55 PM	23	5.00	24	5	29	223.52
1	8/7/2019	4:38:35 PM	24	5.00	14277	35766	50043	1014.88

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	33	Dates:	8/11/2019-8/17/2019	Counted On	8/14/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm2 = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.89E-08 uCi/cm ²	QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK		
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	32	*****	*****	
2.	BACKGROUND 2	28	*****	*****	
3.	BACKGROUND 3	30	*****	*****	
4.	BACKGROUND 4	26	*****	*****	
5.	BACKGROUND 5	27	*****	*****	
6.	BACKGROUND 6	33	*****	*****	
7.	B021 ELEVATOR ENTRANCE	31	M	M	
8.	B021 FENCED GATE	32	M	M	
9.	B021 RAM SAFE	28	M	M	
10.	B021 EAST FLOOR	24	M	M	
11.	B021 NORTH BENCH TOP	32	M	M	
12.	B021 WEST LAB FLOOR	28	M	M	
13.	B021 WORK PLATFORM	28	M	M	
14.	B021 CALORIMETER	30	M	M	
15.	B021 COMPUTERS	32	M	M	
16.	B021 DSC/TGA	29	M	M	
17.	B021 SOUTH BENCH TOP	26	M	M	
18.	121 DOORWAY	29	M	M	
19.	121 GLOVE BOX- FLOOR	32	M	M	
20.	121 GLOVE BOX	24	M	M	
21.	121 WEST BENCH TOP	23	M	M	
22.	121 CENTER OF LAB	24	M	M	
23.	122 EAST BENCH TOP	23	M	M	
³⁶ CI SOURCE 09/14/1971		50038	*****	*****	
		100.89%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE:  8/15/2019			SIGN/DATE:  9/11/19		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190814_1642

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190814_1642\20190814_1642.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	8/14/2019	4:43:37 PM	1	5.00	25	6	32	318.65		
1	8/14/2019	4:49:46 PM	2	5.00	23	5	28	258.66		
1	8/14/2019	4:56:03 PM	3	5.00	24	6	30	293.20		
1	8/14/2019	5:02:13 PM	4	5.00	21	6	26	312.48		
1	8/14/2019	5:08:28 PM	5	5.00	22	4	27	227.13		
1	8/14/2019	5:14:45 PM	6	5.00	27	6	33	272.16		
1	8/14/2019	5:21:04 PM	7	5.00	25	6	31	269.46		
1	8/14/2019	5:27:18 PM	8	5.00	24	8	32	288.87		
1	8/14/2019	5:33:35 PM	9	5.00	22	6	28	251.52		
1	8/14/2019	5:39:48 PM	10	5.00	19	5	24	266.73		
1	8/14/2019	5:46:03 PM	11	5.00	23	9	32	361.15		
1	8/14/2019	5:52:19 PM	12	5.00	20	7	28	302.31		



Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	8/14/2019	5:58:39 PM	13	5.00	21	6	28	288.08
1	8/14/2019	6:04:54 PM	14	5.00	24	6	30	278.49
1	8/14/2019	6:11:03 PM	15	5.00	26	6	32	250.24
1	8/14/2019	6:17:16 PM	16	5.00	25	4	29	173.92
1	8/14/2019	6:23:33 PM	17	5.00	21	5	26	266.22
1	8/14/2019	6:29:44 PM	18	5.00	24	6	29	252.10
1	8/14/2019	6:36:02 PM	19	5.00	25	7	32	308.72
1	8/14/2019	6:42:17 PM	20	5.00	21	4	24	235.30
1	8/14/2019	6:48:34 PM	21	5.00	18	5	23	297.82
1	8/14/2019	6:54:53 PM	22	5.00	20	5	24	179.23
1	8/14/2019	7:01:03 PM	23	5.00	19	4	23	239.14
1	8/14/2019	7:06:43 PM	24	5.00	14369	35669	50038	1022.65

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	34	Dates:	8/18/2019-8/24/2019	Counted On	8/20/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.90E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	26	*****	*****	
2.	BACKGROUND 2	28	*****	*****	
3.	BACKGROUND 3	28	*****	*****	
4.	BACKGROUND 4	32	*****	*****	
5.	BACKGROUND 5	32	*****	*****	
6.	BACKGROUND 6	31	*****	*****	
7.	B021 ELEVATOR ENTRANCE	31	M	M	
8.	B021 FENCED GATE	28	M	M	
9.	B021 RAM SAFE	28	M	M	
10.	B021 EAST FLOOR	24	M	M	
11.	B021 NORTH BENCH TOP	26	M	M	
12.	B021 WEST LAB FLOOR	30	M	M	
13.	B021 WORK PLATFORM	25	M	M	
14.	B021 CALORIMETER	29	M	M	
15.	B021 COMPUTERS	26	M	M	
16.	B021 DSC/TGA	31	M	M	
17.	B021 SOUTH BENCH TOP	23	M	M	
18.	121 DOORWAY	30	M	M	
19.	121 GLOVE BOX- FLOOR	30	M	M	
20.	121 GLOVE BOX	26	M	M	
21.	121 WEST BENCH TOP	23	M	M	
22.	121 CENTER OF LAB	28	M	M	
23.	122 EAST BENCH TOP	32	M	M	
³⁶ CI SOURCE 09/14/1971		49962	*****	*****	
		100.74%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE:  8/20/2019			SIGN/DATE:  8/21/19		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190820_1318

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190820_1318\20190820_1318.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	8/20/2019	1:19:30 PM	1	5.00	22	4	26	211.46		
1	8/20/2019	1:25:39 PM	2	5.00	22	6	28	290.73		
1	8/20/2019	1:31:57 PM	3	5.00	23	5	28	229.17		
1	8/20/2019	1:38:06 PM	4	5.00	26	6	32	216.13		
1	8/20/2019	1:44:21 PM	5	5.00	25	7	32	284.73		
1	8/20/2019	1:50:38 PM	6	5.00	26	5	31	224.41		
1	8/20/2019	1:56:57 PM	7	5.00	26	5	31	247.04		
1	8/20/2019	2:03:12 PM	8	5.00	23	5	28	214.68		
1	8/20/2019	2:09:29 PM	9	5.00	24	4	28	200.50		
1	8/20/2019	2:15:45 PM	10	5.00	19	5	24	277.64		
1	8/20/2019	2:21:59 PM	11	5.00	20	6	26	326.07		
1	8/20/2019	2:28:16 PM	12	5.00	24	6	30	231.97		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	8/20/2019	2:34:36 PM	13	5.00	20	5	25	282.63
1	8/20/2019	2:40:50 PM	14	5.00	22	7	29	322.92
1	8/20/2019	2:47:00 PM	15	5.00	22	4	26	203.36
1	8/20/2019	2:53:13 PM	16	5.00	24	7	31	321.54
1	8/20/2019	2:59:31 PM	17	5.00	19	5	23	264.30
1	8/20/2019	3:05:47 PM	18	5.00	24	6	30	246.22
1	8/20/2019	3:12:05 PM	19	5.00	26	4	30	209.37
1	8/20/2019	3:18:23 PM	20	5.00	21	6	26	333.73
1	8/20/2019	3:24:32 PM	21	5.00	18	5	23	318.59
1	8/20/2019	3:30:43 PM	22	5.00	22	6	28	254.16
1	8/20/2019	3:37:01 PM	23	5.00	27	6	32	291.15
1	8/20/2019	3:42:41 PM	24	5.00	14492	35470	49962	1006.81


LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 8/20/19	Time 1110	Swipe Instrument Quantasart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) <input checked="" type="checkbox"/> ON <input type="checkbox"/> OFF <input type="checkbox"/>					

LSC Printout Date: 8/20/2019

NOTE SWIPE AND SURVEY LOCATIONS ON MAP



Survey #	Location	CPM	Swipe # and Locations	Notes
1	Background	80		
2	Elevator Entrance	150		
3	South Floor	120		
4	RAM Safe	1200		
5	East Lab Floor	80		
6	North Bench Top	120		
7	West Floor	90		
8	Work Platform	100		
9	Calorimeter	100		
10	Computers	100		
11	DSC/TGA	80		
12	South Bench Top	100		
1-6	Background		16 DSC/TGA	
7	Elevator Entrance		17 South Bench Top	
8	South Floor			
9	RAM Safe			
10	East Lab Floor			
11	North Bench Top			
12	West Lab Floor			
13	Work Platform			
14	Calorimeter			
15	Computers			

Surveyed By: SC Date: 8/20/19

Swiped By: CMH Date: 8/20/19

Reviewed By: M. Heine Date: 8/21/19

M. Heine

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	35	Dates:	8/25/2019-8/31/2019	Counted On	8/27/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.81E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	28	*****	*****	
2.	BACKGROUND 2	27	*****	*****	
3.	BACKGROUND 3	22	*****	*****	
4.	BACKGROUND 4	32	*****	*****	
5.	BACKGROUND 5	30	*****	*****	
6.	BACKGROUND 6	31	*****	*****	
7.	B021 ELEVATOR ENTRANCE	25	M	M	
8.	B021 FENCED GATE	27	M	M	
9.	B021 RAM SAFE	27	M	M	
10.	B021 EAST FLOOR	32	M	M	
11.	B021 NORTH BENCH TOP	27	M	M	
12.	B021 WEST LAB FLOOR	29	M	M	
13.	B021 WORK PLATFORM	29	M	M	
14.	B021 CALORIMETER	25	M	M	
15.	B021 COMPUTERS	30	M	M	
16.	B021 DSC/TGA	26	M	M	
17.	B021 SOUTH BENCH TOP	30	M	M	
18.	121 DOORWAY	25	M	M	
19.	121 GLOVE BOX- FLOOR	29	M	M	
20.	121 GLOVE BOX	27	M	M	
21.	121 WEST BENCH TOP	26	M	M	
22.	121 CENTER OF LAB	28	M	M	
23.	122 EAST BENCH TOP	28	M	M	
³⁶ CI SOURCE 09/14/1971		50143	*****	*****	
		101.10%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE: <i>B. Tanner</i> / 8/28/2019			SIGN/DATE: <i>M. Heine</i> 8/28/19		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190827_1646

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190827_1646\20190827_1646.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	8/27/2019	4:48:10 PM	1	5.00	23	5	28	194.04		
1	8/27/2019	4:54:20 PM	2	5.00	19	8	27	388.45		
1	8/27/2019	5:00:38 PM	3	5.00	19	3	22	193.68		
1	8/27/2019	5:06:48 PM	4	5.00	27	5	32	241.85		
1	8/27/2019	5:13:04 PM	5	5.00	24	6	30	272.47		
1	8/27/2019	5:19:21 PM	6	5.00	25	6	31	232.66		
1	8/27/2019	5:25:41 PM	7	5.00	20	5	25	273.95		
1	8/27/2019	5:31:56 PM	8	5.00	21	6	27	362.28		
1	8/27/2019	5:38:13 PM	9	5.00	21	6	27	268.48		
1	8/27/2019	5:44:28 PM	10	5.00	24	8	32	258.76		
1	8/27/2019	5:50:43 PM	11	5.00	21	6	27	246.15		
1	8/27/2019	5:56:57 PM	12	5.00	23	6	29	255.68		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	8/27/2019	6:03:18 PM	13	5.00	24	5	29	258.33
1	8/27/2019	6:09:33 PM	14	5.00	20	5	25	282.27
1	8/27/2019	6:15:42 PM	15	5.00	25	5	30	276.22
1	8/27/2019	6:21:55 PM	16	5.00	22	5	26	273.09
1	8/27/2019	6:28:06 PM	17	5.00	23	7	30	335.64
1	8/27/2019	6:34:17 PM	18	5.00	21	4	25	158.82
1	8/27/2019	6:40:38 PM	19	5.00	25	4	29	134.89
1	8/27/2019	6:46:52 PM	20	5.00	22	5	27	293.15
1	8/27/2019	6:53:10 PM	21	5.00	21	5	26	297.90
1	8/27/2019	6:59:29 PM	22	5.00	24	4	28	234.92
1	8/27/2019	7:05:40 PM	23	5.00	21	7	28	311.07
1	8/27/2019	7:11:20 PM	24	5.00	14423	35719	50143	1019.08

LABORATORY SURVEY


Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 8/27/19	Time 14:39	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma

Reactor Status (Check One)
 ON OFF

LSC Printout Date 8/27/2019

NOTE SWIPE AND SURVEY LOCATIONS ON MAP



Survey #	Location	CPM
1	Background	≤ 100
2	Elevator Entrance	≤ 100
3	South Floor	≤ 100
4	RAM Safe	1500
5	East Lab Floor	≤ 100
6	North Bench Top	260
7	West Floor	≤ 100
8	Work Platform	≤ 100
9	Calorimeter	≤ 100
10	Computers	450 NW 2, 100
11	DSC/TGA	150
12	South Bench Top	150


Swipe # and Locations	
1-6	Background 16 DSC/TGA
7	Elevator Entrance 17 South Bench Top
8	South Floor
9	RAM Safe
10	East Lab Floor
11	North Bench Top
12	West Lab Floor
13	Work Platform
14	Calorimeter
15	Computers

Surveyed By: MJ	Date: 8/22/19	Notes
Swiped By: MJ	Date: 8/27/19	
Reviewed By: <i>M. Heine</i>	Date: 8/28/19	

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	36	Dates:	9/01/2019-9/07/2019	Counted On	9/3/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.97E-08 uCi/cm ²	QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK		
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	31	*****	*****	
2.	BACKGROUND 2	29	*****	*****	
3.	BACKGROUND 3	31	*****	*****	
4.	BACKGROUND 4	32	*****	*****	
5.	BACKGROUND 5	27	*****	*****	
6.	BACKGROUND 6	32	*****	*****	
7.	B021 ELEVATOR ENTRANCE	35	M	M	
8.	B021 FENCED GATE	36	M	M	
9.	B021 RAM SAFE	26	M	M	
10.	B021 EAST FLOOR	31	M	M	
11.	B021 NORTH BENCH TOP	28	M	M	
12.	B021 WEST LAB FLOOR	29	M	M	
13.	B021 WORK PLATFORM	35	M	M	
14.	B021 CALORIMETER	29	M	M	
15.	B021 COMPUTERS	29	M	M	
16.	B021 DSC/TGA	33	M	M	
17.	B021 SOUTH BENCH TOP	32	M	M	
18.	121 DOORWAY	39	1.21E-05	1.21E-07	
19.	121 GLOVE BOX- FLOOR	32	M	M	
20.	121 GLOVE BOX	28	M	M	
21.	121 WEST BENCH TOP	26	M	M	
22.	121 CENTER OF LAB	26	M	M	
23.	122 EAST BENCH TOP	32	M	M	
³⁶ CI SOURCE 09/14/1971		50092	*****	*****	
		101.00%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE:  / 9/4/2019			SIGN/DATE:  9/4/19		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190903_1203

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190903_1203\20190903_1203.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	9/3/2019	12:08:11 PM	1	5.00	28	3	31	96.71		
1	9/3/2019	12:14:41 PM	2	5.00	23	5	29	250.03		
1	9/3/2019	12:20:58 PM	3	5.00	24	7	31	274.24		
1	9/3/2019	12:27:15 PM	4	5.00	25	7	32	234.64		
1	9/3/2019	12:33:33 PM	5	5.00	21	6	27	306.94		
1	9/3/2019	12:39:51 PM	6	5.00	26	5	32	240.13		
1	9/3/2019	12:46:11 PM	7	5.00	27	8	35	258.03		
1	9/3/2019	12:52:33 PM	8	5.00	26	10	36	260.09		
1	9/3/2019	12:58:53 PM	9	5.00	21	6	26	223.33		
1	9/3/2019	1:05:05 PM	10	5.00	26	5	31	193.09		
1	9/3/2019	1:11:21 PM	11	5.00	23	5	28	212.65		
1	9/3/2019	1:17:42 PM	12	5.00	23	6	29	218.64		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	9/3/2019	1:24:05 PM	13	5.00	29	6	35	222.18
1	9/3/2019	1:30:24 PM	14	5.00	22	8	29	242.16
1	9/3/2019	1:36:36 PM	15	5.00	25	5	29	222.93
1	9/3/2019	1:42:47 PM	16	5.00	25	8	33	207.27
1	9/3/2019	1:49:04 PM	17	5.00	25	7	32	265.60
1	9/3/2019	1:55:26 PM	18	5.00	26	13	39	284.72
1	9/3/2019	2:01:46 PM	19	5.00	26	7	32	210.78
1	9/3/2019	2:08:06 PM	20	5.00	22	6	28	254.47
1	9/3/2019	2:14:24 PM	21	5.00	20	6	26	246.04
1	9/3/2019	2:20:36 PM	22	5.00	21	5	26	203.49
1	9/3/2019	2:26:55 PM	23	5.00	26	5	32	160.89
1	9/3/2019	2:32:36 PM	24	5.00	14556	35536	50092	1016.23

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	37	Dates:	9/08/2019-9/14/2019	Counted On	9/10/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.71E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	30	*****	*****	
2.	BACKGROUND 2	27	*****	*****	
3.	BACKGROUND 3	26	*****	*****	
4.	BACKGROUND 4	31	*****	*****	
5.	BACKGROUND 5	23	*****	*****	
6.	BACKGROUND 6	25	*****	*****	
7.	B021 ELEVATOR ENTRANCE	26	M	M	
8.	B021 FENCED GATE	27	M	M	
9.	B021 RAM SAFE	28	M	M	
10.	B021 EAST FLOOR	30	M	M	
11.	B021 NORTH BENCH TOP	28	M	M	
12.	B021 WEST LAB FLOOR	29	M	M	
13.	B021 WORK PLATFORM	25	M	M	
14.	B021 CALORIMETER	25	M	M	
15.	B021 COMPUTERS	31	M	M	
16.	B021 DSC/TGA	29	M	M	
17.	B021 SOUTH BENCH TOP	25	M	M	
18.	121 DOORWAY	32	M	M	
19.	121 GLOVE BOX- FLOOR	28	M	M	
20.	121 GLOVE BOX	25	M	M	
21.	121 WEST BENCH TOP	31	M	M	
22.	121 CENTER OF LAB	30	M	M	
23.	122 EAST BENCH TOP	23	M	M	
³⁶ CI SOURCE 09/14/1971		50265	*****	*****	
		101.35%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE: <i>[Signature]</i> / 9/10/2019			SIGN/DATE: <i>Modolisen Price</i> 9/11/19		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190910_0856

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190910_0856\20190910_0856.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	9/10/2019	8:57:40 AM	1		5.00	25	6	30	251.27	
1	9/10/2019	9:03:58 AM	2		5.00	21	6	27	279.98	
1	9/10/2019	9:10:14 AM	3		5.00	20	6	26	337.17	
1	9/10/2019	9:16:29 AM	4		5.00	26	5	31	262.34	
1	9/10/2019	9:22:47 AM	5		5.00	20	3	23	218.53	
1	9/10/2019	9:29:03 AM	6		5.00	21	5	25	251.27	
1	9/10/2019	9:35:19 AM	7		5.00	19	6	26	340.90	
1	9/10/2019	9:41:38 AM	8		5.00	22	6	27	236.22	
1	9/10/2019	9:47:56 AM	9		5.00	22	7	28	319.08	
1	9/10/2019	9:54:06 AM	10		5.00	25	5	30	204.48	
1	9/10/2019	10:00:22 AM	11		5.00	23	5	28	195.64	
1	9/10/2019	10:06:42 AM	12		5.00	25	4	29	174.56	



Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	9/10/2019	10:13:02 AM	13	5.00	21	4	25	259.11
1	9/10/2019	10:19:20 AM	14	5.00	20	5	25	252.87
1	9/10/2019	10:25:39 AM	15	5.00	27	4	31	242.08
1	9/10/2019	10:31:51 AM	16	5.00	24	5	29	285.17
1	9/10/2019	10:38:09 AM	17	5.00	19	6	25	263.10
1	9/10/2019	10:44:30 AM	18	5.00	25	7	32	167.79
1	9/10/2019	10:50:50 AM	19	5.00	23	4	28	147.81
1	9/10/2019	10:57:05 AM	20	5.00	21	3	25	161.37
1	9/10/2019	11:03:24 AM	21	5.00	23	7	31	214.75
1	9/10/2019	11:09:44 AM	22	5.00	25	5	30	262.19
1	9/10/2019	11:15:55 AM	23	5.00	19	5	23	269.85
1	9/10/2019	11:21:36 AM	24	5.00	14554	35711	50265	1012.03

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	38	Dates:	9/15/2019-9/21/2019	Counted On	9/20/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		9.26E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	39	*****	*****	
2.	BACKGROUND 2	33	*****	*****	
3.	BACKGROUND 3	40	*****	*****	
4.	BACKGROUND 4	37	*****	*****	
5.	BACKGROUND 5	32	*****	*****	
6.	BACKGROUND 6	27	*****	*****	
7.	B021 ELEVATOR ENTRANCE	27	M	M	
8.	B021 FENCED GATE	26	M	M	
9.	B021 RAM SAFE	26	M	M	
10.	B021 EAST FLOOR	28	M	M	
11.	B021 NORTH BENCH TOP	25	M	M	
12.	B021 WEST LAB FLOOR	27	M	M	
13.	B021 WORK PLATFORM	27	M	M	
14.	B021 CALORIMETER	28	M	M	
15.	B021 COMPUTERS	30	M	M	
16.	B021 DSC/TGA	24	M	M	
17.	B021 SOUTH BENCH TOP	27	M	M	
18.	121 DOORWAY	30	M	M	
19.	121 GLOVE BOX- FLOOR	29	M	M	
20.	121 GLOVE BOX	28	M	M	
21.	121 WEST BENCH TOP	31	M	M	
22.	121 CENTER OF LAB	33	M	M	
23.	122 EAST BENCH TOP	30	M	M	
³⁶ CI SOURCE 09/14/1971		50155		*****	*****
			101.13%	CALIBRATION CHECK OK? YES	
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE:  / 9/24/19			SIGN/DATE:  9/24/19		

Assay Definition-

Assay Description:

Assay Type: CPM
Report Name: Report1
Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190920_1051
Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190920_1051\20190920_1051.results
Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor
Quench Indicator: tSIE/AEC
External Std Terminator (sec): 0.5 2s%
Pre-Count Delay (min): 0.00
Quench Set: n/a
Count Time (min): 5.00
Count Mode: Normal
Assay Count Cycles: 1 Repeat Sample Count: 1
#Vials/Sample: 1 Calculate % Reference: Off

Background Subtract: Off
Low CPM Threshold: Off
2 Sigma % Terminator: Off

Table with 3 columns: Regions, LL, UL. Rows A, B, C.

Count Corrections-

Static Controller: On Luminescence Correction: Off
Colored Samples: n/a Heterogeneity Monitor: n/a
Coincidence Time (nsec): 18 Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Table with 5 columns: Regions, Half Life, Units, Reference Date, Reference Time. Rows A, B, C.

Cycle 1 Results

Table with 10 columns: P#, DATE, TIME, S#, Count, Time, CPMA, CPMB, CPMC, SIS, MESSAGES. Rows 1-12.

1	9/20/2019	12:08:15 PM	13	5.00	22	5	27	183.03
1	9/20/2019	12:14:33 PM	14	5.00	22	6	28	231.55
1	9/20/2019	12:20:51 PM	15	5.00	23	6	30	247.82
1	9/20/2019	12:27:07 PM	16	5.00	21	3	24	182.89
1	9/20/2019	12:33:23 PM	17	5.00	20	7	27	293.45
1	9/20/2019	12:39:43 PM	18	5.00	24	7	30	232.74
1	9/20/2019	12:46:01 PM	19	5.00	24	5	29	237.84
1	9/20/2019	12:52:15 PM	20	5.00	23	5	28	236.69
1	9/20/2019	12:58:34 PM	21	5.00	25	6	31	254.50
1	9/20/2019	1:04:51 PM	22	5.00	27	6	33	201.05
1	9/20/2019	1:11:02 PM	23	5.00	25	5	30	131.81
1	9/20/2019	1:16:44 PM	24	5.00	14611	35544	50155	1008.71


LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 9/17/19	Time 10:37 10:18	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One)					
ON <input checked="" type="checkbox"/> OFF <input type="checkbox"/>					

LSC Printout Date 9/20/2019

NOTE SWIPE AND SURVEY LOCATIONS ON MAP



Survey #	Location	CPM	Swiped By:	Date:	Notes
13	Background	80	BT	9/17/19	SWIPES TAKEN WITH REACTOR OFF ON 9/20/19 AT 1046.
14	121 Doorway	140			
15	121 Center of lab	100			
16	121 Glove Box	170			
17					
18	121 Doorway				
19	121 Glove Box-floor				
20	121 Glove Box				
21	West Bench Top				
22	121 Center of lab				
23	East Bench Top				

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	39	Dates:	9/22/2019-9/28/2019	Counted On	9/25/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.83E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	27	*****	*****	
2.	BACKGROUND 2	30	*****	*****	
3.	BACKGROUND 3	27	*****	*****	
4.	BACKGROUND 4	31	*****	*****	
5.	BACKGROUND 5	30	*****	*****	
6.	BACKGROUND 6	26	*****	*****	
7.	B021 ELEVATOR ENTRANCE	26	M	M	
8.	B021 FENCED GATE	27	M	M	
9.	B021 RAM SAFE	24	M	M	
10.	B021 EAST FLOOR	25	M	M	
11.	B021 NORTH BENCH TOP	30	M	M	
12.	B021 WEST LAB FLOOR	27	M	M	
13.	B021 WORK PLATFORM	28	M	M	
14.	B021 CALORIMETER	24	M	M	
15.	B021 COMPUTERS	26	M	M	
16.	B021 DSC/TGA	27	M	M	
17.	B021 SOUTH BENCH TOP	26	M	M	
18.	121 DOORWAY	28	M	M	
19.	121 GLOVE BOX- FLOOR	27	M	M	
20.	121 GLOVE BOX	29	M	M	
21.	121 WEST BENCH TOP	29	M	M	
22.	121 CENTER OF LAB	26	M	M	
23.	122 EAST BENCH TOP	30	M	M	
³⁶ CI SOURCE 09/14/1971		50154	*****	*****	
		101.13%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE:  / 9/26/19			SIGN/DATE:  9/26/19		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190925_1337

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20190925_1337\20190925_1337.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	9/25/2019	1:38:58 PM	1		5.00	23	4	27	174.43	
1	9/25/2019	1:45:05 PM	2		5.00	26	5	30	167.78	
1	9/25/2019	1:51:20 PM	3		5.00	22	5	27	233.93	
1	9/25/2019	1:57:28 PM	4		5.00	25	6	31	167.87	
1	9/25/2019	2:03:42 PM	5		5.00	27	3	30	197.21	
1	9/25/2019	2:09:58 PM	6		5.00	21	5	26	212.18	
1	9/25/2019	2:16:17 PM	7		5.00	21	5	26	158.08	
1	9/25/2019	2:22:31 PM	8		5.00	22	5	27	260.51	
1	9/25/2019	2:28:48 PM	9		5.00	20	4	24	189.27	
1	9/25/2019	2:35:01 PM	10		5.00	20	5	25	211.84	
1	9/25/2019	2:41:17 PM	11		5.00	24	6	30	210.77	
1	9/25/2019	2:47:32 PM	12		5.00	23	3	27	141.09	

Protocol# 1 - Reactor Weekly.lsa


User: CSE

1	9/25/2019	2:53:51 PM	13	5.00	22	6	28	190.94
1	9/25/2019	3:00:07 PM	14	5.00	20	3	24	180.33
1	9/25/2019	3:06:25 PM	15	5.00	21	5	26	224.05
1	9/25/2019	3:12:39 PM	16	5.00	21	5	27	221.55
1	9/25/2019	3:18:55 PM	17	5.00	21	5	26	184.37
1	9/25/2019	3:25:14 PM	18	5.00	24	4	28	171.88
1	9/25/2019	3:31:30 PM	19	5.00	22	4	27	277.23
1	9/25/2019	3:37:44 PM	20	5.00	24	5	29	209.44
1	9/25/2019	3:44:00 PM	21	5.00	23	6	29	260.95
1	9/25/2019	3:50:14 PM	22	5.00	20	6	26	269.12
1	9/25/2019	3:56:24 PM	23	5.00	24	6	30	106.03
1	9/25/2019	4:02:10 PM	24	5.00	14484	35670	50154	1011.96

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 9/24/19	Time 0945	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One)		LSC Printout Date 9/25/2019			
ON <input checked="" type="checkbox"/>	OFF <input type="checkbox"/>				




Survey #	Location	CPM	Surveyed By:	Date:	Notes
13	Background	5100	AG	9/24/19	
14	121 Doorway	5100	BT	9/24/19	
15	121 Center of lab	5100	M. Heine	9/26/19	
16	121 Glove Box	1100			
18	121 Doorway				
19	121 Glove Box-floor				
20	121 Glove Box				
21	West Bench Top				
22	121 Center of lab				
23	East Bench Top				

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	40	Dates:	9/29/2019-10/05/2019	Counted On	10/2/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		9.14E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	37	*****	*****	
2.	BACKGROUND 2	34	*****	*****	
3.	BACKGROUND 3	29	*****	*****	
4.	BACKGROUND 4	30	*****	*****	
5.	BACKGROUND 5	36	*****	*****	
6.	BACKGROUND 6	31	*****	*****	
7.	B021 ELEVATOR ENTRANCE	28	M	M	
8.	B021 FENCED GATE	45	1.70E-05	1.70E-07	
9.	B021 RAM SAFE	31	M	M	
10.	B021 EAST FLOOR	25	M	M	
11.	B021 NORTH BENCH TOP	30	M	M	
12.	B021 WEST LAB FLOOR	28	M	M	
13.	B021 WORK PLATFORM	23	M	M	
14.	B021 CALORIMETER	27	M	M	
15.	B021 COMPUTERS	31	M	M	
16.	B021 DSC/TGA	38	M	M	
17.	B021 SOUTH BENCH TOP	32	M	M	
18.	121 DOORWAY	29	M	M	
19.	121 GLOVE BOX- FLOOR	30	M	M	
20.	121 GLOVE BOX	22	M	M	
21.	121 WEST BENCH TOP	29	M	M	
22.	121 CENTER OF LAB	30	M	M	
23.	122 EAST BENCH TOP	31	M	M	
³⁶ Cl SOURCE 09/14/1971		50026	*****	*****	
		100.87%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE:  10/3/2019			SIGN/DATE:  10/3/19		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191002_1659

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191002_1659\20191002_1659.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	10/2/2019	5:00:40 PM	1	5.00		33	5	37	232.92	
1	10/2/2019	5:06:50 PM	2	5.00		27	7	34	252.81	
1	10/2/2019	5:13:08 PM	3	5.00		24	5	29	285.50	
1	10/2/2019	5:19:19 PM	4	5.00		23	7	30	377.48	
1	10/2/2019	5:25:35 PM	5	5.00		30	6	36	275.60	
1	10/2/2019	5:31:52 PM	6	5.00		23	8	31	230.38	
1	10/2/2019	5:38:11 PM	7	5.00		23	5	28	237.79	
1	10/2/2019	5:44:26 PM	8	5.00		30	15	45	354.58	
1	10/2/2019	5:50:45 PM	9	5.00		25	6	31	269.77	
1	10/2/2019	5:57:01 PM	10	5.00		21	4	25	241.26	
1	10/2/2019	6:03:17 PM	11	5.00		23	7	30	372.87	
1	10/2/2019	6:09:33 PM	12	5.00		24	4	28	153.24	

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	10/2/2019	6:15:53 PM	13	5.00	18	5	23	272.22
1	10/2/2019	6:22:11 PM	14	5.00	23	5	27	284.87
1	10/2/2019	6:28:29 PM	15	5.00	24	6	31	230.49
1	10/2/2019	6:34:43 PM	16	5.00	28	11	38	330.65
1	10/2/2019	6:41:01 PM	17	5.00	27	5	32	256.55
1	10/2/2019	6:47:19 PM	18	5.00	25	4	29	237.57
1	10/2/2019	6:53:39 PM	19	5.00	24	5	30	221.40
1	10/2/2019	6:59:54 PM	20	5.00	17	5	22	321.67
1	10/2/2019	7:06:13 PM	21	5.00	25	4	29	186.36
1	10/2/2019	7:12:30 PM	22	5.00	24	6	30	242.54
1	10/2/2019	7:18:48 PM	23	5.00	28	3	31	110.58
1	10/2/2019	7:24:28 PM	24	5.00	14458	35568	50026	1020.88


LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 10/2/2019	Time 1632	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) <input type="checkbox"/> ON <input checked="" type="checkbox"/> OFF					

LSC Printout Date 10/2/2019

NOTE SWIPE AND SURVEY LOCATIONS ON MAP



Survey #	Location	CPM
1	Background	<100
2	Elevator Entrance	<100
3	South Floor	120
4	RAM Safe	900
5	East Lab Floor	<100
6	North Bench Top	<100
7	West Floor	<100
8	Work Platform	110
9	Calorimeter	<100
10	Computers	<100
11	DSC/TGA	200
12	South Bench Top	<100
13		
14		
15		

Swipe # and Locations	
1-6	Background 16 DSC/TGA
7	Elevator Entrance 17 South Bench Top
8	South Floor
9	RAM Safe
10	East Lab Floor
11	North Bench Top
12	West Lab Floor
13	Work Platform
14	Calorimeter
15	Computers

Surveyed By: <i>LD</i>	Date: <u>10/2/2019</u>	Notes
Swiped By: <i>B7</i>	Date: <u>10/2/2019</u>	
Reviewed By: <i>M. Herrin</i>	Date: <u>10/3/19</u>	

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	41	Dates:	10/06/2019-10/12/2019	Counted On	10/8/2019
INSTRUMENT USED: Quantasart (Serial# 073396)			CONTROL FACTOR: 1.40E-06		
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.85E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	29	*****	*****	
2.	BACKGROUND 2	30	*****	*****	
3.	BACKGROUND 3	30	*****	*****	
4.	BACKGROUND 4	26	*****	*****	
5.	BACKGROUND 5	26	*****	*****	
6.	BACKGROUND 6	32	*****	*****	
7.	B021 ELEVATOR ENTRANCE	27	M	M	
8.	B021 FENCED GATE	33	M	M	
9.	B021 RAM SAFE	23	M	M	
10.	B021 EAST FLOOR	28	M	M	
11.	B021 NORTH BENCH TOP	24	M	M	
12.	B021 WEST LAB FLOOR	25	M	M	
13.	B021 WORK PLATFORM	28	M	M	
14.	B021 CALORIMETER	27	M	M	
15.	B021 COMPUTERS	30	M	M	
16.	B021 DSC/TGA	28	M	M	
17.	B021 SOUTH BENCH TOP	26	M	M	
18.	121 DOORWAY	29	M	M	
19.	121 GLOVE BOX- FLOOR	30	M	M	
20.	121 GLOVE BOX	30	M	M	
21.	121 WEST BENCH TOP	29	M	M	
22.	121 CENTER OF LAB	30	M	M	
23.	122 EAST BENCH TOP	30	M	M	
³⁶ CI SOURCE 09/14/1971		50233		*****	*****
		101.28%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE: <i>B. Tanner</i> 10/9/19			SIGN/DATE: <i>Maddisen Heine</i> 10/10/19		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

Assay Definition-

Assay Description:

Assay Type: CPM
 Report Name: Report1
 Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191008_1409
 Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191008_1409\20191008_1409.results
 Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor
 Quench Indicator: tSIE/AEC
 External Std Terminator (sec): 0.5 2s%
 Pre-Count Delay (min): 0.00
 Quench Set: n/a
 Count Time (min): 5.00
 Count Mode: Normal
 Assay Count Cycles: 1 Repeat Sample Count: 1
 #Vials/Sample: 1 Calculate % Reference: Off

Background Subtract: Off
 Low CPM Threshold: Off
 2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On Luminescence Correction: Off
 Colored Samples: n/a Heterogeneity Monitor: n/a
 Coincidence Time (nsec): 18 Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	10/8/2019	2:10:28 PM	1	5.00		23	6	29	358.50	
1	10/8/2019	2:16:38 PM	2	5.00		26	4	30	212.69	
1	10/8/2019	2:22:55 PM	3	5.00		24	6	30	327.61	
1	10/8/2019	2:29:03 PM	4	5.00		21	5	26	317.71	
1	10/8/2019	2:35:19 PM	5	5.00		22	4	26	236.87	
1	10/8/2019	2:41:36 PM	6	5.00		28	4	32	180.74	
1	10/8/2019	2:47:55 PM	7	5.00		22	6	27	209.34	
1	10/8/2019	3:16:31 PM	8	5.00		27	6	33	201.46	
1	10/8/2019	3:22:50 PM	9	5.00		18	6	23	263.20	
1	10/8/2019	3:29:07 PM	10	5.00		24	4	28	262.39	
1	10/8/2019	3:35:19 PM	11	5.00		18	5	24	299.45	
1	10/8/2019	3:41:34 PM	12	5.00		20	4	25	249.27	

Protocol# 1 - Reactor Weekly.lsa


User: CSE

1	10/8/2019	3:47:50 PM	13	5.00	22	6	28	310.23
1	10/8/2019	3:54:06 PM	14	5.00	22	4	27	248.26
1	10/8/2019	4:00:14 PM	15	5.00	24	6	30	307.66
1	10/8/2019	4:06:28 PM	16	5.00	23	6	28	266.63
1	10/8/2019	4:12:43 PM	17	5.00	21	4	26	200.90
1	10/8/2019	4:19:00 PM	18	5.00	22	7	29	263.96
1	10/8/2019	4:25:19 PM	19	5.00	25	5	30	222.93
1	10/8/2019	4:31:36 PM	20	5.00	24	6	30	236.99
1	10/8/2019	4:37:54 PM	21	5.00	24	5	29	234.38
1	10/8/2019	4:44:08 PM	22	5.00	24	5	30	194.58
1	10/8/2019	4:50:31 PM	23	5.00	25	5	30	173.45
1	10/8/2019	4:56:11 PM	24	5.00	14609	35624	50233	1012.76

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 10/8/19	Time 1046	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One)		LSC Printout Date 10/8/2019			
ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>					



Survey #	Location	CPM
1	Background	≤100
2	Elevator Entrance	≤100
3	South Floor	≤100
4	RAM Safe	1200
5	East Lab Floor	≤100
6	North Bench Top	1100
7	West Floor	≤100
8	Work Platform	≤100
9	Calorimeter	≤100
10	Computers	≤100
11	DSC/TGA	≤100
12	South Bench Top	≤100
13	Background	
14	Elevator Entrance	
15	South Floor	
16	RAM Safe	
17	East Lab Floor	
18	North Bench Top	
19	West Lab Floor	
20	Work Platform	
21	Calorimeter	
22	Computers	

Surveyed By: AG	Date: 10/8/19
Swiped By: JS	Date: 10/8/19
Reviewed By: Madsen/Heine	Date: 10/10/19

Notes Survey #6 has high CPM due to radioactive samples on bench top

m. Heine

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 10/8/19	Time 1100	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One)		LSC Printout Date 10/8/2019			
ON <input type="checkbox"/>	OFF <input checked="" type="checkbox"/>				

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP		
13	Background	≤ 100			
14	121 Doorway	≤ 100			
15	121 Center of lab	≤ 100			
16	121 Glove Box	≤ 100			
Swipe # and Locations					
18	121 Doorway		DODG 121 GUO LAB Fume Hood East Bench Top West Bench Top Glove Box Surveyed By: JS Date: 10/8/19 Swiped By: Ag/JS Date: 10/8/19 Reviewed By: Madsen Date: 10/10/19 m. Heine		
19	121 Glove Box-floor				
20	121 Glove Box				
21	West Bench Top				
22	121 Center of lab				
23	East Bench Top				

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	42	Dates:	10/13/2019-10/19/2019	Counted On	10/15/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		9.11E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	30	*****	*****	
2.	BACKGROUND 2	38	*****	*****	
3.	BACKGROUND 3	33	*****	*****	
4.	BACKGROUND 4	27	*****	*****	
5.	BACKGROUND 5	34	*****	*****	
6.	BACKGROUND 6	32	*****	*****	
7.	B021 ELEVATOR ENTRANCE	45	1.77E-05	1.77E-07	
8.	B021 FENCED GATE	53	2.89E-05	2.89E-07	
9.	B021 RAM SAFE	37	M	M	
10.	B021 EAST FLOOR	50	2.47E-05	2.47E-07	
11.	B021 NORTH BENCH TOP	31	M	M	
12.	B021 WEST LAB FLOOR	56	3.31E-05	3.31E-07	
13.	B021 WORK PLATFORM	37	M	M	
14.	B021 CALORIMETER	35	M	M	
15.	B021 COMPUTERS	36	M	M	
16.	B021 DSC/TGA	31	M	M	
17.	B021 SOUTH BENCH TOP	35	M	M	
18.	121 DOORWAY	52	2.75E-05	2.75E-07	
19.	121 GLOVE BOX- FLOOR	28	M	M	
20.	121 GLOVE BOX	32	M	M	
21.	121 WEST BENCH TOP	30	M	M	
22.	121 CENTER OF LAB	40	1.07E-05	1.07E-07	
23.	122 EAST BENCH TOP	33	M	M	
³⁶ CI SOURCE 09/14/1971		50131		*****	*****
		101.08%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE: <i>B. Tanner</i> 10/16/2019			SIGN/DATE: <i>M. Heine</i> 10/23/19		

Assay Definition-

Assay Description:

Assay Type: CPM
Report Name: Report1
Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191015_1341
Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191015_1341\20191015_1341.results
Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor
Quench Indicator: tSIE/AEC
External Std Terminator (sec): 0.5 2s%
Pre-Count Delay (min): 0.00
Quench Set: n/a
Count Time (min): 5.00
Count Mode: Normal
Assay Count Cycles: 1 Repeat Sample Count: 1
#Vials/Sample: 1 Calculate % Reference: Off

Background Subtract: Off
Low CPM Threshold: Off
2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On Luminescence Correction: Off
Colored Samples: n/a Heterogeneity Monitor: n/a
Coincidence Time (nsec): 18 Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	10/15/2019	1:43:04 PM	1	5.00		25	5	30	226.48	
1	10/15/2019	1:49:14 PM	2	5.00		33	6	38	186.61	
1	10/15/2019	1:55:34 PM	3	5.00		26	7	33	301.87	
1	10/15/2019	2:01:45 PM	4	5.00		22	5	27	257.77	
1	10/15/2019	2:08:01 PM	5	5.00		28	6	34	213.57	
1	10/15/2019	2:14:17 PM	6	5.00		26	6	32	253.76	
1	10/15/2019	2:20:40 PM	7	5.00		34	11	45	191.43	
1	10/15/2019	2:26:56 PM	8	5.00		43	10	53	111.92	
1	10/15/2019	2:33:22 PM	9	5.00		27	10	37	245.11	
1	10/15/2019	2:39:42 PM	10	5.00		36	15	50	210.59	
1	10/15/2019	2:45:58 PM	11	5.00		24	7	31	294.09	
1	10/15/2019	2:52:18 PM	12	5.00		46	9	56	120.21	

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	10/15/2019	2:58:36 PM	13	5.00	29	8	37	186.55
1	10/15/2019	3:04:55 PM	14	5.00	26	9	35	307.60
1	10/15/2019	3:11:06 PM	15	5.00	29	7	36	176.60
1	10/15/2019	3:17:21 PM	16	5.00	24	7	31	307.26
1	10/15/2019	3:23:35 PM	17	5.00	27	8	35	287.06
1	10/15/2019	3:29:54 PM	18	5.00	36	15	52	251.32
1	10/15/2019	3:36:14 PM	19	5.00	24	5	28	165.85
1	10/15/2019	3:42:35 PM	20	5.00	26	6	32	210.26
1	10/15/2019	3:48:57 PM	21	5.00	23	7	30	210.27
1	10/15/2019	3:55:12 PM	22	5.00	32	8	40	117.11
1	10/15/2019	4:01:36 PM	23	5.00	27	6	33	190.54
1	10/15/2019	4:07:15 PM	24	5.00	14481	35650	50131	1014.49


LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 10/15/19	Time 1:22Z	Swipe Instrument Quantasart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input checked="" type="checkbox"/> OFF <input type="checkbox"/>					

LSC Printout Date 10/15/19

NOTE SWIPE AND SURVEY LOCATIONS ON MAP



Survey #	Location	CPM	Notes
1	Background	60	
2	Elevator Entrance	50	
3	South Floor	100	
4	RAM Safe	210	
5	East Lab Floor	100	
6	North Bench Top	120	
7	West Floor	120	
8	Work Platform	80	
9	Calorimeter	70	
10	Computers	150	
11	DSC/TGA	70	
12	South Bench Top	90	
13	work platform		
14	Calorimeter		
15	Computers		
16	Background		
17	Elevator Entrance		

Surveyed By: AH Date: 10/15/2019

Swiped By: AH AH CS Date: 10/15/2019

Reviewed By: MH Date: 10/23/2019

Maddison Devin m. heine

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.


LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 10/15/19	Time 1335	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) <input checked="" type="checkbox"/> ON <input type="checkbox"/> OFF <input type="checkbox"/>					

LSC Printout Date 10/15/19

NOTE SWIPE AND SURVEY LOCATIONS ON MAP



Survey #	Location	CPM	Notes
18	121 Doorway		
19	121 Glove Box-floor		
20	121 Glove Box		
21	West Bench Top		
22	121 Center of lab		
23	East Bench Top		
Surveyed By: <i>AH</i>			Date: <u>10/15/2019</u>
Swiped By: <i>CS</i>			Date: <u>10/15/2019</u>
Reviewed By: <i>Madison Dovic</i>			Date: <u>10/23/2019</u>
			<i>M. Heine</i>

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	43	Dates:	10/20/2019-10/26/2019	Counted On	10/23/2019
INSTRUMENT USED: Quantasart (Serial# 073396)			CONTROL FACTOR:		1.40E-06
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.93E-08 uCi/cm ²	QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK		
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	29	*****	*****	
2.	BACKGROUND 2	37	*****	*****	
3.	BACKGROUND 3	35	*****	*****	
4.	BACKGROUND 4	27	*****	*****	
5.	BACKGROUND 5	26	*****	*****	
6.	BACKGROUND 6	26	*****	*****	
7.	B021 ELEVATOR ENTRANCE	32	M	M	
8.	B021 FENCED GATE	31	M	M	
9.	B021 RAM SAFE	28	M	M	
10.	B021 EAST FLOOR	30	M	M	
11.	B021 NORTH BENCH TOP	29	M	M	
12.	B021 WEST LAB FLOOR	31	M	M	
13.	B021 WORK PLATFORM	33	M	M	
14.	B021 CALORIMETER	27	M	M	
15.	B021 COMPUTERS	26	M	M	
16.	B021 DSC/TGA	31	M	M	
17.	B021 SOUTH BENCH TOP	27	M	M	
18.	121 DOORWAY	29	M	M	
19.	121 GLOVE BOX- FLOOR	32	M	M	
20.	121 GLOVE BOX	28	M	M	
21.	121 WEST BENCH TOP	29	M	M	
22.	121 CENTER OF LAB	27	M	M	
23.	122 EAST BENCH TOP	23	M	M	
³⁶ Cl SOURCE 09/14/1971		50165	*****	*****	
			101.15%	CALIBRATION CHECK OK?	YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE: <i>B. Tanner</i> / 10/23/19			SIGN/DATE: <i>M. Heine</i> / 10/23/19		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191023_1046

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191023_1046\20191023_1046.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	10/23/2019	10:47:22 AM	1	5.00	23	6	29	286.31		
1	10/23/2019	10:53:40 AM	2	5.00	32	5	37	189.57		
1	10/23/2019	10:59:58 AM	3	5.00	31	4	35	195.18		
1	10/23/2019	11:06:12 AM	4	5.00	23	5	27	248.32		
1	10/23/2019	11:12:32 AM	5	5.00	23	4	26	205.58		
1	10/23/2019	11:18:50 AM	6	5.00	22	4	26	267.43		
1	10/23/2019	11:25:10 AM	7	5.00	26	6	32	213.35		
1	10/23/2019	11:31:28 AM	8	5.00	25	6	31	188.95		
1	10/23/2019	11:37:46 AM	9	5.00	23	5	28	241.25		
1	10/23/2019	11:44:02 AM	10	5.00	25	5	30	190.79		
1	10/23/2019	11:50:12 AM	11	5.00	24	5	29	276.13		
1	10/23/2019	11:56:31 AM	12	5.00	25	6	31	164.56		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	10/23/2019	12:02:48 PM	13	5.00	28	6	33	237.23
1	10/23/2019	12:09:06 PM	14	5.00	22	5	27	256.21
1	10/23/2019	12:15:16 PM	15	5.00	21	5	26	259.45
1	10/23/2019	12:21:31 PM	16	5.00	22	8	31	305.44
1	10/23/2019	12:27:45 PM	17	5.00	20	7	27	298.25
1	10/23/2019	12:34:02 PM	18	5.00	25	4	29	182.87
1	10/23/2019	12:40:21 PM	19	5.00	25	7	32	214.51
1	10/23/2019	12:46:39 PM	20	5.00	23	5	28	220.45
1	10/23/2019	12:52:55 PM	21	5.00	25	5	29	256.35
1	10/23/2019	12:59:12 PM	22	5.00	22	5	27	140.67
1	10/23/2019	1:05:28 PM	23	5.00	18	5	23	263.79
1	10/23/2019	1:11:08 PM	24	5.00	14257	35908	50165	1023.25

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 10/23/19	Time 1032	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One)		LSC Printout Date 10/23/2019			
ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>					

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP		
13	Background	50	<p>DODG 121 GUO LAB</p>		
14	121 Doorway	60			
15	121 Center of lab	60			
16	121 Glove Box	60			
18	121 Doorway		West Bench Top		
19	121 Glove Box-floor		Glove Box		
20	121 Glove Box				
21	West Bench Top				
22	121 Center of lab				
23	East Bench Top				
Swipe # and Locations			Notes		
			Surveyed By:	LL	Date: 10-23-19
			Swiped By:	DW	Date: 10-23-19
			Reviewed By:	Mobley	Date: 10/23/19
			m. Heine		

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	44	Dates:	10/27/2019-11/2/2019	Counted On	10/28/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.		uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100	
Limit of quantification (LOQ)=		9.17E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	40	*****	*****	
2.	BACKGROUND 2	36	*****	*****	
3.	BACKGROUND 3	27	*****	*****	
4.	BACKGROUND 4	28	*****	*****	
5.	BACKGROUND 5	31	*****	*****	
6.	BACKGROUND 6	38	*****	*****	
7.	B021 ELEVATOR ENTRANCE	36	M	M	
8.	B021 FENCED GATE	34	M	M	
9.	B021 RAM SAFE	29	M	M	
10.	B021 EAST FLOOR	30	M	M	
11.	B021 NORTH BENCH TOP	27	M	M	
12.	B021 WEST LAB FLOOR	29	M	M	
13.	B021 WORK PLATFORM	28	M	M	
14.	B021 CALORIMETER	32	M	M	
15.	B021 COMPUTERS	31	M	M	
16.	B021 DSC/TGA	24	M	M	
17.	B021 SOUTH BENCH TOP	30	M	M	
18.	121 DOORWAY	27	M	M	
19.	121 GLOVE BOX- FLOOR	24	M	M	
20.	121 GLOVE BOX	30	M	M	
21.	121 WEST BENCH TOP	25	M	M	
22.	121 CENTER OF LAB	25	M	M	
23.	122 EAST BENCH TOP	31	M	M	
³⁶ Cl SOURCE 09/14/1971		50112		*****	*****
		101.04%		CALIBRATION CHECK OK? YES	
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: A. Gomez			REVIEWED BY: M. Heine		
SIGN/DATE:  / 10-30-2019			SIGN/DATE:  / 10/30/2019		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191028_1507

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191028_1507\20191028_1507.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	10/28/2019	3:08:34 PM	1		5.00	31	9	40	275.45	
1	10/28/2019	3:14:47 PM	2		5.00	28	7	36	261.67	
1	10/28/2019	3:21:03 PM	3		5.00	23	4	27	252.98	
1	10/28/2019	3:27:19 PM	4		5.00	21	6	28	272.78	
1	10/28/2019	3:33:38 PM	5		5.00	25	5	31	317.94	
1	10/28/2019	3:39:57 PM	6		5.00	33	6	38	218.06	
1	10/28/2019	3:46:14 PM	7		5.00	30	6	36	251.50	
1	10/28/2019	3:52:28 PM	8		5.00	28	6	34	299.06	
1	10/28/2019	3:58:44 PM	9		5.00	22	7	29	352.75	
1	10/28/2019	4:05:00 PM	10		5.00	23	7	30	337.18	
1	10/28/2019	4:11:10 PM	11		5.00	21	6	27	280.70	
1	10/28/2019	4:17:26 PM	12		5.00	22	7	29	319.11	

Protocol# 1 - Reactor Weekly.lsa

User: CSE


1	10/28/2019	4:23:50 PM	13	5.00	22	7	28	295.75
1	10/28/2019	4:30:08 PM	14	5.00	26	6	32	336.23
1	10/28/2019	4:36:25 PM	15	5.00	26	5	31	258.54
1	10/28/2019	4:42:39 PM	16	5.00	19	5	24	324.57
1	10/28/2019	4:48:55 PM	17	5.00	23	7	30	314.36
1	10/28/2019	4:55:10 PM	18	5.00	22	5	27	279.77
1	10/28/2019	5:01:25 PM	19	5.00	18	6	24	303.00
1	10/28/2019	5:07:42 PM	20	5.00	25	5	30	299.59
1	10/28/2019	5:14:00 PM	21	5.00	20	5	25	276.32
1	10/28/2019	5:20:16 PM	22	5.00	21	4	25	280.34
1	10/28/2019	5:26:33 PM	23	5.00	27	4	31	236.85
1	10/28/2019	5:32:14 PM	24	5.00	14163	35949	50112	1034.79

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 10/28/19	Time 15:34	Swipe Instrument Quantasmart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>					

LSC Printout Date **10/28/19**



Survey #	Location	CPM	Notes
1	Background	20	
2	Elevator Entrance	30	
3	South Floor	100	
4	RAM Safe	450	
5	East Lab Floor	80	
6	North Bench Top	120	
7	West Floor	80	
8	Work Platform	80	
9	Calorimeter	100	
10	Computers	80	
11	DSC/TGA	100	
12	South Bench Top	100	
Swipe # and Locations			
1-6	Background	16	DSC/TGA
7	Elevator Entrance	17	South Bench Top
8	South Floor		
9	RAM Safe		
10	East Lab Floor		
11	North Bench Top		
12	West Lab Floor		
13	Work Platform		
14	Calorimeter		
15	Computers		

Surveyed By: <i>BT</i>	Date: 10/28/19
Swiped By: <i>TN</i>	Date: 10/28/19
Reviewed By: <i>M. Heine</i>	Date: 10/30/19

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916


Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 10/28/19	Time 15:53	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>		LSC Printout Date 10/28/19			

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP		
13	Background	80	<p>DODG 121 GUO LAB</p>		
14	121 Doorway	80			
15	121 Center of lab	146			
16	121 Glove Box	100			
Swipe # and Locations					
18	121 Doorway		Surveyed By:	BT	Date: 10/28/19
19	121 Glove Box-floor		Swiped By:	TN	Date: 10/28/19
20	121 Glove Box		Reviewed By:	<i>Matthew Davis</i>	Date: 10/30/19
21	West Bench Top		<i>m-there</i>		
22	121 Center of lab				
23	East Bench Top				

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	45	Dates:	11/3/2019-11/9/2019	Counted On	11/5/2019
INSTRUMENT USED: Quantasart (Serial# 073396)			CONTROL FACTOR: 1.40E-06		
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.75E-08 uCi/cm ²	QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK		
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	28	*****	*****	
2.	BACKGROUND 2	27	*****	*****	
3.	BACKGROUND 3	28	*****	*****	
4.	BACKGROUND 4	27	*****	*****	
5.	BACKGROUND 5	26	*****	*****	
6.	BACKGROUND 6	29	*****	*****	
7.	B021 ELEVATOR ENTRANCE	27	M	M	
8.	B021 FENCED GATE	25	M	M	
9.	B021 RAM SAFE	30	M	M	
10.	B021 EAST FLOOR	25	M	M	
11.	B021 NORTH BENCH TOP	30	M	M	
12.	B021 WEST LAB FLOOR	28	M	M	
13.	B021 WORK PLATFORM	29	M	M	
14.	B021 CALORIMETER	22	M	M	
15.	B021 COMPUTERS	25	M	M	
16.	B021 DSC/TGA	27	M	M	
17.	B021 SOUTH BENCH TOP	27	M	M	
18.	121 DOORWAY	29	M	M	
19.	121 GLOVE BOX- FLOOR	25	M	M	
20.	121 GLOVE BOX	25	M	M	
21.	121 WEST BENCH TOP	31	M	M	
22.	121 CENTER OF LAB	32	M	M	
23.	122 EAST BENCH TOP	26	M	M	
³⁶ Cl SOURCE 09/14/1971		50121	*****	*****	
		101.06%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE:  11/5/2019			SIGN/DATE:  11/12/19		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191105_1321

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191105_1321\20191105_1321.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	11/5/2019	1:22:46 PM	1		5.00	22	7	28	258.88	
1	11/5/2019	1:29:04 PM	2		5.00	21	6	27	270.55	
1	11/5/2019	1:35:23 PM	3		5.00	22	6	28	290.46	
1	11/5/2019	1:41:39 PM	4		5.00	21	6	27	304.61	
1	11/5/2019	1:47:58 PM	5		5.00	20	6	26	327.01	
1	11/5/2019	1:54:18 PM	6		5.00	25	5	29	269.17	
1	11/5/2019	2:00:36 PM	7		5.00	21	7	27	276.26	
1	11/5/2019	2:06:52 PM	8		5.00	21	4	25	253.79	
1	11/5/2019	2:13:10 PM	9		5.00	23	7	30	261.91	
1	11/5/2019	2:19:27 PM	10		5.00	20	5	25	277.33	
1	11/5/2019	2:25:36 PM	11		5.00	24	6	30	319.26	
1	11/5/2019	2:31:54 PM	12		5.00	22	6	28	234.23	

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	11/5/2019	2:38:14 PM	13	5.00	25	4	29	217.36
1	11/5/2019	2:44:32 PM	14	5.00	18	4	22	263.30
1	11/5/2019	2:50:43 PM	15	5.00	21	4	25	223.64
1	11/5/2019	2:56:53 PM	16	5.00	22	5	27	305.77
1	11/5/2019	3:03:11 PM	17	5.00	22	5	27	313.52
1	11/5/2019	3:09:29 PM	18	5.00	23	6	29	260.83
1	11/5/2019	3:15:48 PM	19	5.00	20	5	25	260.31
1	11/5/2019	3:22:07 PM	20	5.00	19	5	25	346.86
1	11/5/2019	3:28:27 PM	21	5.00	24	7	31	297.44
1	11/5/2019	3:34:39 PM	22	5.00	27	5	32	190.22
1	11/5/2019	3:40:56 PM	23	5.00	22	4	26	173.29
1	11/5/2019	3:46:37 PM	24	5.00	14605	35516	50121	1014.84

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 11/5/2019	Time 1122	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One)		LSC Printout Date 11/5/2019			
ON <input type="checkbox"/>	OFF <input checked="" type="checkbox"/>				

NOTE SWIPE AND SURVEY LOCATIONS ON MAP

DODG B0021 GUO LAB

Survey #	Location	CPM
1	Background	85
2	Elevator Entrance	80
3	South Floor	120
4	RAM Safe	700
5	East Lab Floor	80
6	North Bench Top	80
7	West Floor	60
8	Work Platform	100
9	Calorimeter	80
10	Computers	80
11	DSC/TGA	100
12	South Bench Top	30
13		
14		
15		

Swipe # and Locations	
1-6	Background 16 DSC/TGA
7	Elevator Entrance 17 South Bench Top
8	South Floor
9	RAM Safe
10	East Lab Floor
11	North Bench Top
12	West Lab Floor
13	Work Platform
14	Calorimeter
15	Computers

Surveyed By: BT	Date: 11/5/19	Notes
Swiped By: AB	Date: 11/05/19	
Reviewed By: M. Heine	Date: 11/12/19	

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	46	Dates:	11/10/2019-11/16/2019	Counted On	11/12/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.90E-08 uCi/cm ²	QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK		
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	31	*****	*****	
2.	BACKGROUND 2	28	*****	*****	
3.	BACKGROUND 3	28	*****	*****	
4.	BACKGROUND 4	29	*****	*****	
5.	BACKGROUND 5	27	*****	*****	
6.	BACKGROUND 6	34	*****	*****	
7.	B021 ELEVATOR ENTRANCE	28	M	M	
8.	B021 FENCED GATE	29	M	M	
9.	B021 RAM SAFE	29	M	M	
10.	B021 EAST FLOOR	29	M	M	
11.	B021 NORTH BENCH TOP	27	M	M	
12.	B021 WEST LAB FLOOR	31	M	M	
13.	B021 WORK PLATFORM	30	M	M	
14.	B021 CALORIMETER	30	M	M	
15.	B021 COMPUTERS	28	M	M	
16.	B021 DSC/TGA	27	M	M	
17.	B021 SOUTH BENCH TOP	28	M	M	
18.	121 DOORWAY	31	M	M	
19.	121 GLOVE BOX- FLOOR	29	M	M	
20.	121 GLOVE BOX	25	M	M	
21.	121 WEST BENCH TOP	26	M	M	
22.	121 CENTER OF LAB	28	M	M	
23.	122 EAST BENCH TOP	28	M	M	
³⁶ Cl SOURCE 09/14/1971		50193	*****	*****	
		101.20%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE: <i>B. Tanner</i> / 11/12/19			SIGN/DATE: <i>Maddisen Dru</i> 11/14/19		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191112_1335

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191112_1335\20191112_1335.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	11/12/2019	1:36:19 PM	1	5.00		24	7	31	266.80	
1	11/12/2019	1:42:29 PM	2	5.00		23	5	28	203.59	
1	11/12/2019	1:48:45 PM	3	5.00		23	5	28	256.55	
1	11/12/2019	1:54:53 PM	4	5.00		24	5	29	256.73	
1	11/12/2019	2:01:08 PM	5	5.00		23	5	27	231.88	
1	11/12/2019	2:07:24 PM	6	5.00		26	8	34	273.40	
1	11/12/2019	2:13:41 PM	7	5.00		21	6	28	281.71	
1	11/12/2019	2:19:54 PM	8	5.00		24	5	29	251.84	
1	11/12/2019	2:26:11 PM	9	5.00		23	5	29	191.45	
1	11/12/2019	2:32:27 PM	10	5.00		23	6	29	240.24	
1	11/12/2019	2:38:40 PM	11	5.00		20	7	27	281.23	
1	11/12/2019	2:44:57 PM	12	5.00		25	6	31	222.20	

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	11/12/2019	2:51:16 PM	13	5.00	24	6	30	298.35
1	11/12/2019	2:57:31 PM	14	5.00	24	5	30	281.05
1	11/12/2019	3:03:39 PM	15	5.00	25	3	28	195.20
1	11/12/2019	3:09:52 PM	16	5.00	22	5	27	212.05
1	11/12/2019	3:16:05 PM	17	5.00	22	5	28	289.82
1	11/12/2019	3:22:15 PM	18	5.00	25	5	31	257.83
1	11/12/2019	3:28:30 PM	19	5.00	22	7	29	302.62
1	11/12/2019	3:34:44 PM	20	5.00	22	3	25	235.49
1	11/12/2019	3:41:00 PM	21	5.00	22	4	26	212.43
1	11/12/2019	3:47:17 PM	22	5.00	22	5	28	221.86
1	11/12/2019	3:53:25 PM	23	5.00	23	6	28	271.48
1	11/12/2019	3:59:05 PM	24	5.00	14561	35632	50193	1018.49


LABORATORY SURVEY

Radiation Safety Office
Washington State University
Pullman, WA 99164-1302
(509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 11/12/2019	Time 0817	Swipe Instrument Quantasart	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>				

LSC Printout Date **11/12/2019**

NOTE SWIPE AND SURVEY LOCATIONS ON MAP



Survey #	Location	CPM	Notes
1	Background	80	
2	Elevator Entrance	80	
3	South Floor	80	
4	RAM Safe	1500	
5	East Lab Floor	96	
6	North Bench Top	100	
7	West Floor	80	
8	Work Platform	60	
9	Calorimeter	106	
10	Computers	80	
11	DSC/TGA	80	
12	South Bench Top	80	
13	Background	16	DSC/TGA
14	Elevator Entrance	17	South Bench Top
15	South Floor		
16	RAM Safe		
17	East Lab Floor		
18	North Bench Top		
19	West Lab Floor		
20	Work Platform		
21	Calorimeter		
22	Computers		

Surveyed By: BT	Date: 11/12/19
Swiped By: JS	Date: 11/12/19
Reviewed By: <i>M. Heine</i>	Date: 11/14/19

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	47	Dates:	11/17/2019-11/23/2019	Counted On	11/20/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.		uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100	
Limit of quantification (LOQ)=		8.99E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	37	*****	*****	
2.	BACKGROUND 2	31	*****	*****	
3.	BACKGROUND 3	28	*****	*****	
4.	BACKGROUND 4	30	*****	*****	
5.	BACKGROUND 5	30	*****	*****	
6.	BACKGROUND 6	28	*****	*****	
7.	B021 ELEVATOR ENTRANCE	33	M	M	
8.	B021 FENCED GATE	31	M	M	
9.	B021 RAM SAFE	23	M	M	
10.	B021 EAST FLOOR	25	M	M	
11.	B021 NORTH BENCH TOP	24	M	M	
12.	B021 WEST LAB FLOOR	27	M	M	
13.	B021 WORK PLATFORM	27	M	M	
14.	B021 CALORIMETER	25	M	M	
15.	B021 COMPUTERS	28	M	M	
16.	B021 DSC/TGA	32	M	M	
17.	B021 SOUTH BENCH TOP	27	M	M	
18.	121 DOORWAY	29	M	M	
19.	121 GLOVE BOX- FLOOR	28	M	M	
20.	121 GLOVE BOX	26	M	M	
21.	121 WEST BENCH TOP	27	M	M	
22.	121 CENTER OF LAB	27	M	M	
23.	122 EAST BENCH TOP	25	M	M	
³⁶ CI SOURCE 09/14/1971		50192	*****	*****	
			101.20%	CALIBRATION CHECK OK?	YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE: <i>B. Tanner</i> / 11/21/19			SIGN/DATE: <i>Maddison Heine</i> 11/22/19		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191120_1410

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191120_1410\20191120_1410.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	11/20/2019	2:11:22 PM	1	5.00		28	9	37	291.42	
1	11/20/2019	2:17:31 PM	2	5.00		26	5	31	270.87	
1	11/20/2019	2:23:50 PM	3	5.00		25	3	28	192.75	
1	11/20/2019	2:30:02 PM	4	5.00		23	7	30	328.27	
1	11/20/2019	2:36:18 PM	5	5.00		24	6	30	260.04	
1	11/20/2019	2:42:36 PM	6	5.00		24	4	28	222.77	
1	11/20/2019	2:48:57 PM	7	5.00		24	8	33	346.91	
1	11/20/2019	2:55:12 PM	8	5.00		27	4	31	142.26	
1	11/20/2019	3:01:32 PM	9	5.00		19	5	23	288.31	
1	11/20/2019	3:07:49 PM	10	5.00		20	5	25	307.99	
1	11/20/2019	3:14:04 PM	11	5.00		19	5	24	262.30	
1	11/20/2019	3:20:19 PM	12	5.00		22	5	27	265.37	

Protocol# 1 - Reactor Weekly.lsa


User: CSE

1	11/20/2019	3:26:38 PM	13	5.00	23	4	27	270.82
1	11/20/2019	3:32:56 PM	14	5.00	21	4	25	230.03
1	11/20/2019	3:39:16 PM	15	5.00	22	6	28	290.31
1	11/20/2019	3:45:33 PM	16	5.00	27	5	32	215.33
1	11/20/2019	3:51:52 PM	17	5.00	22	5	27	311.18
1	11/20/2019	3:58:13 PM	18	5.00	24	6	29	347.73
1	11/20/2019	4:04:31 PM	19	5.00	24	4	28	184.66
1	11/20/2019	4:10:47 PM	20	5.00	21	6	26	300.58
1	11/20/2019	4:17:05 PM	21	5.00	22	5	27	261.32
1	11/20/2019	4:23:21 PM	22	5.00	21	6	27	322.35
1	11/20/2019	4:29:32 PM	23	5.00	20	5	25	278.95
1	11/20/2019	4:35:12 PM	24	5.00	14358	35834	50192	1021.34

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 11/20/19	Time 1340	Swipe Instrument Quantasart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One)		LSC Printout Date 11/20/2019			
ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>					



Survey #	Location	CPM	Swipe # and Locations	Notes
1-6	Background	90	16 DSC/TGA	
7	Elevator Entrance	100	17 South Bench Top	
8	South Floor	80		
9	RAM Safe	1200		
10	East Lab Floor	80		
11	North Bench Top	80		
12	West Floor	100		
13	Work Platform	60		
14	Calorimeter	80		
15	Computers	100		

Surveyed By: BT	Date: 11/20/19
Swiped By: BT	Date: 11/20/19
Reviewed By: Mabelle T. Hu	Date: 11/22/19

m. Heine

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.


LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 11/20/19	Time 1401	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>					

LSC Printout Date 11/22/2019

NOTE SWIPE AND SURVEY LOCATIONS ON MAP




Surveys		Notes	
18	121 Doorway	Surveyed By: BT	Date: 11/20/19
19	121 Glove Box-floor	Swiped By: BT	Date: 11/20/19
20	121 Glove Box	Reviewed By: <i>M. Heine</i>	Date: 11/22/19
21	West Bench Top		
22	121 Center of lab		
23	East Bench Top		

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.


GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	48	Dates:	11/24/2019-11/30/2019	Counted On	11/25/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.87E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	31	*****	*****	
2.	BACKGROUND 2	31	*****	*****	
3.	BACKGROUND 3	27	*****	*****	
4.	BACKGROUND 4	26	*****	*****	
5.	BACKGROUND 5	30	*****	*****	
6.	BACKGROUND 6	29	*****	*****	
7.	B021 ELEVATOR ENTRANCE	35	M	M	
8.	B021 FENCED GATE	46	2.38E-05	2.38E-07	
9.	B021 RAM SAFE	29	M	M	
10.	B021 EAST FLOOR	34	M	M	
11.	B021 NORTH BENCH TOP	23	M	M	
12.	B021 WEST LAB FLOOR	26	M	M	
13.	B021 WORK PLATFORM	27	M	M	
14.	B021 CALORIMETER	28	M	M	
15.	B021 COMPUTERS	34	M	M	
16.	B021 DSC/TGA	29	M	M	
17.	B021 SOUTH BENCH TOP	24	M	M	
18.	121 DOORWAY	33	M	M	
19.	121 GLOVE BOX- FLOOR	28	M	M	
20.	121 GLOVE BOX	23	M	M	
21.	121 WEST BENCH TOP	27	M	M	
22.	121 CENTER OF LAB	29	M	M	
23.	122 EAST BENCH TOP	31	M	M	
³⁶ CI SOURCE 09/14/1971		50094	*****	*****	
		101.00%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE:  / 11/25/19			SIGN/DATE:  11/26/19		

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	48	Dates:	11/24/2019-11/30/2019	Counted On	11/25/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.		uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100	
Limit of quantification (LOQ)=		8.87E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	31	*****	*****	
2.	BACKGROUND 2	31	*****	*****	
3.	BACKGROUND 3	27	*****	*****	
4.	BACKGROUND 4	26	*****	*****	
5.	BACKGROUND 5	30	*****	*****	
6.	BACKGROUND 6	29	*****	*****	
7.	B021 ELEVATOR ENTRANCE	35	M	M	
8.	B021 FENCED GATE	46	2.38E-05	2.38E-07	
9.	B021 RAM SAFE	29	M	M	
10.	B021 EAST FLOOR	34	M	M	
11.	B021 NORTH BENCH TOP	23	M	M	
12.	B021 WEST LAB FLOOR	26	M	M	
13.	B021 WORK PLATFORM	27	M	M	
14.	B021 CALORIMETER	28	M	M	
15.	B021 COMPUTERS	34	M	M	
16.	B021 DSC/TGA	29	M	M	
17.	B021 SOUTH BENCH TOP	24	M	M	
18.	121 DOORWAY	33	M	M	
19.	121 GLOVE BOX- FLOOR	28	M	M	
20.	121 GLOVE BOX	23	M	M	
21.	121 WEST BENCH TOP	27	M	M	
22.	121 CENTER OF LAB	29	M	M	
23.	122 EAST BENCH TOP	31	M	M	
³⁶ CI SOURCE 09/14/1971		50094		*****	*****
		101.00%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE:  / 11/25/19			SIGN/DATE:  11/26/19		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191125_1011

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191125_1011\20191125_1011.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	11/25/2019	10:12:41 AM	1	5.00		28	3	31	154.01	
1	11/25/2019	10:18:52 AM	2	5.00		27	4	31	200.03	
1	11/25/2019	10:25:11 AM	3	5.00		23	4	27	208.71	
1	11/25/2019	10:31:22 AM	4	5.00		21	5	26	259.44	
1	11/25/2019	10:37:41 AM	5	5.00		25	5	30	263.52	
1	11/25/2019	10:43:59 AM	6	5.00		22	6	29	304.21	
1	11/25/2019	10:50:19 AM	7	5.00		30	5	35	170.70	
1	11/25/2019	10:56:33 AM	8	5.00		40	6	46	173.37	
1	11/25/2019	11:02:52 AM	9	5.00		25	4	29	77.99	
1	11/25/2019	11:09:14 AM	10	5.00		28	7	34	231.44	
1	11/25/2019	11:15:30 AM	11	5.00		18	5	23	183.21	
1	11/25/2019	11:21:45 AM	12	5.00		21	6	26	309.97	

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	11/25/2019	11:28:07 AM	13	5.00	22	5	27	205.99
1	11/25/2019	11:34:22 AM	14	5.00	23	5	28	250.96
1	11/25/2019	11:40:34 AM	15	5.00	27	7	34	194.15
1	11/25/2019	11:46:46 AM	16	5.00	21	8	29	351.57
1	11/25/2019	11:53:01 AM	17	5.00	20	4	24	243.68
1	11/25/2019	11:59:13 AM	18	5.00	26	7	33	245.36
1	11/25/2019	12:05:37 PM	19	5.00	23	5	28	141.98
1	11/25/2019	12:11:54 PM	20	5.00	19	4	23	300.09
1	11/25/2019	12:18:13 PM	21	5.00	22	4	27	247.06
1	11/25/2019	12:24:33 PM	22	5.00	22	7	29	299.15
1	11/25/2019	12:30:44 PM	23	5.00	26	5	31	193.37
1	11/25/2019	12:36:25 PM	24	5.00	14445	35649	50094	1017.29


LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 11/25/19	Time 0948	Swipe Instrument Quantasart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input type="checkbox"/> OFF <input checked="" type="checkbox"/>					

LSC Printout Date 11/25/19

NOTE SWIPE AND SURVEY LOCATIONS ON MAP



Survey #	Location	CPM	Surveyed By:	Date:	Notes
1-6	Background	90	BT	11/25/19	
7	Elevator Entrance	90	BT	11/25/19	
8	South Floor	100	BT	11/25/19	
9	RAM Safe	1000	M. Heine	11/26/19	
10	East Lab Floor	80			
11	North Bench Top	80			
12	West Lab Floor	120			
13	Work Platform	80			
14	Calorimeter	90			
15	Computers	80			

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	49	Dates:	12/01/2019-12/07/2019	Counted On	12/4/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm2 = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		9.59E-08 uCi/cm ²	QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK		
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	46	*****	*****	
2.	BACKGROUND 2	40	*****	*****	
3.	BACKGROUND 3	38	*****	*****	
4.	BACKGROUND 4	39	*****	*****	
5.	BACKGROUND 5	38	*****	*****	
6.	BACKGROUND 6	37	*****	*****	
7.	B021 ELEVATOR ENTRANCE	37	M	M	
8.	B021 FENCED GATE	32	M	M	
9.	B021 RAM SAFE	33	M	M	
10.	B021 EAST FLOOR	28	M	M	
11.	B021 NORTH BENCH TOP	27	M	M	
12.	B021 WEST LAB FLOOR	31	M	M	
13.	B021 WORK PLATFORM	30	M	M	
14.	B021 CALORIMETER	27	M	M	
15.	B021 COMPUTERS	29	M	M	
16.	B021 DSC/TGA	30	M	M	
17.	B021 SOUTH BENCH TOP	28	M	M	
18.	121 DOORWAY	30	M	M	
19.	121 GLOVE BOX- FLOOR	28	M	M	
20.	121 GLOVE BOX	24	M	M	
21.	121 WEST BENCH TOP	27	M	M	
22.	121 CENTER OF LAB	31	M	M	
23.	122 EAST BENCH TOP	31	M	M	
³⁶ CI SOURCE 09/14/1971		50157	*****	*****	
		101.13%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE: <i>B. Tanner</i> 12/5/2019			SIGN/DATE: <i>Moodijen Heine</i> 12/5/19		

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191204_1605

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191204_1605\20191204_1605.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	12/4/2019	4:06:31 PM	1		5.00	34	13	46	327.46	
1	12/4/2019	4:12:39 PM	2		5.00	32	8	40	249.33	
1	12/4/2019	4:18:57 PM	3		5.00	27	11	38	343.37	
1	12/4/2019	4:25:07 PM	4		5.00	29	10	39	287.01	
1	12/4/2019	4:31:22 PM	5		5.00	30	8	38	267.91	
1	12/4/2019	4:37:39 PM	6		5.00	28	9	37	296.94	
1	12/4/2019	4:43:58 PM	7		5.00	27	10	37	276.83	
1	12/4/2019	4:50:13 PM	8		5.00	25	7	32	292.78	
1	12/4/2019	4:56:29 PM	9		5.00	26	7	33	263.79	
1	12/4/2019	5:02:44 PM	10		5.00	23	5	28	178.04	
1	12/4/2019	5:08:58 PM	11		5.00	21	6	27	216.55	
1	12/4/2019	5:15:15 PM	12		5.00	26	6	31	184.85	

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	12/4/2019	5:21:34 PM	13	5.00	23	7	30	316.12
1	12/4/2019	5:27:51 PM	14	5.00	22	6	27	255.42
1	12/4/2019	5:34:09 PM	15	5.00	24	4	29	171.30
1	12/4/2019	5:40:19 PM	16	5.00	23	7	30	317.91
1	12/4/2019	5:46:38 PM	17	5.00	23	4	28	166.47
1	12/4/2019	5:52:58 PM	18	5.00	24	6	30	213.48
1	12/4/2019	5:59:13 PM	19	5.00	23	5	28	164.13
1	12/4/2019	6:05:25 PM	20	5.00	21	4	24	156.70
1	12/4/2019	6:11:40 PM	21	5.00	22	5	27	192.22
1	12/4/2019	6:17:55 PM	22	5.00	23	7	31	233.08
1	12/4/2019	6:24:05 PM	23	5.00	25	6	31	234.82
1	12/4/2019	6:29:45 PM	24	5.00	14539	35617	50157	1014.42


LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 12/4/19	Time 1537	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) <input checked="" type="checkbox"/> ON <input type="checkbox"/> OFF					

LSC Printout Date 12/4/2019

NOTE SWIPE AND SURVEY LOCATIONS ON MAP



Survey #	Location	CPM	Notes
1-6	Background	80	
7	Elevator Entrance	160	DSC/TGA
8	South Floor	160	South Bench Top
9	RAM Safe	100	
10	East Lab Floor	80	
11	North Bench Top	80	
12	West Lab Floor	80	
13	Work Platform	80	
14	Calorimeter	80	
15	Computers	120	

Surveyed By: BT Date: 12/4/19
 Swiped By: FM Date: 12/4/19
 Reviewed By: Madsen Date: 12/5/19
m. Heine

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.


LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 12/4/19	Time 1555	Swipe Instrument Quantasart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input checked="" type="checkbox"/> OFF <input type="checkbox"/>					

LSC Printout Date: 12/4/2019

NOTE SWIPE AND SURVEY LOCATIONS ON MAP




Survey #	Location	CPM	Surveyed By:	Date:	Notes
13	Background	100	B.T.	12/4/19	
14	121 Doorway	80			
15	121 Center of lab	160			
16	121 Glove Box	100			
17					
18	121 Doorway				
19	121 Glove Box-floor				
20	121 Glove Box				
21	West Bench Top				
22	121 Center of lab				
23	East Bench Top				
			Swiped By: FM	Date: 12/4/19	
			Reviewed By: M. Heim	Date: 12/5/19	

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	50	Dates:	12/08/2019-12/14/2019	Counted On	12/10/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		8.80E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE:
1.	BACKGROUND 1	30	*****	*****	
2.	BACKGROUND 2	29	*****	*****	
3.	BACKGROUND 3	29	*****	*****	
4.	BACKGROUND 4	27	*****	*****	
5.	BACKGROUND 5	26	*****	*****	
6.	BACKGROUND 6	28	*****	*****	
7.	B021 ELEVATOR ENTRANCE	28	M	M	
8.	B021 FENCED GATE	30	M	M	
9.	B021 RAM SAFE	29	M	M	
10.	B021 EAST FLOOR	23	M	M	
11.	B021 NORTH BENCH TOP	21	M	M	
12.	B021 WEST LAB FLOOR	29	M	M	
13.	B021 WORK PLATFORM	26	M	M	
14.	B021 CALORIMETER	26	M	M	
15.	B021 COMPUTERS	30	M	M	
16.	B021 DSC/TGA	27	M	M	
17.	B021 SOUTH BENCH TOP	25	M	M	
18.	121 DOORWAY	32	M	M	
19.	121 GLOVE BOX- FLOOR	28	M	M	
20.	121 GLOVE BOX	28	M	M	
21.	121 WEST BENCH TOP	26	M	M	
22.	121 CENTER OF LAB	25	M	M	
23.	122 EAST BENCH TOP	27	M	M	
³⁶ CI SOURCE 09/14/1971		50188	*****	*****	
		101.19%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE:  / 12/10/2019			SIGN/DATE:  12/11/19		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191210_1406

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191210_1406\20191210_1406.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	12/10/2019	2:07:53 PM	1	5.00		26	5	30	229.88	
1	12/10/2019	2:14:02 PM	2	5.00		23	6	29	251.07	
1	12/10/2019	2:20:19 PM	3	5.00		24	6	29	279.85	
1	12/10/2019	2:26:29 PM	4	5.00		21	5	27	281.65	
1	12/10/2019	2:32:47 PM	5	5.00		21	5	26	231.14	
1	12/10/2019	2:39:03 PM	6	5.00		24	4	28	240.21	
1	12/10/2019	2:45:21 PM	7	5.00		23	6	28	216.42	
1	12/10/2019	2:51:35 PM	8	5.00		24	6	30	233.53	
1	12/10/2019	2:57:52 PM	9	5.00		22	7	29	206.59	
1	12/10/2019	3:04:06 PM	10	5.00		19	4	23	185.23	
1	12/10/2019	3:10:21 PM	11	5.00		17	4	21	164.50	
1	12/10/2019	3:16:38 PM	12	5.00		20	9	29	364.16	

Protocol# 1 - Reactor Weekly.lsa

User: CSE

1	12/10/2019	3:22:58	PM	13	5.00	20	5	26	221.17
1	12/10/2019	3:29:14	PM	14	5.00	22	4	26	172.07
1	12/10/2019	3:35:31	PM	15	5.00	26	4	30	187.75
1	12/10/2019	3:41:42	PM	16	5.00	22	5	27	250.86
1	12/10/2019	3:48:01	PM	17	5.00	21	4	25	131.70
1	12/10/2019	3:54:19	PM	18	5.00	27	5	32	154.25
1	12/10/2019	4:00:35	PM	19	5.00	22	6	28	196.17
1	12/10/2019	4:06:48	PM	20	5.00	21	7	28	221.46
1	12/10/2019	4:13:02	PM	21	5.00	21	5	26	166.28
1	12/10/2019	4:19:16	PM	22	5.00	19	6	25	291.71
1	12/10/2019	4:25:25	PM	23	5.00	21	5	27	262.56
1	12/10/2019	4:31:05	PM	24	5.00	14433	35756	50188	1014.26

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 12/10/2019	Time 10:47	Swipe Instrument Quantasmart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) <input checked="" type="checkbox"/> ON <input type="checkbox"/> OFF					

LSC Printout Date: 12/10/2019

NOTE SWIPE AND SURVEY LOCATIONS ON MAP

DODG B0021 GUO LAB

Map labels: Calorimeter (9), work platform (8), RAM Safe (4), Controller, Computers (10), DSC/TGA (11), Bench Top (6), Elevator Entrance (7), Bench Top (12), Bench Top (17).

Survey #	Location	CPM	Notes
1	Background	60	
2	Elevator Entrance	100	
3	South Floor	80	
4	RAM Safe	12-60	
5	East Lab Floor	60	
6	North Bench Top	100	
7	West Floor	120	
8	Work Platform	60	
9	Calorimeter	80	
10	Computers	80	
11	DSC/TGA	80	
12	South Bench Top	80	
Swipe # and Locations			
1-6	Background	16	DSC/TGA
7	Elevator Entrance	17	South Bench Top
8	South Floor		
9	RAM Safe		
10	East Lab Floor		
11	North Bench Top		
12	West Lab Floor		
13	Work Platform		
14	Calorimeter		
15	Computers		

Surveyed By: BT Date: 12/10/19

Swiped By: BT Date: 12/10/19

Reviewed By: *Maddison Liu* Date: 12/11/19

vn-Heine

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room 121- GUO	Lab Class D	Survey Instrument GM Detector Model 12	Serial No. 172178	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 12/10/2019	Time 1137	Swipe Instrument Quantasmat	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One) ON <input checked="" type="checkbox"/> OFF <input type="checkbox"/>					



LSC Printout Date 12/10/19

Survey #	Location	CPM	NOTE SWIPE AND SURVEY LOCATIONS ON MAP		
13	Background	36	<p>DODG 121 GUO LAB</p> <p>East Bench Top</p> <p>Fume Hood</p> <p>West Bench Top</p> <p>Glove Box</p> <p>14, 15, 16, 18, 19, 20, 21, 22, 23</p>		
14	121 Doorway	100			
15	121 Center of lab	80			
16	121 Glove Box	100			
Swipe # and Locations					
18	121 Doorway		Surveyed By:	DT	Date: 12/10/19
19	121 Glove Box-floor		Swiped By:	DT	Date: 12/10/19
20	121 Glove Box		Reviewed By:	M. Heine	Date: 12/11/19
21	West Bench Top				
22	121 Center of lab				
23	East Bench Top				

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

GUO LAB DODGEN B021/121 WEEKLY SWIPES & SURVEYS

Revised 7/9/2019

Week #	51	Dates:	12/15/2019-12/21/2019	Counted On	12/17/2019
INSTRUMENT USED: Quantasart (Serial# 073396)		CONTROL FACTOR:		1.40E-06	
CHANNEL ENERGY RANGE (keV)		CH A: 4-200	CH B: 200-2000	CH C: 4-2000	
COUNTING TIME: 5 Min.	uCi/swipe = CPM:C * Control Factor		uCi/cm ² = CPM:C * Control Factor / 100		
Limit of quantification (LOQ)=		9.24E-08 uCi/cm ²		QUANTIFICATION BELOW REG GUIDE 8.23 LIMITS OK	
Swipe #	LOCATION: ROOM # & AREA/OBJECT	GROSS CPM CHANNEL C	uCi/swipe	uCi/cm ²	NOTE
1.	BACKGROUND 1	39	*****	*****	
2.	BACKGROUND 2	32	*****	*****	
3.	BACKGROUND 3	37	*****	*****	
4.	BACKGROUND 4	33	*****	*****	
5.	BACKGROUND 5	31	*****	*****	
6.	BACKGROUND 6	33	*****	*****	
7.	B021 ELEVATOR ENTRANCE	29	M	M	
8.	B021 FENCED GATE	32	M	M	
9.	B021 RAM SAFE	32	M	M	
10.	B021 EAST FLOOR	25	M	M	
11.	B021 NORTH BENCH TOP	30	M	M	
12.	B021 WEST LAB FLOOR	29	M	M	
13.	B021 WORK PLATFORM	33	M	M	
14.	B021 CALORIMETER	29	M	M	
15.	B021 COMPUTERS	30	M	M	
16.	B021 DSC/TGA	28	M	M	
17.	B021 SOUTH BENCH TOP	29	M	M	
18.	121 DOORWAY	26	M	M	
19.	121 GLOVE BOX- FLOOR	27	M	M	
20.	121 GLOVE BOX	25	M	M	
21.	121 WEST BENCH TOP	30	M	M	
22.	121 CENTER OF LAB	26	M	M	
23.	122 EAST BENCH TOP	30	M	M	
³⁶ CI SOURCE 09/14/1971		50133	*****	*****	
		101.08%	CALIBRATION CHECK OK?		YES
Notes(s): M = below limit of quantification on the LSC. The limit of quantification is checked against lowest allowable contamination limit of 1 x 10 ⁻⁶ uCi/cm ² .					
SURVEY PREPARED BY: B. Tanner			REVIEWED BY: M. Heine		
SIGN/DATE:  12/18/2019			SIGN/DATE:  12/20/19		

Protocol# 1 - Reactor Weekly.lsa

User: CSE

Assay Definition-

Assay Description:

Assay Type: CPM

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191217_1526

Raw Results Path: C:\Packard\Tricarb\Results\CSE\Reactor Weekly\20191217_1526\20191217_1526.results

Assay File Name: C:\Packard\TriCarb\Assays\Reactor Weekly.lsa

Count Conditions-

Nuclide: Reactor

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set: n/a

Count Time (min): 5.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

#Vials/Sample: 1

Calculate % Reference: Off

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	4.0	200.0
B	200.0	2000.0
C	4.0	2000.0

Count Corrections-

Static Controller: On

Luminescence Correction: Off

Colored Samples: n/a

Heterogeneity Monitor: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

Half Life-

Half Life Correction: Off

Regions	Half Life	Units	Reference Date	Reference Time
A				
B				
C				

Cycle 1 Results

P#	DATE	TIME	S#	Count	Time	CPMA	CPMB	CPMC	SIS	MESSAGES
1	12/17/2019	3:28:05 PM	1	5.00		30	9	39	255.97	
1	12/17/2019	3:34:14 PM	2	5.00		25	7	32	231.38	
1	12/17/2019	3:40:30 PM	3	5.00		30	7	37	334.24	
1	12/17/2019	3:46:40 PM	4	5.00		26	7	33	242.16	
1	12/17/2019	3:52:56 PM	5	5.00		26	5	31	228.16	
1	12/17/2019	3:59:12 PM	6	5.00		25	7	33	295.85	
1	12/17/2019	4:05:29 PM	7	5.00		22	7	29	272.77	
1	12/17/2019	4:11:44 PM	8	5.00		25	7	32	251.45	
1	12/17/2019	4:17:59 PM	9	5.00		26	7	32	212.35	
1	12/17/2019	4:24:13 PM	10	5.00		20	5	25	196.64	
1	12/17/2019	4:30:28 PM	11	5.00		23	7	30	277.89	
1	12/17/2019	4:36:43 PM	12	5.00		24	5	29	230.04	

Protocol# 1 - Reactor Weekly.lsa


User: CSE

1	12/17/2019	4:43:03	PM	13	5.00	26	7	33	285.72
1	12/17/2019	4:49:19	PM	14	5.00	23	6	29	264.08
1	12/17/2019	4:55:34	PM	15	5.00	23	7	30	278.94
1	12/17/2019	5:01:47	PM	16	5.00	21	7	28	377.91
1	12/17/2019	5:08:05	PM	17	5.00	23	6	29	226.23
1	12/17/2019	5:14:23	PM	18	5.00	22	5	26	166.29
1	12/17/2019	5:20:39	PM	19	5.00	23	4	27	238.92
1	12/17/2019	5:26:52	PM	20	5.00	20	5	25	196.05
1	12/17/2019	5:33:08	PM	21	5.00	24	6	30	164.79
1	12/17/2019	5:39:22	PM	22	5.00	20	6	26	225.94
1	12/17/2019	5:45:31	PM	23	5.00	25	5	30	156.63
1	12/17/2019	5:51:10	PM	24	5.00	14411	35721	50133	1013.09

LABORATORY SURVEY

Radiation Safety Office
 Washington State University
 Pullman, WA 99164-1302
 (509) 335 8916

Building Dodgen Research Facility	Room B0021 - GUO	Lab Class D	Survey Instrument GM Detector Model 3	Serial No. 58680	Radiation Detected Gamma
Authorized User Xiaofeng Guo	Date 12/17/2019	Time 1500	Swipe Instrument Quantasmart	Serial No. 073396	Radiation Detected Beta/Gamma
Reactor Status (Check One)		LSC Printout Date 12/17/2019			
ON <input checked="" type="checkbox"/> OFF <input type="checkbox"/>					



Survey #	Location	CPM
1	Background	80
2	Elevator Entrance	80
3	South Floor	100
4	RAM Safe	966
5	East Lab Floor	80
6	North Bench Top	106
7	West Floor	80
8	Work Platform	80
9	Calorimeter	80
10	Computers	66
11	DSC/TGA	80
12	South Bench Top	67
Swipe # and Locations		
1-6	Background	16 DSC/TGA
7	Elevator Entrance	17 South Bench Top
8	South Floor	
9	RAM Safe	
10	East Lab Floor	
11	North Bench Top	
12	West Lab Floor	
13	Work Platform	
14	Calorimeter	
15	Computers	

Notes

Surveyed By: BT	Date: 12/17/2019
Swiped By: BT	Date: 12/17/2019
Reviewed By: <i>Matt Heine</i>	Date: 12/20/2019

Circled numbers on map indicate a swipe area of 100 square cm. Numbers not circled indicate a survey location.

